



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102685

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/10/2019	13:32
2	K2795-02		1	8	100.00	>212.0	>212.0	05/10/2019	13:55
3	K2795-02DUP		1	8	100.00	>212.0	>212.0	05/10/2019	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 102685

WorkList Name : FP51019

WorkList ID : 125898

Date : 05-10-2019 09:09:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/14/2019	Water	K2795-02	Flash Point	Cool 4 deg C	PSEG01	K31	TRUCK-803	05/09/2019	1010A

Date/Time 05/10/19 13:10
 Received by: KS
 Relinquished by: CP

Date/Time 05/10/19 14:25
 Received by: CP
 Relinquished by: KS