

LB102693


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

62306



A= 0.0000e+000

B= 1.6209e-004

C= -6.5168e-002

Rho= 0.9996359

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.034	-0.034	195	0.000	195				
0.2	0.200	0.152	-0.048	1337	0.0 %	1337				
2.5	2.500	2.581	0.081	16324	0.0 %	16324				
5.0	5.000	5.138	0.138	32103	0.0 %	32103				
7.5	7.500	7.329	-0.171	45619	0.0 %	45619				
10.0	10.000	10.034	0.034	62306	0.0 %	62306				

LB102693 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 May 2019 09:22:20

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	195	-	0.0000	7471B	PPB	11 May 2019 09:25:50	S	Std
0.2 - 1	50.2	1337	-	0.2000	7471B	PPB	11 May 2019 09:28:07	S	Std
2.5 - 1	52.5	16324	-	2.5000	7471B	PPB	11 May 2019 09:30:25	S	Std
5.0 - 1	55.0	32103	-	5.0000	7471B	PPB	11 May 2019 09:32:42	S	Std
7.5 - 1	57.5	45619	-	7.5000	7471B	PPB	11 May 2019 09:34:58	S	Std
10.0 - 1	510.0	62306	-	10.0000	7471B	PPB	11 May 2019 09:37:16	S	Std
ICV38 - 1	ICV38	24305	3.8744	-	7471B	PPB	11 May 2019 09:41:10	U	SMPL
ICB38 - 1	ICB38	164	-0.0386	-	7471B	PPB	11 May 2019 09:43:25	U	SMPL
CCV84 - 1	CCV84	31965	5.1160	-	7471B	PPB	11 May 2019 09:45:42	U	SMPL
CCB84 - 1	CCB84	138	-0.0428	-	7471B	PPB	11 May 2019 09:47:57	U	SMPL
CRA - 1	CRA	1371	0.1571	-	7471B	PPB	11 May 2019 09:50:12	U	SMPL
HighStd - 1	HighStd	63090	10.1609	-	7471B	PPB	11 May 2019 09:52:27	U	SMPL
ChkStd - 1	ChkStd	43131	6.9258	-	7471B	PPB	11 May 2019 09:54:43	U	SMPL
PB119652BL - 1	PBS	69	-0.0540	-	7471B	PPB	11 May 2019 09:57:00	U	SMPL
PB119652BS - 1	LCSS	24358	3.8830	-	7471B	PPB	11 May 2019 09:59:18	U	SMPL
K2764-01 - 1	P021-SS008-0206-02	4899	0.7289	-	7471B	PPB	11 May 2019 10:01:35	U	SMPL
K2764-02 - 1	P021-SS008-0612-01	5020	0.7485	-	7471B	PPB	11 May 2019 10:03:52	U	SMPL
K2764-03 - 1	P021-SS008-1218-01	4754	0.7054	-	7471B	PPB	11 May 2019 10:06:10	U	SMPL
K2764-04 - 1	P021-SS008-1824-01	2162	0.2853	-	7471B	PPB	11 May 2019 10:08:27	U	SMPL
K2764-05 - 1	P021-SS009-0002-01	6787	1.0349	-	7471B	PPB	11 May 2019 10:10:42	U	SMPL
CCV85 - 1	CCV85	31857	5.0985	-	7471B	PPB	11 May 2019 10:12:58	U	SMPL
CCB85 - 1	CCB85	-15	-0.0676	-	7471B	PPB	11 May 2019 10:15:13	U	SMPL
K2764-06 - 1	P021-SS009-0206-01	6664	1.0150	-	7471B	PPB	11 May 2019 10:17:30	U	SMPL
K2764-07 - 1	P021-SS009-0612-01	5841	0.8816	-	7471B	PPB	11 May 2019 10:19:45	U	SMPL
K2764-08 - 1	P021-SS010-0002-01	3069	0.4323	-	7471B	PPB	11 May 2019 10:22:01	U	SMPL
K2764-09 - 1	P021-SS010-0206-01	3561	0.5120	-	7471B	PPB	11 May 2019 10:24:16	U	SMPL
K2764-10 - 1	P021-SS010-0612-01	3523	0.5059	-	7471B	PPB	11 May 2019 10:26:32	U	SMPL
K2764-11 - 1	P026-SS001-0002-01	5212	0.7796	-	7471B	PPB	11 May 2019 10:28:50	U	SMPL
K2764-12 - 1	P026-SS001-0206-01	3431	0.4910	-	7471B	PPB	11 May 2019 10:31:06	U	SMPL
K2764-13 - 1	P026-SS001-0612-01	2633	0.3616	-	7471B	PPB	11 May 2019 10:33:23	U	SMPL
K2764-14 - 1	P026-SS001-1218-01	2380	0.3206	-	7471B	PPB	11 May 2019 10:35:40	U	SMPL
K2764-15 - 1	P026-SS001-1824-01	2068	0.2700	-	7471B	PPB	11 May 2019 10:37:56	U	SMPL
CCV86 - 1	CCV86	32123	5.1416	-	7471B	PPB	11 May 2019 10:40:11	U	SMPL
CCB86 - 1	CCB86	0	-0.0652	-	7471B	PPB	11 May 2019 10:42:27	U	SMPL
K2764-16 - 1	P026-SS002-0002-01	4290	0.6302	-	7471B	PPB	11 May 2019 10:44:43	U	SMPL
K2764-17 - 1	P026-SS002-0206-01	4593	0.6793	-	7471B	PPB	11 May 2019 10:46:59	U	SMPL
K2764-18 - 1	P026-SS002-0612-01	5256	0.7868	-	7471B	PPB	11 May 2019 10:49:15	U	SMPL
K2764-19 - 1	P026-SS002-1218-01	4233	0.6209	-	7471B	PPB	11 May 2019 10:51:31	U	SMPL
K2764-19DUP - 1	P026-SS002-1218-01DUP	4182	0.6127	-	7471B	PPB	11 May 2019 10:53:47	U	SMPL
K2764-20 - 1	K2764-19MS	28975	4.6313	-	7471B	PPB	11 May 2019 10:56:03	U	SMPL
K2764-21 - 1	K2764-19MSD	28517	4.5571	-	7471B	PPB	11 May 2019 10:58:19	U	SMPL
PB119653BL - 1	PBS	32	-0.0600	-	7471B	PPB	11 May 2019 11:00:37	U	SMPL
PB119653BS - 1	LCSS	24993	3.9859	-	7471B	PPB	11 May 2019 11:02:54	U	SMPL
K2765-01 - 1	P026-SS002-1218-02	4762	0.7067	-	7471B	PPB	11 May 2019 11:05:10	U	SMPL
CCV87 - 1	CCV87	33320	5.3356	-	7471B	PPB	11 May 2019 11:07:27	U	SMPL
CCB87 - 1	CCB87	8	-0.0639	-	7471B	PPB	11 May 2019 11:09:42	U	SMPL
K2765-02 - 1	P026-SS002-1824-01	3097	0.4368	-	7471B	PPB	11 May 2019 11:11:59	U	SMPL
K2765-03 - 1	P026-SS003-0002-01	3963	0.5772	-	7471B	PPB	11 May 2019 11:14:15	U	SMPL
K2765-04 - 1	P026-SS003-0206-01	4825	0.7169	-	7471B	PPB	11 May 2019 11:16:30	U	SMPL
K2765-05 - 1	P026-SS003-0612-01	3403	0.4864	-	7471B	PPB	11 May 2019 11:18:46	U	SMPL
K2765-06 - 1	P026-SS003-1218-01	1888	0.2409	-	7471B	PPB	11 May 2019 11:21:02	U	SMPL
K2765-07 - 1	P026-SS003-1824-01	1745	0.2177	-	7471B	PPB	11 May 2019 11:23:18	U	SMPL
K2765-08 - 1	P026-SS004-0002-01	4817	0.7156	-	7471B	PPB	11 May 2019 11:25:34	U	SMPL
K2765-09 - 1	P026-SS004-0206-01	2407	0.3250	-	7471B	PPB	11 May 2019 11:27:52	U	SMPL
K2765-10 - 1	P026-SS004-0612-01	1958	0.2522	-	7471B	PPB	11 May 2019 11:30:09	U	SMPL
K2765-11 - 1	P026-SS004-1218-01	2568	0.3511	-	7471B	PPB	11 May 2019 11:32:25	U	SMPL
CCV88 - 1	CCV88	33414	5.3508	-	7471B	PPB	11 May 2019 11:34:41	U	SMPL
CCB88 - 1	CCB88	26	-0.0610	-	7471B	PPB	11 May 2019 11:36:57	U	SMPL
K2765-12 - 1	P026-SS004-1824-01	3856	0.5598	-	7471B	PPB	11 May 2019 11:39:13	U	SMPL
K2765-13 - 1	P026-SS005-0002-01	3202	0.4538	-	7471B	PPB	11 May 2019 11:41:29	U	SMPL
K2765-14 - 1	P026-SS005-0206-01	3946	0.5744	-	7471B	PPB	11 May 2019 11:43:45	U	SMPL
K2765-15 - 1	P026-SS005-0612-01	3548	0.5099	-	7471B	PPB	11 May 2019 11:46:01	U	SMPL
K2765-16 - 1	P026-SS005-1218-01	3329	0.4744	-	7471B	PPB	11 May 2019 11:48:16	U	SMPL
K2765-17 - 1	P026-SS005-1824-01	4481	0.6611	-	7471B	PPB	11 May 2019 11:50:33	U	SMPL
K2765-18 - 1	P026-SS006-0002-01	3452	0.4944	-	7471B	PPB	11 May 2019 11:52:50	U	SMPL
K2765-18DUP - 1	P026-SS006-0002-01DUP	3445	0.4932	-	7471B	PPB	11 May 2019 11:55:07	U	SMPL
K2765-19 - 1	K2765-18MS	29196	4.6671	-	7471B	PPB	11 May 2019 11:57:24	U	SMPL
K2765-20 - 1	K2765-18MSD	28979	4.6320	-	7471B	PPB	11 May 2019 11:59:40	U	SMPL
CCV89 - 1	CCV89	33244	5.3233	-	7471B	PPB	11 May 2019 12:01:57	U	SMPL
CCB89 - 1	CCB89	4	-0.0645	-	7471B	PPB	11 May 2019 12:04:15	U	SMPL
K2765-21 - 1	P026-SS006-0002-02	3563	0.5123	-	7471B	PPB	11 May 2019 12:06:32	U	SMPL
PB119654BL - 1	PBS	72	-0.0535	-	7471B	PPB	11 May 2019 12:08:48	U	SMPL
PB119654BS - 1	LCSS	24725	3.9424	-	7471B	PPB	11 May 2019 12:11:05	U	SMPL
K2766-01 - 1	P026-SS006-0206-01	2316	0.3102	-	7471B	PPB	11 May 2019 12:13:21	U	SMPL
K2766-02 - 1	P026-SS006-0612-01	2546	0.3475	-	7471B	PPB	11 May 2019 12:15:38	U	SMPL
K2766-03 - 1	P026-SS006-1218-01	5435	0.8158	-	7471B	PPB	11 May 2019 12:17:54	U	SMPL
K2766-04 - 1	P026-SS006-1824-01	3107	0.4384	-	7471B	PPB	11 May 2019 12:20:11	U	SMPL
K2766-05 - 1	P026-SS007-0002-01	5650	0.8506	-	7471B	PPB	11 May 2019 12:22:28	U	SMPL
K2766-06 - 1	P026-SS007-0206-01	5395	0.8093	-	7471B	PPB	11 May 2019 12:24:45	U	SMPL
K2766-07 - 1	P026-SS007-0612-01	3582	0.5154	-	7471B	PPB	11 May 2019 12:27:01	U	SMPL
CCV90 - 1	CCV90	31859	5.0988	-	7471B	PPB	11 May 2019 12:29:17	U	SMPL

LB102693 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 May 2019 09:22:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB90 - 1	CCB90	14	-0.0629		-7471B	PPB	11 May 2019 12:31:34	U	SMPL
K2766-08 - 1	P026-SS007-1218-01	3607	0.5195		-7471B	PPB	11 May 2019 12:33:51	U	SMPL
K2766-09 - 1	P026-SS007-1824-01	2724	0.3764		-7471B	PPB	11 May 2019 12:36:07	U	SMPL
K2766-10 - 1	P026-SS008-0002-01	8023	1.2353		-7471B	PPB	11 May 2019 12:38:24	U	SMPL
K2766-11 - 1	P026-SS008-0206-01	9870	1.5346		-7471B	PPB	11 May 2019 12:40:40	U	SMPL
K2766-11DUP - 1	P026-SS008-0206-01DUP	9888	1.5376		-7471B	PPB	11 May 2019 12:42:56	U	SMPL
K2766-12 - 1	K2766-11MS	33138	5.3061		-7471B	PPB	11 May 2019 12:45:13	U	SMPL
K2766-13 - 1	K2766-11MSD	32593	5.2177		-7471B	PPB	11 May 2019 12:47:29	U	SMPL
K2766-14 - 1	P026-SS008-0206-02	9336	1.4481		-7471B	PPB	11 May 2019 12:49:47	U	SMPL
K2766-15 - 1	P026-SS008-0612-01	7904	1.2160		-7471B	PPB	11 May 2019 12:52:06	U	SMPL
K2766-16 - 1	P026-SS008-1218-01	3686	0.5323		-7471B	PPB	11 May 2019 12:54:25	U	SMPL
CCV91 - 1	CCV91	32252	5.1625		-7471B	PPB	11 May 2019 12:56:41	U	SMPL
CCB91 - 1	CCB91	47	-0.0575		-7471B	PPB	11 May 2019 12:58:58	U	SMPL
K2766-17 - 1	P026-SS008-1824-01	6064	0.9177		-7471B	PPB	11 May 2019 13:01:15	U	SMPL
K2766-18 - 1	P026-SS009-0002-01	3560	0.5119		-7471B	PPB	11 May 2019 13:03:32	U	SMPL
K2766-19 - 1	P026-SS009-0206-01	4042	0.5900		-7471B	PPB	11 May 2019 13:05:48	U	SMPL
K2766-20 - 1	P026-SS009-0612-01	2658	0.3657		-7471B	PPB	11 May 2019 13:08:04	U	SMPL
K2766-21 - 1	P026-SS009-1218-01	2668	0.3673		-7471B	PPB	11 May 2019 13:10:21	U	SMPL
K2764-19LX5 - 1		846	0.0720		-7471B	PPB	11 May 2019 13:12:37	U	SMPL
K2764-19A - 1		28856	4.6120		-7471B	PPB	11 May 2019 13:14:53	U	SMPL
K2765-18LX5 - 1		662	0.0421		-7471B	PPB	11 May 2019 13:17:10	U	SMPL
K2765-18A - 1		28699	4.5866		-7471B	PPB	11 May 2019 13:19:27	U	SMPL
K2766-11LX5 - 1		1967	0.2537		-7471B	PPB	11 May 2019 13:21:44	U	SMPL
CCV92 - 1	CCV92	31493	5.0395		-7471B	PPB	11 May 2019 13:24:01	U	SMPL
CCB92 - 1	CCB92	-13	-0.0673		-7471B	PPB	11 May 2019 13:26:18	U	SMPL
K2766-11A - 1		32380	5.1832		-7471B	PPB	11 May 2019 13:28:36	U	SMPL
CCV93 - 1	CCV93	30845	4.9344		-7471B	PPB	11 May 2019 13:30:52	U	SMPL
CCB93 - 1	CCB93	3	-0.0647		-7471B	PPB	11 May 2019 13:33:10	U	SMPL