

LB102705


7471B

INSTRUMENT ID : CU1

Linear

 μ Abs. :

60071



A= 0.0000e+000

B= 1.6576e-004

C= -3.6847e-002

Rho= 0.9998755

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	106	0.000	106				
0.2	0.200	0.160	-0.040	1190	0.0 %	1190				
2.5	2.500	2.499	-0.001	15298	0.0 %	15298				
5.0	5.000	5.104	0.104	31015	0.0 %	31015				
7.5	7.500	7.535	0.035	45680	0.0 %	45680				
10.0	10.000	9.921	-0.079	60071	0.0 %	60071				

LB102705 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 May 2019 12:03:24

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50-0	106	-	0.00007471B	PPB	13 May 2019 12:08:48	S	Std
0.2 - 1	50-2	1190	-	0.20007471B	PPB	13 May 2019 12:11:05	S	Std
2.5 - 1	52-5	15298	-	2.50007471B	PPB	13 May 2019 12:13:22	S	Std
5.0 - 1	55-0	31015	-	5.00007471B	PPB	13 May 2019 12:15:40	S	Std
7.5 - 1	57-5	45680	-	7.50007471B	PPB	13 May 2019 12:17:56	S	Std
10.0 - 1	510-0	60071	-	10.00007471B	PPB	13 May 2019 12:23:14	S	Std
ICV39 - 1	ICV39	23863	3.9187	-7471B	PPB	13 May 2019 12:26:24	U	SMPL
ICB39 - 1	ICB39	-1	-0.0370	-7471B	PPB	13 May 2019 12:28:40	U	SMPL
CCV94 - 1	CCV94	30957	5.0946	-7471B	PPB	13 May 2019 12:30:56	U	SMPL
CCB94 - 1	CCB94	-19	-0.0400	-7471B	PPB	13 May 2019 12:33:11	U	SMPL
CRA - 1	CRA	1150	0.1538	-7471B	PPB	13 May 2019 12:35:27	U	SMPL
HighStd - 1	HighStd	59914	9.8945	-7471B	PPB	13 May 2019 12:37:43	U	SMPL
ChkStd - 1	ChkStd	42074	6.9374	-7471B	PPB	13 May 2019 12:39:58	U	SMPL
PB119683BL - 1	PBS	-75	-0.0493	-7471B	PPB	13 May 2019 12:42:17	U	SMPL
PB119683BS - 1	LCSS	22783	3.7397	-7471B	PPB	13 May 2019 12:44:34	U	SMPL
K2769-01 - 1	P026-SS017-0002-01	4326	0.6802	-7471B	PPB	13 May 2019 12:46:51	U	SMPL
K2769-02 - 1	P026-SS017-0206-01	4182	0.6564	-7471B	PPB	13 May 2019 12:49:07	U	SMPL
K2769-03 - 1	P026-SS017-0612-01	1923	0.2819	-7471B	PPB	13 May 2019 12:51:25	U	SMPL
K2769-04 - 1	P026-SS017-1218-01	1977	0.2909	-7471B	PPB	13 May 2019 12:53:42	U	SMPL
K2769-05 - 1	P026-SS017-1824-01	1607	0.2295	-7471B	PPB	13 May 2019 12:55:57	U	SMPL
CCV95 - 1	CCV95	31133	5.1238	-7471B	PPB	13 May 2019 12:58:13	U	SMPL
CCB95 - 1	CCB95	37	-0.0307	-7471B	PPB	13 May 2019 13:00:28	U	SMPL
K2801-01 - 1	RT-1873	4007	0.6274	-7471B	PPB	13 May 2019 13:02:44	U	SMPL
K2801-01DUP - 1	RT-1873DUP	3934	0.6153	-7471B	PPB	13 May 2019 13:05:00	U	SMPL
K2801-01MS - 1	RT-1873MS	28009	4.6059	-7471B	PPB	13 May 2019 13:07:16	U	SMPL
K2801-01MSD - 1	RT-1873MSD	27812	4.5733	-7471B	PPB	13 May 2019 13:09:32	U	SMPL
K2801-04 - 1	RBR-200035	1608	0.2297	-7471B	PPB	13 May 2019 13:11:49	U	SMPL
K2801-07 - 1	RT-3456	3573	0.5554	-7471B	PPB	13 May 2019 13:14:06	U	SMPL
K2806-01 - 1	FK-EF-01-009-ABCD	3663	0.5703	-7471B	PPB	13 May 2019 13:16:23	U	SMPL
K2807-01 - 1	FK-GF-02-039-A	7856	1.2654	-7471B	PPB	13 May 2019 13:18:40	U	SMPL
K2807-03 - 1	FK-GF-02-039-B	7813	1.2582	-7471B	PPB	13 May 2019 13:20:57	U	SMPL
K2808-01 - 1	RT-2683-RT-5156	2702	0.4110	-7471B	PPB	13 May 2019 13:23:13	U	SMPL
CCV96 - 1	CCV96	31520	5.1879	-7471B	PPB	13 May 2019 13:25:29	U	SMPL
CCB96 - 1	CCB96	-58	-0.0465	-7471B	PPB	13 May 2019 13:27:45	U	SMPL
K2822-03 - 1	S-3	19403	3.1794	-7471B	PPB	13 May 2019 13:30:01	U	SMPL
K2823-01 - 1	10-B	12217	1.9883	-7471B	PPB	13 May 2019 13:32:17	U	SMPL
K2823-02 - 1	10-B2	115	-0.0178	-7471B	PPB	13 May 2019 13:34:33	U	SMPL
K2823-03 - 1	11-B	1143	0.1526	-7471B	PPB	13 May 2019 13:36:49	U	SMPL
K2823-04 - 1	11-B2	146	-0.0126	-7471B	PPB	13 May 2019 13:39:06	U	SMPL
K2823-05 - 1	12-B	1250	0.1704	-7471B	PPB	13 May 2019 13:41:23	U	SMPL
K2823-06 - 1	12-B2	82	-0.0233	-7471B	PPB	13 May 2019 13:43:39	U	SMPL
PB119684BL - 1	PBS	90	-0.0219	-7471B	PPB	13 May 2019 13:45:57	U	SMPL
PB119684BS - 1	LCSS	23641	3.8819	-7471B	PPB	13 May 2019 13:48:14	U	SMPL
K2767-01 - 1	P026-SS009-1824-01	3189	0.4918	-7471B	PPB	13 May 2019 13:50:30	U	SMPL
CCV97 - 1	CCV97	32221	5.3041	-7471B	PPB	13 May 2019 13:52:47	U	SMPL
CCB97 - 1	CCB97	-18	-0.0398	-7471B	PPB	13 May 2019 13:55:03	U	SMPL
K2767-02 - 1	P026-SS010-0002-01	4094	0.6418	-7471B	PPB	13 May 2019 13:57:20	U	SMPL
K2767-03 - 1	P026-SS010-0206-01	4881	0.7722	-7471B	PPB	13 May 2019 13:59:36	U	SMPL
K2767-04 - 1	P026-SS010-0612-01	4214	0.6617	-7471B	PPB	13 May 2019 14:01:52	U	SMPL
K2767-05 - 1	P026-SS010-1218-01	5940	0.9478	-7471B	PPB	13 May 2019 14:04:07	U	SMPL
K2767-06 - 1	P026-SS010-1824-01	3504	0.5440	-7471B	PPB	13 May 2019 14:06:23	U	SMPL
K2767-07 - 1	P026-SS011-0002-01	8841	1.4286	-7471B	PPB	13 May 2019 14:08:40	U	SMPL
K2767-08 - 1	P026-SS011-0206-01	6074	0.9700	-7471B	PPB	13 May 2019 14:10:56	U	SMPL
K2767-09 - 1	P026-SS011-0612-01	5193	0.8239	-7471B	PPB	13 May 2019 14:13:14	U	SMPL
K2767-10 - 1	P026-SS011-1218-01	7022	1.1271	-7471B	PPB	13 May 2019 14:15:30	U	SMPL
K2767-11 - 1	P026-SS011-1824-01	5309	0.8432	-7471B	PPB	13 May 2019 14:17:46	U	SMPL
CCV98 - 1	CCV98	32425	5.3379	-7471B	PPB	13 May 2019 14:20:02	U	SMPL
CCB98 - 1	CCB98	-32	-0.0422	-7471B	PPB	13 May 2019 14:22:18	U	SMPL
K2767-12 - 1	P026-SS012-0002-01	3614	0.5622	-7471B	PPB	13 May 2019 14:24:34	U	SMPL
K2767-12DUP - 1	P026-SS012-0002-01DUP	3474	0.5390	-7471B	PPB	13 May 2019 14:26:50	U	SMPL
K2767-13 - 1	K2767-12MS	27550	4.5299	-7471B	PPB	13 May 2019 14:29:06	U	SMPL
K2767-14 - 1	K2767-12MSD	27609	4.5396	-7471B	PPB	13 May 2019 14:31:21	U	SMPL
K2767-15 - 1	P026-SS012-0002-02	2911	0.4457	-7471B	PPB	13 May 2019 14:33:39	U	SMPL
K2767-16 - 1	P026-SS012-0206-01	3843	0.6002	-7471B	PPB	13 May 2019 14:35:57	U	SMPL
K2767-17 - 1	P026-SS012-0612-01	3102	0.4773	-7471B	PPB	13 May 2019 14:38:13	U	SMPL
K2767-18 - 1	P026-SS012-1218-01	2627	0.3986	-7471B	PPB	13 May 2019 14:40:30	U	SMPL
K2767-19 - 1	P026-SS012-1824-01	2076	0.3073	-7471B	PPB	13 May 2019 14:42:47	U	SMPL
K2767-20 - 1	P026-SS013-0002-01	3216	0.4962	-7471B	PPB	13 May 2019 14:45:04	U	SMPL
CCV99 - 1	CCV99	32465	5.3446	-7471B	PPB	13 May 2019 14:47:20	U	SMPL
CCB99 - 1	CCB99	-53	-0.0456	-7471B	PPB	13 May 2019 14:49:37	U	SMPL
K2767-21 - 1	P026-SS013-0206-01	4661	0.7358	-7471B	PPB	13 May 2019 14:51:54	U	SMPL
PB119685BL - 1	PBS	74	-0.0246	-7471B	PPB	13 May 2019 14:54:11	U	SMPL
PB119685BS - 1	LCSS	23625	3.8793	-7471B	PPB	13 May 2019 14:56:27	U	SMPL
K2768-01 - 1	P026-SS013-0612-01	6187	0.9887	-7471B	PPB	13 May 2019 14:58:43	U	SMPL
K2768-02 - 1	P026-SS013-1218-01	6267	1.0020	-7471B	PPB	13 May 2019 15:01:00	U	SMPL
K2768-03 - 1	P026-SS013-1824-01	3241	0.5004	-7471B	PPB	13 May 2019 15:03:16	U	SMPL
K2768-04 - 1	P026-SS014-0002-01	7044	1.1308	-7471B	PPB	13 May 2019 15:05:33	U	SMPL
K2768-05 - 1	P026-SS014-0206-01	5905	0.9420	-7471B	PPB	13 May 2019 15:07:50	U	SMPL
K2768-06 - 1	P026-SS014-0612-01	3623	0.5637	-7471B	PPB	13 May 2019 15:10:07	U	SMPL
K2768-07 - 1	P026-SS014-1218-01	4926	0.7797	-7471B	PPB	13 May 2019 15:12:24	U	SMPL
CCV01 - 1	CCV01	32035	5.2733	-7471B	PPB	13 May 2019 15:14:41	U	SMPL

LB102705 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 May 2019 12:03:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB01 - 1	CCB01	-5	-0.0377	-7471B	PPB		13 May 2019 15:16:58	U	SMPL
K2768-08 - 1	P026-SS014-1824-01	4529	0.7139	-7471B	PPB		13 May 2019 15:19:15	U	SMPL
K2768-09 - 1	P026-SS015-0002-01	4561	0.7192	-7471B	PPB		13 May 2019 15:21:32	U	SMPL
K2768-10 - 1	P026-SS015-0206-01	4460	0.7024	-7471B	PPB		13 May 2019 15:23:49	U	SMPL
K2768-11 - 1	P026-SS015-0612-01	3674	0.5722	-7471B	PPB		13 May 2019 15:26:05	U	SMPL
K2768-12 - 1	P026-SS015-1218-01	3532	0.5486	-7471B	PPB		13 May 2019 15:28:22	U	SMPL
K2768-13 - 1	P026-SS015-1824-01	2893	0.4427	-7471B	PPB		13 May 2019 15:30:38	U	SMPL
K2768-13DUP - 1	P026-SS015-1824-01DUP	2883	0.4410	-7471B	PPB		13 May 2019 15:32:54	U	SMPL
K2768-14 - 1	K2768-13MS	28059	4.6142	-7471B	PPB		13 May 2019 15:35:11	U	SMPL
K2768-15 - 1	K2768-13MSD	28323	4.6580	-7471B	PPB		13 May 2019 15:37:29	U	SMPL
K2768-16 - 1	P026-SS015-1824-02	2969	0.4553	-7471B	PPB		13 May 2019 15:39:46	U	SMPL
CCV02 - 1	CCV02	31719	5.2209	-7471B	PPB		13 May 2019 15:42:03	U	SMPL
CCB02 - 1	CCB02	11	-0.0350	-7471B	PPB		13 May 2019 15:44:20	U	SMPL
K2768-17 - 1	P026-SS016-0002-01	4548	0.7170	-7471B	PPB		13 May 2019 15:46:36	U	SMPL
K2768-18 - 1	P026-SS016-0206-01	5323	0.8455	-7471B	PPB		13 May 2019 15:48:53	U	SMPL
K2768-19 - 1	P026-SS016-0612-01	5194	0.8241	-7471B	PPB		13 May 2019 15:51:09	U	SMPL
K2768-20 - 1	P026-SS016-1218-01	4293	0.6748	-7471B	PPB		13 May 2019 15:53:25	U	SMPL
K2768-21 - 1	P026-SS016-1824-01	4099	0.6426	-7471B	PPB		13 May 2019 15:55:42	U	SMPL
K2801-01LX5 - 1		819	0.0989	-7471B	PPB		13 May 2019 15:57:58	U	SMPL
K2801-01A - 1		28700	4.7205	-7471B	PPB		13 May 2019 16:00:14	U	SMPL
K2767-12LX5 - 1		656	0.0719	-7471B	PPB		13 May 2019 16:02:31	U	SMPL
K2767-12A - 1		27767	4.5658	-7471B	PPB		13 May 2019 16:03:49	U	SMPL
K2768-13LX5 - 1		539	0.0525	-7471B	PPB		13 May 2019 16:06:06	U	SMPL
CCV03 - 1	CCV03	31671	5.2130	-7471B	PPB		13 May 2019 16:08:23	U	SMPL
CCB03 - 1	CCB03	-26	-0.0412	-7471B	PPB		13 May 2019 16:10:40	U	SMPL
K2768-13A - 1		27483	4.5188	-7471B	PPB		13 May 2019 16:12:58	U	SMPL
CCV04 - 1	CCV04	31490	5.1830	-7471B	PPB		13 May 2019 16:15:15	U	SMPL
CCB04 - 1	CCB04	-66	-0.0478	-7471B	PPB		13 May 2019 16:17:32	U	SMPL