



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102715

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/13/2019	16:05
2	CAL2	1	Water	NA	NA	20.1	7.00	05/13/2019	16:06
3	CAL3	1	Water	NA	NA	20.1	10.02	05/13/2019	16:07
4	ICV	1	Water	NA	NA	20.1	7.01	05/13/2019	16:10
5	CCV1	1	Water	NA	NA	20.1	2.02	05/13/2019	16:11
6	K2836-03	1	Solid	20.02	20	22.2	7.60	05/13/2019	16:12
7	K2837-02	1	Solid	20.04	20	22.4	9.23	05/13/2019	16:15
8	K2837-02DUP	1	Solid	20.05	20	22.5	9.24	05/13/2019	16:16
9	CCV2	1	Water	NA	NA	20.1	12.02	05/13/2019	16:20

LB 102715

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2837

WorkList ID : 126002

Date : 05-13-2019 13:48:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/10/2019	Solid	K2836-03	Corrosivity	Cool 4 deg C	DIST01		CULVER-CBTC-WC1-051019	05/10/2019	9045D
05/13/2019	Solid	K2837-02	Corrosivity	Cool 4 deg C	PSEG01	O21	SOIL-COMP	05/13/2019	9045D

Date/Time 05-13-19 14:00

Received by: Jo

Relinquished by: [Signature]

Date/Time 05-13-19 16:30

Received by: [Signature]

Relinquished by: JP