



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102734

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/14/2019	14:50
2	CAL2	1	Water	NA	NA	20.1	7.00	05/14/2019	14:51
3	CAL3	1	Water	NA	NA	20.1	10.02	05/14/2019	14:53
4	ICV	1	Water	NA	NA	20.1	6.99	05/14/2019	15:00
5	CCV1	1	Water	NA	NA	20.2	2.01	05/14/2019	15:01
6	K2852-02	1	Water	NA	NA	14.8	8.86	05/14/2019	15:02
7	K2852-02DUP	1	Water	NA	NA	14.9	8.87	05/14/2019	15:03
8	CCV2	1	Water	NA	NA	20.1	12.02	05/14/2019	15:05

173 102434

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2852

WorkList ID : 126042

Date : 05-14-2019 14:34:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/14/2019	Water	K2852-02 -17	Corrosivity	Cool 4 deg C	PSEG01	O33	TP-GROUNDWATER	05/14/2019	9040C

Date/Time 05-14-19 14:40
Received by: JD
Relinquished by: CL

Date/Time 05-14-19 15:15
Received by: CL
Relinquished by: JD