

LB102735


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53273



A= 0.0000e+000

B= 1.8832e-004

C= 1.3395e-002

Rho= 0.9998806

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.030	0.030	88	0.000	88					-
0.05	0.050										8
0.20	0.200	0.215	0.015	1073	0.0 %	1073					1
2.50	2.500	2.522	0.022	13320	0.0 %	13320					-3
5.00	5.000	4.875	-0.125	25815	0.0 %	25815					0
7.50	7.500	7.512	0.012	39818	0.0 %	39818					0
10.0	10.000	10.046	0.046	53273	0.0 %	53273					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	88	-		0 ISM02.4	PPB	5/14/2019 14:56	Std
	0.2 S0.2	1073	-		0.2 ISM02.4	PPB	5/14/2019 14:59	Std
	2.5 S2.5	13320	-		2.5 ISM02.4	PPB	5/14/2019 15:01	Std
	5 S5.0	25815	-		5 ISM02.4	PPB	5/14/2019 15:03	Std
	7.5 S7.5	39818	-		7.5 ISM02.4	PPB	5/14/2019 15:05	Std
	10 S10.0	53273	-		10 ISM02.4	PPB	5/14/2019 15:08	Std
ICV	ICV	21665	4.0934	-	ISM02.4	PPB	5/14/2019 15:13	SMPL
ICB	ICB	9	0.0151	-	ISM02.4	PPB	5/14/2019 15:15	SMPL
CCV017	CCV017	25684	4.8502	-	ISM02.4	PPB	5/14/2019 15:17	SMPL
CCB017	CCB017	7	0.0147	-	ISM02.4	PPB	5/14/2019 15:19	SMPL
PB119708BL	PBS708	42	0.0213	-	ISM02.4	PPB	5/14/2019 15:22	SMPL
K2720-01	MH0AQ3	2130	0.4145	-	ISM02.4	PPB	5/14/2019 15:24	SMPL
K2720-02	MH0AQ4	2503	0.4848	-	ISM02.4	PPB	5/14/2019 15:26	SMPL
K2720-03	MH0AQ5	2273	0.4414	-	ISM02.4	PPB	5/14/2019 15:28	SMPL
K2720-05	MH0AQ9	2897	0.559	-	ISM02.4	PPB	5/14/2019 15:31	SMPL
K2720-06	MH0AR0	2982	0.575	-	ISM02.4	PPB	5/14/2019 15:33	SMPL
K2720-07	MH0AR1	3165	0.6094	-	ISM02.4	PPB	5/14/2019 15:35	SMPL
K2720-08	MH0AR2	3258	0.6269	-	ISM02.4	PPB	5/14/2019 15:37	SMPL
K2720-09	MH0AR3	2138	0.416	-	ISM02.4	PPB	5/14/2019 15:40	SMPL
K2720-10	MH0AR4	2384	0.4624	-	ISM02.4	PPB	5/14/2019 15:42	SMPL
K2720-11	MH0AR5	2360	0.4578	-	ISM02.4	PPB	5/14/2019 15:44	SMPL
K2720-12	MH0AR6	3148	0.6062	-	ISM02.4	PPB	5/14/2019 15:47	SMPL
K2720-13	MH0AR7	2880	0.5558	-	ISM02.4	PPB	5/14/2019 15:49	SMPL
K2720-14	MH0AR8	7517	1.429	-	ISM02.4	PPB	5/14/2019 15:51	SMPL
K2720-15	MH0AR9	12311	2.3318	-	ISM02.4	PPB	5/14/2019 15:53	SMPL
K2720-16	MH0AS0	17467	3.3028	-	ISM02.4	PPB	5/14/2019 15:56	SMPL
K2720-17	MH0AS1	48447	9.137	-	ISM02.4	PPB	5/14/2019 15:58	SMPL
K2720-18	MH0AS2	38148	7.1975	-	ISM02.4	PPB	5/14/2019 16:00	SMPL
K2720-19	MH0AS3	39023	7.3623	-	ISM02.4	PPB	5/14/2019 16:02	SMPL
K2720-20	MH0AS4	20206	3.8186	-	ISM02.4	PPB	5/14/2019 16:05	SMPL
CCV018	CCV018	25782	4.8687	-	ISM02.4	PPB	5/14/2019 16:07	SMPL
CCB018	CCB018	-9	0.0117	-	ISM02.4	PPB	5/14/2019 16:09	SMPL
K2720-21	MH0AS4D	20494	3.8729	-	ISM02.4	PPB	5/14/2019 16:12	SMPL
K2720-22	MH0AS4S	31894	6.0197	-	ISM02.4	PPB	5/14/2019 16:19	SMPL
PB119709BL	PBS709	84	0.0292	-	ISM02.4	PPB	5/14/2019 16:21	SMPL
K2722-01	MH0AS5	35955	6.7845	-	ISM02.4	PPB	5/14/2019 16:24	SMPL
K2722-02	MH0AS5D	35424	6.6845	-	ISM02.4	PPB	5/14/2019 16:26	SMPL
K2722-03	MH0AS5S	46650	8.7986	-	ISM02.4	PPB	5/14/2019 16:45	SMPL
K2722-04	MH0AS6	35223	6.6466	-	ISM02.4	PPB	5/14/2019 16:47	SMPL
K2722-05	MH0AS7	22709	4.29	-	ISM02.4	PPB	5/14/2019 16:49	SMPL
K2722-06	MH0AS8	27031	5.1039	-	ISM02.4	PPB	5/14/2019 16:51	SMPL
K2722-07	MH0AS9	17386	3.2876	-	ISM02.4	PPB	5/14/2019 16:54	SMPL
K2722-08	MH0AT0	19807	3.7435	-	ISM02.4	PPB	5/14/2019 16:56	SMPL
K2722-09	MH0AT1	20166	3.8111	-	ISM02.4	PPB	5/14/2019 16:58	SMPL
K2722-10	MH0AT2	14132	2.6748	-	ISM02.4	PPB	5/14/2019 17:01	SMPL
PB119710BL	PBS710	26	0.0183	-	ISM02.4	PPB	5/14/2019 17:03	SMPL

INSTRUMENT ID : CV1

K2726-01	MHOB17	1681	0.33 -	ISM02.4	PPB	5/14/2019 17:05 SMPL
K2726-02	MHOB18	2114	0.4115 -	ISM02.4	PPB	5/14/2019 17:07 SMPL
K2726-03	MHOB19	3754	0.7204 -	ISM02.4	PPB	5/14/2019 17:10 SMPL
K2726-04	MHOB20	2694	0.5207 -	ISM02.4	PPB	5/14/2019 17:12 SMPL
K2726-05	MHOB21	3019	0.5819 -	ISM02.4	PPB	5/14/2019 17:14 SMPL
K2726-06	MHOB22	3514	0.6752 -	ISM02.4	PPB	5/14/2019 17:16 SMPL
CCV019	CCV019	25587	4.832 -	ISM02.4	PPB	5/14/2019 17:19 SMPL
CCB019	CCB019	-26	0.0085 -	ISM02.4	PPB	5/14/2019 17:21 SMPL
K2726-07	MHOB23	3946	0.7565 -	ISM02.4	PPB	5/14/2019 17:23 SMPL
K2726-08	MHOB24	1686	0.3309 -	ISM02.4	PPB	5/14/2019 17:26 SMPL
K2726-09	MHOB25	4346	0.8318 -	ISM02.4	PPB	5/14/2019 17:28 SMPL
K2726-10	MHOB26	9923	1.8821 -	ISM02.4	PPB	5/14/2019 17:30 SMPL
K2726-11	MHOB27	8608	1.6345 -	ISM02.4	PPB	5/14/2019 17:32 SMPL
K2726-12	MHOB28	8660	1.6443 -	ISM02.4	PPB	5/14/2019 17:35 SMPL
K2726-13	MHOB29	27989	5.2843 -	ISM02.4	PPB	5/14/2019 17:37 SMPL
K2726-14	MHOB30	76033	14.332 -	ISM02.4	PPB	5/14/2019 17:39 SMPL
K2726-15	MHOB31	21780	4.115 -	ISM02.4	PPB	5/14/2019 17:42 SMPL
K2726-16	MHOB32	33818	6.382 -	ISM02.4	PPB	5/14/2019 17:44 SMPL
K2726-17	MHOB33	16687	3.1559 -	ISM02.4	PPB	5/14/2019 17:46 SMPL
K2726-18	MHOB34	18153	3.432 -	ISM02.4	PPB	5/14/2019 17:48 SMPL
K2726-19	MHOB35	23441	4.4278 -	ISM02.4	PPB	5/14/2019 17:51 SMPL
K2726-20	MHOB36	18939	3.58 -	ISM02.4	PPB	5/14/2019 17:53 SMPL
K2726-21	MHOB36D	19252	3.639 -	ISM02.4	PPB	5/14/2019 17:55 SMPL
K2726-22	MHOB36S	34422	6.4958 -	ISM02.4	PPB	5/14/2019 18:06 SMPL
K2726-14DLX2		36319	6.853 -	ISM02.4	PPB	5/14/2019 18:08 SMPL
CCV020	CCV020	25975	4.905 -	ISM02.4	PPB	5/14/2019 18:10 SMPL
CCB020	CCB020	-9	0.0117 -	ISM02.4	PPB	5/14/2019 18:13 SMPL