



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102756

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/15/2019	11:20
2	CAL2	1	Water	NA	NA	20.2	7.01	05/15/2019	11:21
3	CAL3	1	Water	NA	NA	20.2	10.03	05/15/2019	11:22
4	ICV	1	Water	NA	NA	20.1	6.98	05/15/2019	11:25
5	CCV1	1	Water	NA	NA	20.2	2.02	05/15/2019	11:26
6	K2858-01	1	Water	NA	NA	15.5	5.63	05/15/2019	11:27
7	K2861-02	1	Water	NA	NA	14.3	8.41	05/15/2019	11:30
8	K2861-02DUP	1	Water	NA	NA	14.4	8.42	05/15/2019	11:31
9	CCV2	1	Water	NA	NA	20.1	12.02	05/15/2019	11:33

28102756

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2858

WorkList ID : 126084

Date : 05-15-2019 11:11:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/14/2019	Water	K2858-01-17	pH	Cool 4 deg C	STER12	O11	A-CELL-POOL-WATER-2ND-C	05/14/2019	9040C
05/14/2019	Water	K2861-02-17	pH	Cool 4 deg C	TETR06	O31	TT-001-IDWGW-20190514	05/14/2019	9040C

Date/Time 05-15-19 11:15
 Received by: fo
 Relinquished by: ca

Date/Time 05-15-19 12:10
 Received by: cm
 Relinquished by: fo