



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102762

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/15/2019	14:00
2	CAL2	1	Water	NA	NA	20.1	7.01	05/15/2019	14:01
3	CAL3	1	Water	NA	NA	20.1	10.03	05/15/2019	14:02
4	ICV	1	Water	NA	NA	20.1	7.00	05/15/2019	14:03
5	CCV1	1	Water	NA	NA	20.1	2.02	05/15/2019	14:05
6	K2651-01	1	Water	NA	NA	23.0	7.40	05/15/2019	14:06
7	CCV2	1	Water	NA	NA	20.1	12.02	05/15/2019	14:09

LB 102762

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2551

WorkList ID : 126091

Date : 05-15-2019 13:50:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2019	Water	K2651-01	pH	NONE	CHEM02	QA Office	55061-042019	04/30/2019	SM 4500-H B

Date/Time 05-15-19 14:00
Received by: JO
Relinquished by: [Signature]

Date/Time 05-15-19 14:35
Received by: [Signature]
Relinquished by: [Signature]