



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102782

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/16/2019	12:05
2	CAL2	1	Water	NA	NA	20.3	7.00	05/16/2019	12:06
3	CAL3	1	Water	NA	NA	20.3	10.03	05/16/2019	12:07
4	ICV	1	Water	NA	NA	20.2	6.99	05/16/2019	12:08
5	CCV1	1	Water	NA	NA	20.3	2.02	05/16/2019	12:10
6	K2881-01	1	Solid	20.03	20	22.5	5.56	05/16/2019	12:11
7	K2881-01DUP	1	Solid	20.04	20	22.6	5.57	05/16/2019	12:13
8	CCV2	1	Water	NA	NA	20.3	12.02	05/16/2019	12:15

LB 102782

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2881

WorkList ID : 126145

Date : 05-16-2019 11:52:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/15/2019	Solid	K2881-01-17	pH	Cool 4 deg C	TULL02	O41	FK-PS-01-047	05/15/2019	9045D

Date/Time 05-16-19 12:00
 Received by: JH
 Relinquished by: CS

Date/Time 05-16-19 12:40
 Received by: CS
 Relinquished by: JP