



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102787

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP74764
Residual chlorine std, 0PPM	WP74760
Residual chlorine std, 0.1PPM	WP74761
Residual chlorine std, 0.2PPM	WP74762
Residual chlorine std, and CCV , 0.4PPM	WP74763
Residual chlorine std, 1.6PPM	WP74765
Residual chlorine ICV-LCS, 0.4PPM	WP74766
POTASSIUM BROMATE 0.1N 1L	W2449

Intercept: -0.0006

Slope: 1.0076

Regression: 0.999992

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	05/16/2019	10:32
2	CAL2	0.1	1	0.000	0.100	0.100	05/16/2019	10:32
3	CAL3	0.2	1	0.000	0.200	0.200	05/16/2019	10:33
4	CAL4	0.4	1	0.000	0.400	0.400	05/16/2019	10:34
5	CAL5	0.8	1	0.000	0.810	0.810	05/16/2019	10:34
6	CAL6	1.6	1	0.000	1.610	1.610	05/16/2019	10:35



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102787

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4000	0.4000	0.3980	05/16/2019	10:36
2	ICB				1	0.0000	0.0000	0.0000	0.0010	05/16/2019	10:36
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4080	05/16/2019	10:37
4	CCB1				1	0.0000	0.0000	0.0000	0.0010	05/16/2019	10:38
5	LB102787BL	50	50		1	0.0000	0.0000	0.0000	0.0010	05/16/2019	10:39
6	LB102787BS	50	50	0.4	1	0.0000	0.3900	0.3900	0.3880	05/16/2019	10:40
7	K2886-01	50	50		1	0.0000	0.0000	0.0000	0.0010	05/16/2019	10:41
8	K2886-01DUP	50	50		1	0.0000	0.0000	0.0000	0.0010	05/16/2019	10:42
9	K2886-01MS	50	50	0.4	1	0.0000	0.3800	0.3800	0.3780	05/16/2019	10:43
10	K2886-01MSD	50	50	0.4	1	0.0000	0.3900	0.3900	0.3880	05/16/2019	10:44
11	CCV2			0.4	1	0.0000	0.4000	0.4000	0.3980	05/16/2019	10:45
12	CCB2				1	0.0000	0.0000	0.0000	0.0010	05/16/2019	10:46