



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102797

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/16/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	05/16/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.02	05/16/2019	17:17
4	ICV	1	Water	NA	NA	20.1	6.99	05/16/2019	17:20
5	CCV1	1	Water	NA	NA	20.3	2.01	05/16/2019	17:30
6	K2643-06	1	Solid	20.02	20	23.4	8.16	05/16/2019	17:31
7	K2643-06DUP	1	Solid	20.03	20	23.5	8.17	05/16/2019	17:32
8	K2643-08	1	Solid	20.04	20	23.6	7.99	05/16/2019	17:33
9	K2869-12	1	Solid	20.05	20	23.0	8.80	05/16/2019	17:35
10	K2880-03	1	Solid	20.04	20	22.8	9.50	05/16/2019	17:36
11	K2896-04	1	Solid	20.02	20	23.0	10.13	05/16/2019	17:39
12	K2896-08	1	Solid	20.01	20	21.9	8.99	05/16/2019	17:41
13	K2899-01	1	Solid	20.02	20	23.1	7.06	05/16/2019	17:43
14	K2900-03	1	Solid	20.04	20	23.2	8.60	05/16/2019	17:45
15	CCV2	1	Water	NA	NA	20.2	12.02	05/16/2019	17:46

WORKLIST(Hardcopy Internal Chain)

18102797

WorkList Name : CORROSIONITY K2899

WorkList ID : 126157

Date : 05-16-2019 14:54:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2019	Solid	K2643-06	Corrosivity	Cool 4 deg C	PSEG05		OR-02-051619-A	05/16/2019	9045D
05/16/2019	Solid	K2643-08	Corrosivity	Cool 4 deg C	PSEG05		OR-02-051619-B	05/16/2019	9045D
05/16/2019	Solid	K2869-12	Corrosivity	Cool 4 deg C	PSEG01	O21	ROCKY-SOIL	05/16/2019	9045D
05/16/2019	Solid	K2880-03	Corrosivity	Cool 4 deg C	PSEG01	O31	339-RBR-251608	05/16/2019	9045D
05/15/2019	Solid	K2896-04	Corrosivity	Cool 4 deg C	PSEG01	O11	TP-18-S	05/15/2019	9045D
05/15/2019	Solid	K2896-08	Corrosivity	Cool 4 deg C	PSEG01	O11	TP-20-S	05/15/2019	9045D
05/14/2019	Solid	K2899-01	Corrosivity	Cool 4 deg C	CLEA11	O63	OIL-WATER-SEPERATOR(OV	05/14/2019	9045D
05/16/2019	Solid	K2900-03	Corrosivity	Cool 4 deg C	WSP001	O43	P-2(0-10)	05/16/2019	9045D

Date/Time 05.16.19 16:20
 Received by: JD
 Relinquished by: _____

Date/Time 05.16.19 18:25
 Received by: _____
 Relinquished by: JP