



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102798

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/16/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	05/16/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.02	05/16/2019	17:17
4	ICV	1	Water	NA	NA	20.1	6.99	05/16/2019	17:19
5	CCV1	1	Water	NA	NA	20.1	2.02	05/16/2019	17:20
6	K2880-05	1	Water	NA	NA	18.8	7.65	05/16/2019	17:21
7	K2896-10	1	Water	NA	NA	13.2	7.73	05/16/2019	17:24
8	K2896-10DUP	1	Water	NA	NA	13.3	7.74	05/16/2019	17:25
9	CCV2	1	Water	NA	NA	20.1	12.02	05/16/2019	17:26

2B 102798

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2896

WorkList ID : 126174

Date : 05-16-2019 16:49:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2019	Water	K2880-05	-A Corrosivity	Cool 4 deg C	PSEG01	O31	339	05/16/2019	9040C
05/15/2019	Water	K2896-10	-X Corrosivity	Cool 4 deg C	PSEG01	O11	TP-20	05/15/2019	9040C

Date/Time 05.16.19 16:20
 Received by: SD
 Relinquished by: SD

Date/Time 05.16.19 18:20
 Received by: SD
 Relinquished by: SD