

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB102816

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP74163
calibration std. phosphate 1 ppm	WP74778
calibration std. phosphate 0.5 ppm	WP74777
calibration std. phosphate 0.3 ppm	WP74776
calibration std. 0 ppm	WP74773
phosphate ICV std.	WP74780
phosphate CCV std.	WP74779
calibration std. phosphate 0.1 ppm	WP74775
calibration std. phosphate 0.05 ppm	WP74774
Combined reagent	WP74783
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: -0.0007Slope: 0.6469Regression: 0.999874

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/17/2019	13:49
2	CAL2	0.05	1	50	50	0.033	05/17/2019	13:49
3	CAL3	0.10	1	50	50	0.065	05/17/2019	13:50
4	CAL4	0.30	1	50	50	0.186	05/17/2019	13:50
5	CAL5	0.50	1	50	50	0.327	05/17/2019	13:51
6	CAL6	1.00	1	50	50	0.646	05/17/2019	13:51



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB102816

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.499	05/17/2019	13:52
2	ICB		1	50	50	0.000	0.001	05/17/2019	13:52
3	CCV1	0.50	1	50	50	0.326	0.505	05/17/2019	13:53
4	CCB1		1	50	50	0.000	0.001	05/17/2019	13:53
5	PB119826BL		1	50	50	0.000	0.001	05/17/2019	13:54
6	PB119826BS	0.50	1	50	50	0.330	0.511	05/17/2019	13:55
7	K2844-03		1	50	50	0.118	0.183	05/17/2019	13:56
8	K2844-07		1	50	50	0.049	0.077	05/17/2019	13:58
9	K2844-09		1	50	50	0.123	0.191	05/17/2019	13:58
10	K2844-10		1	50	50	0.060	0.094	05/17/2019	13:59
11	K2844-10DUP		1	50	50	0.053	0.083	05/17/2019	13:59
12	K2844-10MS	0.50	1	50	50	0.402	0.623	05/17/2019	14:00
13	K2844-10MSD	0.50	1	50	50	0.400	0.619	05/17/2019	14:00
14	K2864-01		1	50	50	0.018	0.029	05/17/2019	14:02
15	CCV2	0.50	1	50	50	0.328	0.508	05/17/2019	14:03
16	CCB2		1	50	50	0.000	0.001	05/17/2019	14:03
17	K2864-02		1	50	50	0.019	0.030	05/17/2019	14:04
18	K2886-01		1	50	50	0.012	0.020	05/17/2019	14:04
19	K2919-01		1	50	50	3.072	4.750	05/17/2019	14:03
20	K2919-01		100	50	50	0.252	0.391	05/17/2019	14:20
21	CCV3	0.50	1	50	50	0.321	0.497	05/17/2019	14:21
22	CCB3		1	50	50	0.000	0.001	05/17/2019	14:21