

LB102824


ISM02.4

Instrument ID: CVI

Linear

 μ Abs.:

44856



A= 0.0000e+000

B= 2.2158e-004

C= -2.1826e-002

Rho= 0.9997000

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.085	0.085	483	0.000	483					—
0.05	0.050										—
0.20	0.200	0.166	-0.034	846	0.0 %	846					-17.
2.50	2.500	2.404	-0.096	10947	0.0 %	10947					-3.8
5.00	5.000	4.974	-0.026	22547	0.0 %	22547					-0.52.
7.50	7.500	7.654	0.154	34643	0.0 %	34643					2
10.0	10.000	9.917	-0.083	44856	0.0 %	44856					-0.8

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	483	-		0 ISM02.4	PPB	5/17/2019 14:33	Std
	0.2 S0.2	846	-		0.2 ISM02.4	PPB	5/17/2019 14:35	Std
	2.5 S2.5	10947	-		2.5 ISM02.4	PPB	5/17/2019 14:37	Std
	5 S5.0	22547	-		5 ISM02.4	PPB	5/17/2019 14:40	Std
	7.5 S7.5	34643	-		7.5 ISM02.4	PPB	5/17/2019 14:42	Std
	10 S10.0	44856	-		10 ISM02.4	PPB	5/17/2019 14:44	Std
ICV	ICV	17859	3.9353	-	ISM02.4	PPB	5/17/2019 14:51	SMPL
ICB	ICB	407	0.0684	-	ISM02.4	PPB	5/17/2019 14:53	SMPL
CCV037	CCV037	22564	4.9778	-	ISM02.4	PPB	5/17/2019 14:55	SMPL
CCB037	CCB037	396	0.0659	-	ISM02.4	PPB	5/17/2019 14:57	SMPL
PB119846BL	PBW846	431	0.0737	-	ISM02.4	PPB	5/17/2019 15:00	SMPL
K2890-01	MC0B58	477	0.0839	-	ISM02.4	PPB	5/17/2019 15:02	SMPL
K2890-02	MC0B58D	457	0.0794	-	ISM02.4	PPB	5/17/2019 15:04	SMPL
K2890-04	MC0B59	302	0.0451	-	ISM02.4	PPB	5/17/2019 15:09	SMPL
K2890-05	MC0B60	423	0.0719	-	ISM02.4	PPB	5/17/2019 15:11	SMPL
K2890-06	MC0B61	397	0.0661	-	ISM02.4	PPB	5/17/2019 15:13	SMPL
K2890-07	MC0B62	417	0.0706	-	ISM02.4	PPB	5/17/2019 15:15	SMPL
K2890-08	MC0B63	427	0.0728	-	ISM02.4	PPB	5/17/2019 15:18	SMPL
K2890-03	MC0B58S	5335	1.1603	-	ISM02.4	PPB	5/17/2019 15:32	SMPL
CCV038	CCV038	22697	5.0073	-	ISM02.4	PPB	5/17/2019 15:34	SMPL
CCB038	CCB038	338	0.0531	-	ISM02.4	PPB	5/17/2019 15:36	SMPL