



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102831

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/17/2019	13:05
2	K2880-05		1	8	100.00	>212.0	>212.0	05/17/2019	13:35
3	K2880-05DUP		1	8	100.00	>212.0	>212.0	05/17/2019	14:05
4	K2896-10		1	9	100.00	>212.0	>212.0	05/17/2019	14:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 102831

WorkList Name : FP51719

WorkList ID : 126216

Date : 05-17-2019 09:09:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/21/2019	Water	K2880-05	Flash Point	Cool 4 deg C	PSEG01	O31	339	05/16/2019	1010A
05/22/2019	Water	K2896-10	Flash Point	Cool 4 deg C	PSEG01	O11	TP-20	05/15/2019	1010A

Date/Time 05/17/19 12:30
 Received by: KS
 Relinquished by: CP

Date/Time 05/17/19 14:40
 Received by: CP
 Relinquished by: KS