



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 5/21/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 15:00  
In Date: 05/20/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102  
Time OUT: 07:35  
Out Date: 05/21/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1  
Thermometer ID: %SOLIDS-OVEN

QC:LB102857

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K2846-01 | BFHB6            | 1     | 1.13              | 9.88                          | 3.43                              | 26.3    |
| K2846-02 | BFHB7            | 2     | 1.16              | 9.95                          | 3.74                              | 29.4    |
| K2846-03 | VHBLK01          | 3     | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K2846-04 | BFHB2            | 4     | 1.14              | 9.93                          | 6.54                              | 61.4    |
| K2846-05 | BFHB4            | 5     | 1.15              | 9.81                          | 5.13                              | 46.0    |
| K2846-06 | BFHB5            | 6     | 1.13              | 9.73                          | 7.00                              | 68.3    |
| K2846-07 | BFHC2            | 7     | 1.17              | 9.59                          | 6.45                              | 62.7    |
| K2846-08 | BFHC3            | 8     | 1.18              | 9.64                          | 6.64                              | 64.5    |
| K2846-09 | BFH90            | 9     | 1.16              | 9.83                          | 6.53                              | 61.9    |
| K2846-10 | BFH91            | 10    | 1.15              | 9.95                          | 7.02                              | 66.7    |
| K2846-11 | BFHB0            | 11    | 1.17              | 9.85                          | 4.12                              | 34.0    |
| K2846-12 | BFHB1            | 12    | 1.18              | 9.91                          | 5.44                              | 48.8    |
| K2846-13 | BFHB1MS          | 13    | 1.18              | 9.91                          | 5.44                              | 48.8    |
| K2846-14 | BFHB1MSD         | 14    | 1.18              | 9.91                          | 5.44                              | 48.8    |
| K2846-15 | BFHB8            | 15    | 1.14              | 9.87                          | 5.86                              | 54.1    |
| K2846-16 | BFHB9            | 16    | 1.19              | 9.69                          | 5.7                               | 53.1    |
| K2846-17 | BFHC8            | 17    | 1.19              | 9.97                          | 4.37                              | 36.2    |
| K2846-18 | BFHA0            | 18    | 1.18              | 9.98                          | 4.83                              | 41.5    |
| K2846-19 | BFHA1            | 19    | 1.1               | 9.68                          | 4.69                              | 41.8    |
| K2846-20 | BFHA2            | 20    | 1.16              | 9.78                          | 5.38                              | 49.0    |
| K2846-21 | BFHA3            | 21    | 1.18              | 9.79                          | 5.93                              | 55.2    |
| K2846-22 | BFHA4            | 22    | 1.13              | 9.89                          | 6.97                              | 66.7    |
| K2846-23 | BFHA5            | 23    | 1.18              | 9.96                          | 7.96                              | 77.2    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$