



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102869

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/20/2019	13:10
2	K2913-06		1	8	100.00	>212.0	>212.0	05/20/2019	13:35
3	K2913-06DUP		1	8	100.00	>212.0	>212.0	05/20/2019	14:05
4	K2935-01		1	9	100.00	>212.0	>212.0	05/20/2019	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102869

WorkList Name : FP52019

WorkList ID : 126293

Date : 05-20-2019 10:04:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/23/2019	Water	K2913-06	Flash Point	Cool 4 deg C	PSEG01	K11	TP-17	05/16/2019	1010A
05/31/2019	Water	K2935-01	Flash Point	Cool 4 deg C	SEAC01		PURGE-WATER-MUD	05/17/2019	1010A

Date/Time 05/20/19 12:45
 Received by: KS
 Relinquished by: CP

Date/Time 05/20/19 14:45
 Received by: CP
 Relinquished by: KS