

LB102870


7470A

INSTRUMENT ID : CU1

Linear

 μ Abs. :

43647



A= 0.0000e+000

B= 2.2387e-004

C= -2.6142e-002

Rho= 0.9989143

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	88	0.000	88				
0.2	0.200	0.189	-0.011	959	0.0 %	959				
2.5	2.500	2.420	-0.080	10928	0.0 %	10928				
5.0	5.000	5.028	0.028	22577	0.0 %	22577				
7.5	7.500	7.824	0.324	35066	0.0 %	35066				
10.0	10.000	9.745	-0.255	43647	0.0 %	43647				

LB102870 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 May 2019 15:50:22

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50-0	88	-	0.00007470A	PPB	20 May 2019 15:51:17	S	Std
0.2 - 1	50-2	959	-	0.20007470A	PPB	20 May 2019 15:53:25	S	Std
2.5 - 1	52-5	10928	-	2.50007470A	PPB	20 May 2019 15:55:34	S	Std
5.0 - 1	55-0	22577	-	5.00007470A	PPB	20 May 2019 15:57:42	S	Std
7.5 - 1	57-5	35066	-	7.50007470A	PPB	20 May 2019 15:59:50	S	Std
10.0 - 1	510-0	43647	-	10.00007470A	PPB	20 May 2019 16:01:57	S	Std
ICV52 - 1	ICV52	17439	3.8780	-7470A	PPB	20 May 2019 16:05:18	U	SMPL
ICB52 - 1	ICB52	7	-0.0246	-7470A	PPB	20 May 2019 16:07:25	U	SMPL
CCV55 - 1	CCV55	22932	5.1077	-7470A	PPB	20 May 2019 16:09:31	U	SMPL
CCB55 - 1	CCB55	-2	-0.0266	-7470A	PPB	20 May 2019 16:11:37	U	SMPL
CRA - 1	CRA	989	0.1953	-7470A	PPB	20 May 2019 16:13:44	U	SMPL
HighStd - 1	HighStd	43804	9.7803	-7470A	PPB	20 May 2019 16:15:51	U	SMPL
ChkStd - 1	ChkStd	32170	7.1758	-7470A	PPB	20 May 2019 16:17:58	U	SMPL
PB119908BL - 1	PBS	-68	-0.0414	-7470A	PPB	20 May 2019 16:20:06	U	SMPL
PB119908BS - 1	LCS	15985	3.5525	-7470A	PPB	20 May 2019 16:22:13	U	SMPL
K2643-06 - 1	OR-02-051619-A	-11	-0.0286	-7470A	PPB	20 May 2019 16:24:20	U	SMPL
K2643-06DUP - 1	OR-02-051619-ADUP	15	-0.0228	-7470A	PPB	20 May 2019 16:26:28	U	SMPL
K2643-06MS - 1	OR-02-051619-AMS	18897	4.2044	-7470A	PPB	20 May 2019 16:28:35	U	SMPL
K2643-06MSD - 1	OR-02-051619-AMSD	18924	4.2104	-7470A	PPB	20 May 2019 16:30:43	U	SMPL
K2643-08 - 1	OR-02-051619-B	-21	-0.0308	-7470A	PPB	20 May 2019 16:32:50	U	SMPL
CCV56 - 1	CCV56	22652	5.0450	-7470A	PPB	20 May 2019 16:34:57	U	SMPL
CCB56 - 1	CCB56	-8	-0.0279	-7470A	PPB	20 May 2019 16:37:03	U	SMPL
K2648-02 - 1	TR-05-051619	27	-0.0201	-7470A	PPB	20 May 2019 16:39:11	U	SMPL
K2869-12 - 1	ROCKY-SOIL	19	-0.0219	-7470A	PPB	20 May 2019 16:41:18	U	SMPL
K2880-03 - 1	339-RBR-251608	39	-0.0174	-7470A	PPB	20 May 2019 16:43:24	U	SMPL
K2896-04 - 1	TP-18-S	15	-0.0228	-7470A	PPB	20 May 2019 16:45:31	U	SMPL
K2896-08 - 1	TP-20-S	19	-0.0219	-7470A	PPB	20 May 2019 16:47:37	U	SMPL
K2899-01 - 1	OIL-WATER-SEPERATOR(OWS)-4	68	-0.0109	-7470A	PPB	20 May 2019 16:49:44	U	SMPL
K2900-03 - 1	P-2(0-10)	48	-0.0154	-7470A	PPB	20 May 2019 16:51:52	U	SMPL
PB119909BL - 1	PBW	25	-0.0205	-7470A	PPB	20 May 2019 16:54:00	U	SMPL
PB119909BS - 1	LCSW	16126	3.5840	-7470A	PPB	20 May 2019 16:56:08	U	SMPL
K2838-07 - 1	FEQC-FB-051419	93	-0.0053	-7470A	PPB	20 May 2019 16:58:15	U	SMPL
CCV57 - 1	CCV57	22776	5.0728	-7470A	PPB	20 May 2019 17:00:22	U	SMPL
CCB57 - 1	CCB57	-5	-0.0273	-7470A	PPB	20 May 2019 17:02:28	U	SMPL
K2852-01 - 1	TP-GROUNDWATER	4900	1.0708	-7470A	PPB	20 May 2019 17:04:35	U	SMPL
K2861-02 - 1	TT-001-IDWGW-20190514	647	0.1187	-7470A	PPB	20 May 2019 17:06:42	U	SMPL
K2880-04 - 1	339	1208	0.2443	-7470A	PPB	20 May 2019 17:08:48	U	SMPL
K2896-09 - 1	TP-20	320	0.0455	-7470A	PPB	20 May 2019 17:10:55	U	SMPL
K2906-05 - 1	FEQC-FB-051619	31	-0.0192	-7470A	PPB	20 May 2019 17:13:02	U	SMPL
K2909-01 - 1	20190350-FIELD-BLANK	133	0.0036	-7470A	PPB	20 May 2019 17:15:09	U	SMPL
K2913-05 - 1	TP-17	390	0.0612	-7470A	PPB	20 May 2019 17:17:17	U	SMPL
K2913-05DUP - 1	TP-17DUP	403	0.0641	-7470A	PPB	20 May 2019 17:19:25	U	SMPL
K2913-05MS - 1	TP-17MS	19560	4.3528	-7470A	PPB	20 May 2019 17:21:33	U	SMPL
K2913-05MSD - 1	TP-17MSD	19737	4.3924	-7470A	PPB	20 May 2019 17:23:40	U	SMPL
CCV58 - 1	CCV58	22123	4.9266	-7470A	PPB	20 May 2019 17:25:46	U	SMPL
CCB58 - 1	CCB58	-5	-0.0273	-7470A	PPB	20 May 2019 17:27:53	U	SMPL
PB119910BL - 1	PBW	25	-0.0205	-7470A	PPB	20 May 2019 17:30:00	U	SMPL
PB119910BS - 1	LCSW	16080	3.5737	-7470A	PPB	20 May 2019 17:32:06	U	SMPL
K2896-10 - 1	TP-20	98	-0.0042	-7470A	PPB	20 May 2019 17:34:13	U	SMPL
K2913-06 - 1	TP-17	139	0.0050	-7470A	PPB	20 May 2019 17:36:20	U	SMPL
K2913-06DUP - 1	TP-17DUP	156	0.0088	-7470A	PPB	20 May 2019 17:38:27	U	SMPL
K2913-06MS - 1	TP-17MS	17880	3.9767	-7470A	PPB	20 May 2019 17:40:35	U	SMPL
K2913-06MSD - 1	TP-17MSD	17732	3.9436	-7470A	PPB	20 May 2019 17:42:43	U	SMPL
PB119911BL - 1	PBW	-35	-0.0340	-7470A	PPB	20 May 2019 17:44:51	U	SMPL
PB119911BS - 1	LCSW	16101	3.5784	-7470A	PPB	20 May 2019 17:46:59	U	SMPL
K2641-12 - 1	CL-03-051719	116	-0.0002	-7470A	PPB	20 May 2019 17:49:06	U	SMPL
CCV59 - 1	CCV59	22597	5.0327	-7470A	PPB	20 May 2019 17:51:12	U	SMPL
CCB59 - 1	CCB59	-7	-0.0277	-7470A	PPB	20 May 2019 17:53:19	U	SMPL
K2641-12DUP - 1	CL-03-051719DUP	161	0.0099	-7470A	PPB	20 May 2019 17:55:26	U	SMPL
K2641-12MS - 1	CL-03-051719MS	18160	4.0394	-7470A	PPB	20 May 2019 17:57:32	U	SMPL
K2641-12MSD - 1	CL-03-051719MSD	18123	4.0311	-7470A	PPB	20 May 2019 17:59:39	U	SMPL
K2647-10 - 1	SU-01-051719	77	-0.0089	-7470A	PPB	20 May 2019 18:01:46	U	SMPL
K2913-04 - 1	TP-17-S	162	0.0101	-7470A	PPB	20 May 2019 18:03:53	U	SMPL
K2922-04 - 1	SOIL-COMP-1	156	0.0088	-7470A	PPB	20 May 2019 18:06:00	U	SMPL
K2922-08 - 1	SOIL-COMP-2	252	0.0303	-7470A	PPB	20 May 2019 18:08:08	U	SMPL
PB119908TB - 1	PB119797TB	47	-0.0156	-7470A	PPB	20 May 2019 18:10:15	U	SMPL
PB119910TB - 1	PB119850TB	33	-0.0188	-7470A	PPB	20 May 2019 18:12:24	U	SMPL
PB119911TB - 1	PB119855TB	173	0.0126	-7470A	PPB	20 May 2019 18:14:34	U	SMPL
CCV60 - 1	CCV60	22494	5.0096	-7470A	PPB	20 May 2019 18:16:41	U	SMPL
CCB60 - 1	CCB60	-28	-0.0324	-7470A	PPB	20 May 2019 18:18:48	U	SMPL
K2643-06LX5 - 1		26	-0.0203	-7470A	PPB	20 May 2019 18:20:55	U	SMPL
K2643-06A - 1		18883	4.2012	-7470A	PPB	20 May 2019 18:23:03	U	SMPL
K2913-05LX5 - 1		24	-0.0208	-7470A	PPB	20 May 2019 18:25:10	U	SMPL
K2913-05A - 1		19501	4.3396	-7470A	PPB	20 May 2019 18:27:18	U	SMPL
K2913-06LX5 - 1		24	-0.0208	-7470A	PPB	20 May 2019 18:29:27	U	SMPL
K2913-06A - 1		16858	3.7479	-7470A	PPB	20 May 2019 18:31:34	U	SMPL
K2641-12LX5 - 1		-7	-0.0277	-7470A	PPB	20 May 2019 18:33:42	U	SMPL
K2641-12A - 1		18798	4.1822	-7470A	PPB	20 May 2019 18:35:51	U	SMPL
CCV61 - 1	CCV61	22790	5.0759	-7470A	PPB	20 May 2019 18:38:00	U	SMPL
CCB61 - 1	CCB61	-13	-0.0291	-7470A	PPB	20 May 2019 18:40:08	U	SMPL