



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102872

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	05/20/2019	16:00
2	CAL2	1	Water	NA	NA	21.2	7.01	05/20/2019	16:01
3	CAL3	1	Water	NA	NA	21.2	10.03	05/20/2019	16:02
4	ICV	1	Water	NA	NA	21.2	6.99	05/20/2019	16:04
5	CCV1	1	Water	NA	NA	21.2	2.02	05/20/2019	16:05
6	K2913-05	1	Water	NA	NA	15.2	7.29	05/20/2019	16:07
7	K2913-05DUP	1	Water	NA	NA	15.3	7.30	05/20/2019	16:08
8	K2951-15	1	Water	NA	NA	22.6	7.04	05/20/2019	16:10
9	K2956-01	1	Water	NA	NA	23.4	8.42	05/20/2019	16:11
10	K2964-01	1	Water	NA	NA	25.6	5.77	05/20/2019	16:13
11	K2964-02	1	Water	NA	NA	25.9	5.73	05/20/2019	16:15
12	K2964-03	1	Water	NA	NA	25.6	5.46	05/20/2019	16:16
13	K2964-04	1	Water	NA	NA	25.3	6.54	05/20/2019	16:19
14	K2964-05	1	Water	NA	NA	24.9	6.79	05/20/2019	16:20
15	CCV2	1	Water	NA	NA	21.2	12.02	05/20/2019	16:22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2951

WorkList ID : 126322

Date : 05-20-2019 15:33:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2019	Water	K2913-05-1A	pH	Cool 4 deg C	PSEG01	K11	TP-17	05/16/2019	9040C
05/17/2019	Water	K2951-15	pH	Cool 4 deg C	PSEG01	K21	TP-14	05/17/2019	9040C
05/20/2019	Water	K2956-01	pH	Cool 4 deg C	PSEG01	K31	FM5210	05/20/2019	9040C
05/20/2019	Water	K2964-01	pH	Cool 4 deg C	PSEG01	O11	COMP-1	05/20/2019	9040C
05/20/2019	Water	K2964-02	pH	Cool 4 deg C	PSEG01	O11	COMP-2	05/20/2019	9040C
05/20/2019	Water	K2964-03	pH	Cool 4 deg C	PSEG01	O11	COMP-3	05/20/2019	9040C
05/20/2019	Water	K2964-04	pH	Cool 4 deg C	PSEG01	O11	501030	05/20/2019	9040C
05/20/2019	Water	K2964-05	pH	Cool 4 deg C	PSEG01	O11	13818	05/20/2019	9040C

Date/Time 05.20.19 15:35
 Received by: JD
 Relinquished by: JC

Date/Time 05.20.19 18:00
 Received by: JC
 Relinquished by: In