



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102873

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	05/20/2019	16:00
2	CAL2	1	Water	NA	NA	21.2	7.01	05/20/2019	16:01
3	CAL3	1	Water	NA	NA	21.2	10.03	05/20/2019	16:02
4	ICV	1	Water	NA	NA	21.2	6.99	05/20/2019	16:04
5	CCV1	1	Water	NA	NA	21.3	2.02	05/20/2019	16:30
6	K2959-01	1	Solid	20.03	20	22.6	8.54	05/20/2019	16:31
7	K2959-01DUP	1	Solid	20.04	20	22.7	8.55	05/20/2019	16:32
8	K2959-02	1	Solid	20.03	20	22.5	9.08	05/20/2019	16:34
9	K2959-03	1	Solid	20.04	20	22.6	9.10	05/20/2019	16:35
10	K2959-04	1	Solid	20.02	20	22.7	8.61	05/20/2019	16:36
11	K2959-05	1	Solid	20.04	20	22.8	8.52	05/20/2019	16:38
12	K2959-06	1	Solid	20.02	20	23.6	8.41	05/20/2019	16:40
13	K2959-07	1	Solid	20.03	20	23.8	8.47	05/20/2019	16:42
14	CCV2	1	Water	NA	NA	21.3	12.02	05/20/2019	16:45

WORKLIST(Hardcopy Internal Chain)

LB 102873

WorkList Name : PH S K2959

WorkList ID : 126324

Date : 05-20-2019 15:36:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/20/2019	Solid	K2959-01	pH	Cool 4 deg C	PSEG01	K11	20657	05/20/2019	9041A
05/20/2019	Solid	K2959-02	pH	Cool 4 deg C	PSEG01	K11	20658	05/20/2019	9041A
05/20/2019	Solid	K2959-03	pH	Cool 4 deg C	PSEG01	K11	20659	05/20/2019	9041A
05/20/2019	Solid	K2959-04	pH	Cool 4 deg C	PSEG01	K11	20660	05/20/2019	9041A
05/20/2019	Solid	K2959-05	pH	Cool 4 deg C	PSEG01	K11	20661	05/20/2019	9041A
05/20/2019	Solid	K2959-06	pH	Cool 4 deg C	PSEG01	K11	20662	05/20/2019	9041A
05/20/2019	Solid	K2959-07	pH	Cool 4 deg C	PSEG01	K11	20663	05/20/2019	9041A

Date/Time 05/20/19 15:40
 Received by: JD
 Relinquished by: JC

Date/Time 05/20/19 18:10
 Received by: JC
 Relinquished by: JD