



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102874

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	05/20/2019	16:00
2	CAL2	1	Water	NA	NA	21.2	7.01	05/20/2019	16:01
3	CAL3	1	Water	NA	NA	21.2	10.03	05/20/2019	16:02
4	ICV	1	Water	NA	NA	21.2	6.99	05/20/2019	16:04
5	CCV1	1	Water	NA	NA	21.2	2.02	05/20/2019	17:20
6	K2951-04	1	Solid	20.03	20	23.9	9.75	05/20/2019	17:21
7	K2951-04DUP	1	Solid	20.04	20	23.8	9.76	05/20/2019	17:22
8	K2951-10	1	Solid	20.05	20	23.9	7.96	05/20/2019	17:25
9	K2951-14	1	Solid	20.03	20	23.5	8.33	05/20/2019	17:26
10	CCV2	1	Water	NA	NA	21.3	12.02	05/20/2019	17:30

WORKLIST(Hardcopy Internal Chain)

LB 102874

WorkList Name : corrosivity K2962

WorkList ID : 126327

Date : 05-20-2019 15:38:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/17/2019	Solid	K2951-04	Corrosivity	Cool 4 deg C	PSEG01	K21	TP-12-S	05/17/2019	9045D
05/17/2019	Solid	K2951-10	Corrosivity	Cool 4 deg C	PSEG01	K21	TP-14-S	05/17/2019	9045D
05/17/2019	Solid	K2951-14	Corrosivity	Cool 4 deg C	PSEG01	K21	TP-15-S	05/17/2019	9045D

Date/Time 05.20.19 15:40
 Received by: Jg
 Relinquished by: Jc

Date/Time 05.20.19 18:15
 Received by: Jc
 Relinquished by: Jg