

LB102885


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51723



A= 0.0000e+000

B= 1.9374e-004

C= -5.4021e-002

Rho= 0.9998908

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.042	-0.042	64	0.000	64					
0.05	0.050										
0.20	0.200	0.160	-0.040	1107	0.0 %	1107					
2.50	2.500	2.562	0.062	13503	0.0 %	13503					-20
5.00	5.000	5.090	0.090	26552	0.0 %	26552					2
7.50	7.500	7.462	-0.038	38793	0.0 %	38793					2
10.0	10.000	9.967	-0.033	51723	0.0 %	51723					-1
											0

RS

05.21.19

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	64	-		0 ISM02.4	PPB	5/21/2019 9:45	Std
	0.2 S0.2	1107	-		0.2 ISM02.4	PPB	5/21/2019 9:47	Std
	2.5 S2.5	13503	-		2.5 ISM02.4	PPB	5/21/2019 9:49	Std
	5 S5.0	26552	-		5 ISM02.4	PPB	5/21/2019 9:52	Std
	7.5 S7.5	38793	-		7.5 ISM02.4	PPB	5/21/2019 9:54	Std
	10 S10.0	51723	-		10 ISM02.4	PPB	5/21/2019 9:56	Std
ICV	ICV	19919	3.8052	-	ISM02.4	PPB	5/21/2019 9:59	SMPL
ICB	ICB	12	-0.0517	-	ISM02.4	PPB	5/21/2019 10:02	SMPL
CCV062	CCV062	26814	5.141	-	ISM02.4	PPB	5/21/2019 10:04	SMPL
CCB062	CCB062	11	-0.0519	-	ISM02.4	PPB	5/21/2019 10:06	SMPL
PB119708BL	PBS708	47	-0.0449	-	ISM02.4	PPB	5/21/2019 10:08	SMPL
K2720-20	MH0AS4	20142	3.8484	-	ISM02.4	PPB	5/21/2019 10:11	SMPL
K2720-21	MH0AS4D	19131	3.6525	-	ISM02.4	PPB	5/21/2019 10:13	SMPL
K2720-22	MH0AS4S	27655	5.3039	-	ISM02.4	PPB	5/21/2019 10:18	SMPL
PB119709BL	PBS709	27	-0.0488	-	ISM02.4	PPB	5/21/2019 10:25	SMPL
K2722-01	MH0AS5	29302	5.623	-	ISM02.4	PPB	5/21/2019 10:28	SMPL
K2722-02	MH0AS5D	28999	5.5643	-	ISM02.4	PPB	5/21/2019 10:30	SMPL
K2722-03	MH0AS5S	36713	7.0589	-	ISM02.4	PPB	5/21/2019 10:32	SMPL
K2722-04	MH0AS6	31678	6.0834	-	ISM02.4	PPB	5/21/2019 10:35	SMPL
K2722-05	MH0AS7	19828	3.7875	-	ISM02.4	PPB	5/21/2019 10:37	SMPL
K2722-06	MH0AS8	22164	4.2401	-	ISM02.4	PPB	5/21/2019 10:39	SMPL
K2722-07	MH0AS9	15838	3.0145	-	ISM02.4	PPB	5/21/2019 10:41	SMPL
K2722-08	MH0AT0	18336	3.4985	-	ISM02.4	PPB	5/21/2019 10:44	SMPL
K2722-09	MH0AT1	18336	3.4985	-	ISM02.4	PPB	5/21/2019 10:46	SMPL
K2722-10	MH0AT2	12809	2.4276	-	ISM02.4	PPB	5/21/2019 10:48	SMPL
PB119710BL	PBS710	-32	-0.0602	-	ISM02.4	PPB	5/21/2019 10:50	SMPL
K2726-01	MH0B17	1570	0.2502	-	ISM02.4	PPB	5/21/2019 10:53	SMPL
K2726-02	MH0B18	1983	0.3302	-	ISM02.4	PPB	5/21/2019 10:55	SMPL
K2726-03	MH0B19	2976	0.5226	-	ISM02.4	PPB	5/21/2019 10:57	SMPL
K2726-04	MH0B20	2251	0.3821	-	ISM02.4	PPB	5/21/2019 11:00	SMPL
CCV063	CCV063	26172	5.0166	-	ISM02.4	PPB	5/21/2019 11:02	SMPL
CCB063	CCB063	-17	-0.0573	-	ISM02.4	PPB	5/21/2019 11:04	SMPL
K2726-05	MH0B21	2884	0.5047	-	ISM02.4	PPB	5/21/2019 11:06	SMPL
K2726-06	MH0B22	3067	0.5402	-	ISM02.4	PPB	5/21/2019 11:09	SMPL
K2726-07	MH0B23	1284	0.1947	-	ISM02.4	PPB	5/21/2019 11:11	SMPL
K2726-08	MH0B24	2552	0.4404	-	ISM02.4	PPB	5/21/2019 11:13	SMPL
K2726-09	MH0B25	2884	0.5047	-	ISM02.4	PPB	5/21/2019 11:15	SMPL
K2726-10	MH0B26	5244	0.962	-	ISM02.4	PPB	5/21/2019 11:18	SMPL
K2726-11	MH0B27	8635	1.619	-	ISM02.4	PPB	5/21/2019 11:20	SMPL
K2726-12	MH0B28	8455	1.5841	-	ISM02.4	PPB	5/21/2019 11:22	SMPL
K2726-13	MH0B29	34137	6.5598	-	ISM02.4	PPB	5/21/2019 11:25	SMPL
K2726-14	MH0B30	76886	14.8421	-	ISM02.4	PPB	5/21/2019 11:27	SMPL
K2726-15	MH0B31	21164	4.0464	-	ISM02.4	PPB	5/21/2019 11:29	SMPL
K2726-16	MH0B32	32110	6.1671	-	ISM02.4	PPB	5/21/2019 11:31	SMPL
K2726-17	MH0B33	16210	3.0866	-	ISM02.4	PPB	5/21/2019 11:34	SMPL
K2726-18	MH0B34	17616	3.359	-	ISM02.4	PPB	5/21/2019 11:36	SMPL

LB102885

INSTRUMENT ID : CV1

K2726-19	MH0B35	20462	3.9104 -	ISM02.4	PPB	5/21/2019 11:38 SMPL
K2726-20	MH0B36	20549	3.9272 -	ISM02.4	PPB	5/21/2019 11:40 SMPL
K2726-21	MH0B36D	20213	3.8621 -	ISM02.4	PPB	5/21/2019 11:43 SMPL
K2726-22	MH0B36S	33005	6.3405 -	ISM02.4	PPB	5/21/2019 11:45 SMPL
K2726-14DLX4		18530	3.536 -	ISM02.4	PPB	5/21/2019 11:47 SMPL
CCV064	CCV064	25983	4.98 -	ISM02.4	PPB	5/21/2019 11:50 SMPL
CCB064	CCB064	-32	-0.0602 -	ISM02.4	PPB	5/21/2019 11:52 SMPL