



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB102887
Analysis Date: 05/21/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 05/21/2019
Extraction IN Time: 07:20
Extraction OUT Time: 07:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB102887BL	LB102887BL	WATER	1.3	1000	100	2.5697	2.5697	3.02	2.5697	2.5697	0.0000	0
2	LB102887BS	LB102887BS	WATER	1.3	1000	100	2.1032	2.1032	3.01	2.1219	2.1219	0.0187	18.7
3	LB102887BSD	LB102887BSD	WATER	1.3	1000	100	2.5487	2.5487	3.03	2.5676	2.5676	0.0189	18.9
4	K2913-05	TP-17	WATER	1.3	1000	100	3.1427	3.1427	3.04	3.1456	3.1456	0.0029	2.9
5	K2951-15	TP-14	WATER	1.3	1000	100	3.0969	3.0969	3.03	3.0984	3.0984	0.0015	1.5
6	K2956-01	FM5210	WATER	1.3	1000	100	3.0999	3.0999	3.05	3.1016	3.1016	0.0017	1.7



QC Batch# LB102887 Test: TPH Analysis Date: 05/21/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2515
pH Paper 0-14	M4240
Sodium Sulfate	EP1957
1:1 HCL	WP74116
Silica Gel	W2466
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP74117
LCSWD	10.00 ML	WP74118
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:11</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>08:10</u>		
Bal Check Time:	<u>07:30</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>09:40</u>
		Out Time:	<u>09:10</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>10:46</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>10:15</u>		
Bal Check Time:	<u>11:45</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:20</u>
		Out Time:	<u>10:45</u>		

LB 102887

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K2913

WorkList ID : 126331

Date : 05-21-2019 7:15:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/23/2019	Water	K2913-05-17	TPH	Conc H2SO4 tr	PSEG01	K11	TP-17	05/16/2019	1664A
05/24/2019	Water	K2951-15-C	TPH	Conc H2SO4 tr	PSEG01	K21	TP-14	05/17/2019	1664A
05/23/2019	Water	K2956-01 -K	TPH	Conc H2SO4 tr	PSEG01	K31	FM5210	05/20/2019	1664A

Date/Time 05.21.19 7:20
 Received by: Jo
 Relinquished by: Jo

Date/Time 05.21.19 10:00
 Received by: Jo
 Relinquished by: Jo