



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB102894

ANALYST: Aparna
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC PH METER-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include various calibration standards and reagents like Hexachrome, Hexavalent Chromium ICV Std, and 5N sulfuric acid.

Intercept: 0.0007 Slope: 0.7721 Regression: 0.999975

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Aparna

Parameter: Hexavalent Chromiu

SUPERVISOR REVIEW BY: apatel

Run Number: LB102894

pH Meter ID: WC PH METER-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH		Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
						HN03	H2SO4	Backgrnd	Color				
1	ICV	0.5	1	100	100		2.05	0.000	0.392	0.392	0.507	05/21/2019	12:03
2	ICB		1	100	100		2.10	0.000	0.000	0.000	-0.001	05/21/2019	12:04
3	CCV1	0.5	1	100	100		2.03	0.000	0.400	0.400	0.517	05/21/2019	12:04
4	CCB1		1	100	100		2.00	0.000	0.000	0.000	-0.001	05/21/2019	12:05
5	PB119918BL		1	2.50	100	7.25	1.90	0.000	0.000	0.000	-0.001	05/21/2019	12:05
6	PB119918BS	20	1	2.50	100	7.65	1.75	0.029	0.378	0.349	0.451	05/21/2019	12:06
7	K2951-03		1	2.53	100	7.50	2.01	0.007	0.007	0.000	-0.001	05/21/2019	12:06
8	K2951-03DUP		1	2.51	100	7.23	2.13	0.007	0.007	0.000	-0.001	05/21/2019	12:07
9	K2951-03MSPre	40	2	2.50	100	7.12	2.31	0.021	0.022	0.001	0.000	05/21/2019	12:07
10	K2951-03MS2Ins	1284	40	2.50	100	7.15	2.25	0.002	0.679	0.677	0.876	05/21/2019	12:08
11	K2951-03MS3Post	40	2	2.54	100	7.89	2.05	0.014	0.014	0.000	-0.001	05/21/2019	12:08
12	K2951-09		1	2.52	100	7.19	1.99	0.010	0.015	0.005	0.006	05/21/2019	12:09
13	K2951-13		1	2.55	100	7.50	1.80	0.020	0.020	0.000	-0.001	05/21/2019	12:09
14	K2954-01		1	2.53	100	7.23	2.11	0.041	0.041	0.000	-0.001	05/21/2019	12:10
15	CCV2	0.5	1	100	100		2.14	0.000	0.401	0.401	0.518	05/21/2019	12:10
16	CCB2		1	100	100		1.78	0.000	0.000	0.000	-0.001	05/21/2019	12:11
17	K2954-02		1	2.55	100	7.23	2.13	0.012	0.012	0.000	-0.001	05/21/2019	12:11
18	K2954-03		1	2.50	100	7.65	1.96	0.234	0.234	0.000	-0.001	05/21/2019	12:12
19	K2954-04		1	2.51	100	7.12	2.08	0.075	0.075	0.000	-0.001	05/21/2019	12:12
20	K2954-05		1	2.53	100	7.50	2.18	0.081	0.081	0.000	-0.001	05/21/2019	12:13
21	CCV3	0.5	1	100	100		2.15	0.000	0.402	0.402	0.520	05/21/2019	12:13
22	CCB3		1	100	100		2.10	0.000	0.000	0.000	-0.001	05/21/2019	12:14