



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102901

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/21/2019	09:40
2	K2951-16		1	8	100.00	>212.0	>212.0	05/21/2019	10:05
3	K2951-16DUP		1	8	100.00	>212.0	>212.0	05/21/2019	10:40
4	K2956-02		1	9	100.00	>212.0	>212.0	05/21/2019	11:10
5	K2964-01		1	7	100.00	>212.0	>212.0	05/21/2019	11:45
6	K2964-02		1	8	100.00	>212.0	>212.0	05/21/2019	12:20
7	K2964-03		1	9	100.00	>212.0	>212.0	05/21/2019	12:50
8	K2964-04		1	7	100.00	>212.0	>212.0	05/21/2019	13:20
9	K2964-05		1	7	100.00	>212.0	>212.0	05/21/2019	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 102901

WorkList Name : FP52119

WorkList ID : 126360

Date : 05-21-2019 09:08:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/24/2019	Water	K2951-16	B	Flash Point	Cool 4 deg C	PSEG01	K21 TP-14	05/17/2019	1010A
05/23/2019	Water	K2956-02	C	Flash Point	Cool 4 deg C	PSEG01	K31 FM5210	05/20/2019	1010A
05/27/2019	Water	K2964-01	F	Flash Point	Cool 4 deg C	PSEG01	O11 COMP-1	05/20/2019	1010A
05/27/2019	Water	K2964-02	E	Flash Point	Cool 4 deg C	PSEG01	O11 COMP-2	05/20/2019	1010A
05/27/2019	Water	K2964-03	F	Flash Point	Cool 4 deg C	PSEG01	O11 COMP-3	05/20/2019	1010A
05/27/2019	Water	K2964-04	E	Flash Point	Cool 4 deg C	PSEG01	O11 501030	05/20/2019	1010A
05/27/2019	Water	K2964-05	F	Flash Point	Cool 4 deg C	PSEG01	O11 13818	05/20/2019	1010A

Date/Time 05/21/19 09:10

Received by: KS

Relinquished by: CP

Date/Time 05/21/19 13:50

Received by: CP

Relinquished by: KS