


LB102903

7471B

INSTRUMENT ID : CU1

 μ Abs. :

64081



Linear

A= 0.0000e+000

B= 1.5645e-004

C= -3.3656e-002

Rho= 0.9999846

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.025	-0.025	55	0.000	55				
0.2	0.200	0.229	0.029	1676	0.0 %	1676				
2.5	2.500	2.503	0.003	16215	0.0 %	16215				
5.0	5.000	4.978	-0.022	32035	0.0 %	32035				
7.5	7.500	7.523	0.023	48299	0.0 %	48299				
10.0	10.000	9.992	-0.008	64081	0.0 %	64081				

LB102903 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 May 2019 14:20:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	55	-	0.00007471B	PPB		21 May 2019 14:25:15	S	Std
0.2 - 1	50-2	1676	-	0.20007471B	PPB		21 May 2019 14:27:32	S	Std
2.5 - 1	52-5	16215	-	2.50007471B	PPB		21 May 2019 14:29:50	S	Std
5.0 - 1	55-0	32035	-	5.00007471B	PPB		21 May 2019 14:32:07	S	Std
7.5 - 1	57-5	48299	-	7.50007471B	PPB		21 May 2019 14:34:24	S	Std
10.0 - 1	510-0	64081	-	10.00007471B	PPB		21 May 2019 14:36:41	S	Std
ICV54 - 1	ICV54	24439	3.7899	-7471B	PPB		21 May 2019 14:40:38	U	SMPL
ICB54 - 1	ICB54	-29	-0.0382	-7471B	PPB		21 May 2019 14:42:53	U	SMPL
CCV65 - 1	CCV65	31913	4.9592	-7471B	PPB		21 May 2019 14:45:09	U	SMPL
CCB65 - 1	CCB65	-45	-0.0407	-7471B	PPB		21 May 2019 14:47:24	U	SMPL
CRA - 1	CRA	1667	0.2272	-7471B	PPB		21 May 2019 14:49:41	U	SMPL
HighStd - 1	HighStd	64062	9.9891	-7471B	PPB		21 May 2019 14:51:55	U	SMPL
ChkStd - 1	ChkStd	44534	6.9338	-7471B	PPB		21 May 2019 14:54:10	U	SMPL
PB119926BL - 1	PB119926BL	-95	-0.0485	-7471B	PPB		21 May 2019 14:56:28	U	SMPL
PB119926BS - 1	PB119926BS	23263	3.6059	-7471B	PPB		21 May 2019 14:58:45	U	SMPL
K2935-04 - 1	PURGE-WATER-MUD	136	-0.0124	-7471B	PPB		21 May 2019 15:01:02	U	SMPL
K2935-04DUP - 1	PURGE-WATER-MUDDUP	204	-0.0017	-7471B	PPB		21 May 2019 15:03:19	U	SMPL
K2935-04MS - 1	PURGE-WATER-MUDMS	26009	4.0355	-7471B	PPB		21 May 2019 15:05:36	U	SMPL
K2935-04MSD - 1	PURGE-WATER-MUDMSD	25776	3.9991	-7471B	PPB		21 May 2019 15:07:53	U	SMPL
K2951-03 - 1	TP-12-S	4226	0.6275	-7471B	PPB		21 May 2019 15:10:10	U	SMPL
CCV66 - 1	CCV66	32014	4.9750	-7471B	PPB		21 May 2019 15:12:25	U	SMPL
CCB66 - 1	CCB66	-54	-0.0421	-7471B	PPB		21 May 2019 15:14:40	U	SMPL
K2951-09 - 1	TP-14-S	9639	1.4744	-7471B	PPB		21 May 2019 15:16:56	U	SMPL
K2951-13 - 1	TP-15-S	53285	8.3030	-7471B	PPB		21 May 2019 15:19:11	U	SMPL
K2954-01 - 1	20190405-COMP-CH-82	1810	0.2495	-7471B	PPB		21 May 2019 15:21:27	U	SMPL
K2954-02 - 1	20190406-COMP-CH-83	6657	1.0079	-7471B	PPB		21 May 2019 15:23:45	U	SMPL
K2954-03 - 1	20190332A-GLP-1	218	0.0005	-7471B	PPB		21 May 2019 15:26:01	U	SMPL
K2954-04 - 1	20190333A-GLP-2	1044	0.1297	-7471B	PPB		21 May 2019 15:28:17	U	SMPL
K2954-05 - 1	20190334A-GLP-3	8204	1.2499	-7471B	PPB		21 May 2019 15:30:34	U	SMPL
K2962-02 - 1	DB-2	597	0.0597	-7471B	PPB		21 May 2019 15:32:51	U	SMPL
K2962-06 - 1	DB-3	506	0.0455	-7471B	PPB		21 May 2019 15:35:08	U	SMPL
K2962-10 - 1	DB-4	647	0.0676	-7471B	PPB		21 May 2019 15:37:24	U	SMPL
CCV67 - 1	CCV67	32143	4.9952	-7471B	PPB		21 May 2019 15:39:40	U	SMPL
CCB67 - 1	CCB67	-25	-0.0376	-7471B	PPB		21 May 2019 15:41:55	U	SMPL
K2966-01 - 1	FESB-13(9.5-10)	3082	0.4485	-7471B	PPB		21 May 2019 15:44:11	U	SMPL
K2966-02 - 1	FESB-14(4.5-5)	2634	0.3784	-7471B	PPB		21 May 2019 15:46:27	U	SMPL
K2966-03 - 1	FESB-15(1.2-1.7)	2815	0.4068	-7471B	PPB		21 May 2019 15:48:43	U	SMPL
K2966-04 - 1	FESB-16(2.4-2.9)	273	0.0091	-7471B	PPB		21 May 2019 15:50:59	U	SMPL
K2966-05 - 1	FESB-17(1.9-2.4)	2360	0.3356	-7471B	PPB		21 May 2019 15:53:15	U	SMPL
K2966-06 - 1	FESB-18(2.5-3.0)	20053	3.1037	-7471B	PPB		21 May 2019 15:55:32	U	SMPL
K2935-04LX5 - 1		13	-0.0316	-7471B	PPB		21 May 2019 15:57:48	U	SMPL
K2935-04A - 1		24798	3.8461	-7471B	PPB		21 May 2019 16:00:05	U	SMPL
CCV68 - 1	CCV68	32060	4.9822	-7471B	PPB		21 May 2019 16:02:22	U	SMPL
CCB68 - 1	CCB68	412	0.0308	-7471B	PPB		21 May 2019 16:04:38	U	SMPL