

LB102909


ISM02.4

INSTRUMENT ID : CU1

Linear

 μ Abs.:

44474



A= 0.0000e+000

B= 2.2284e-004

C= 8.5391e-003

Rho= 0.9998097

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.033	0.033	112	0.000	112					—
0.05	0.050										
0.20	0.200	0.198	-0.002	849	0.0 %	849					-1
2.50	2.500	2.412	-0.088	10786	0.0 %	10786					-4
5.00	5.000	5.013	0.013	22460	0.0 %	22460					0
7.50	7.500	7.624	0.124	34177	0.0 %	34177					2
10.0	10.000	9.919	-0.081	44474	0.0 %	44474					-1

LB102909
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	112	-		0 ISM02.4	PPB	5/21/2019 16:32	Std
	0.2 S0.2	849	-		0.2 ISM02.4	PPB	5/21/2019 16:34	Std
	2.5 S2.5	10786	-		2.5 ISM02.4	PPB	5/21/2019 16:37	Std
	5 S5.0	22460	-		5 ISM02.4	PPB	5/21/2019 16:39	Std
	7.5 S7.5	34177	-		7.5 ISM02.4	PPB	5/21/2019 16:41	Std
	10 S10.0	44474	-		10 ISM02.4	PPB	5/21/2019 16:43	Std
ICV	ICV	18368	4.1016 -		ISM02.4	PPB	5/21/2019 16:47	SMPL
ICB	ICB	7	0.0101 -		ISM02.4	PPB	5/21/2019 16:50	SMPL
CCV069	CCV069	22654	5.0567 -		ISM02.4	PPB	5/21/2019 16:52	SMPL
CCB069	CCB069	17	0.0123 -		ISM02.4	PPB	5/21/2019 16:54	SMPL
PB119947BL	PBW947	93	0.0293 -		ISM02.4	PPB	5/21/2019 16:56	SMPL
K2787-07	MX2G89	12748	2.8492 -		ISM02.4	PPB	5/21/2019 16:59	SMPL
K2787-08	MX2G89D	12617	2.8201 -		ISM02.4	PPB	5/21/2019 17:01	SMPL
K2787-09	MX2G89S	17727	3.9587 -		ISM02.4	PPB	5/21/2019 17:03	SMPL
PB119948BL	PBW948	-49	-0.0024 -		ISM02.4	PPB	5/21/2019 17:05	SMPL
K2812-04	MA4247	7243	1.6225 -		ISM02.4	PPB	5/21/2019 17:08	SMPL
K2812-05	MA4247D	6107	1.3694 -		ISM02.4	PPB	5/21/2019 17:10	SMPL
K2812-06	MA4247S	11305	2.5277 -		ISM02.4	PPB	5/21/2019 17:12	SMPL
PB119949BL	PBW949	-316	-0.0619 -		ISM02.4	PPB	5/21/2019 17:15	SMPL
K2785-02	MA5112	16261	3.6321 -		ISM02.4	PPB	5/21/2019 17:17	SMPL
K2785-03	MA5113	117	0.0346 -		ISM02.4	PPB	5/21/2019 17:19	SMPL
K2785-04	MA5115	142	0.0402 -		ISM02.4	PPB	5/21/2019 17:22	SMPL
K2785-05	MA5116	313	0.0783 -		ISM02.4	PPB	5/21/2019 17:24	SMPL
K2785-06	MA5117	116	0.0344 -		ISM02.4	PPB	5/21/2019 17:26	SMPL
K2785-07	MA5118	86	0.0277 -		ISM02.4	PPB	5/21/2019 17:28	SMPL
K2785-08	MA5119	78	0.0259 -		ISM02.4	PPB	5/21/2019 17:31	SMPL
K2785-09	MA5121	124	0.0362 -		ISM02.4	PPB	5/21/2019 17:33	SMPL
K2785-10	MA5122	172	0.0469 -		ISM02.4	PPB	5/21/2019 17:35	SMPL
K2785-11	MA5122D	111	0.0333 -		ISM02.4	PPB	5/21/2019 17:37	SMPL
K2785-12	MA5122S	5265	1.1818 -		ISM02.4	PPB	5/21/2019 17:40	SMPL
CCV070	CCV070	22821	5.0939 -		ISM02.4	PPB	5/21/2019 17:42	SMPL
CCB070	CCB070	-19	0.0043 -		ISM02.4	PPB	5/21/2019 17:44	SMPL
PB119950BL	PBW950	66	0.0232 -		ISM02.4	PPB	5/21/2019 17:47	SMPL
K2863-01	METC20	142	0.0402 -		ISM02.4	PPB	5/21/2019 17:49	SMPL
K2863-02	METC20D	139	0.0395 -		ISM02.4	PPB	5/21/2019 17:51	SMPL
K2863-03	METC20S	5314	1.1927 -		ISM02.4	PPB	5/21/2019 17:53	SMPL
PB119951BL	PBW951	-7	0.007 -		ISM02.4	PPB	5/21/2019 17:56	SMPL
K2784-01	MA5102	104	0.0317 -		ISM02.4	PPB	5/21/2019 17:58	SMPL
K2784-02	MA5103	146	0.0411 -		ISM02.4	PPB	5/21/2019 18:00	SMPL
K2784-03	MA5103D	134	0.0384 -		ISM02.4	PPB	5/21/2019 18:02	SMPL
K2784-04	MA5103S	5481	1.2299 -		ISM02.4	PPB	5/21/2019 18:05	SMPL
K2784-05	MA5104	84	0.0273 -		ISM02.4	PPB	5/21/2019 18:07	SMPL
K2784-06	MA5105	200	0.0531 -		ISM02.4	PPB	5/21/2019 18:09	SMPL
K2784-07	MA5106	240	0.062 -		ISM02.4	PPB	5/21/2019 18:12	SMPL
K2784-08	MA5107	201	0.0533 -		ISM02.4	PPB	5/21/2019 18:14	SMPL
K2784-10	MA5111	17406	3.8872 -		ISM02.4	PPB	5/21/2019 18:16	SMPL

INSTRUMENT ID : CV1

K2784-11	MA5114	155	0.0431 -	ISM02.4	PPB	5/21/2019 18:18 SMPL
K2784-12	MA5120	183	0.0493 -	ISM02.4	PPB	5/21/2019 18:21 SMPL
PB119952BL	PBW952	5	0.0097 -	ISM02.4	PPB	5/21/2019 18:23 SMPL
K2722-11	MH0BH7	38	0.017 -	ISM02.4	PPB	5/21/2019 18:25 SMPL
K2722-12	MH0BH8	70	0.0241 -	ISM02.4	PPB	5/21/2019 18:27 SMPL
K2722-13	MH0BH9	83	0.027 -	ISM02.4	PPB	5/21/2019 18:30 SMPL
CCV071	CCV071	22688	5.0642 -	ISM02.4	PPB	5/21/2019 18:32 SMPL
CCB071	CCB071	-19	0.0043 -	ISM02.4	PPB	5/21/2019 18:34 SMPL
K2722-14	MH0BJ0	64	0.0228 -	ISM02.4	PPB	5/21/2019 18:37 SMPL
CCV072	CCV072	22976	5.1284 -	ISM02.4	PPB	5/21/2019 18:46 SMPL
CCB072	CCB072	-14	0.0054 -	ISM02.4	PPB	5/21/2019 18:48 SMPL