

LB102920


7470A

INSTRUMENT ID : CU1

Linear

 μ Abs.:

4331E



A= 0.0000e+000

B= 2.3200e-004

C= -4.1365e-002

Rho= 0.9998402

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	58	0.000	58				
0.2	0.200	0.167	-0.033	896	0.0 %	896				
2.5	2.500	2.546	0.046	11153	0.0 %	11153				
5.0	5.000	5.108	0.108	22196	0.0 %	22196				
7.5	7.500	7.399	-0.101	32068	0.0 %	32068				
10.0	10.000	10.009	0.009	43318	0.0 %	43318				

LB102920 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 May 2019 13:27:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	58	-	0.00007470A	PPB		22 May 2019 13:30:06	S	Std
0.2 - 1	50-2	896	-	0.20007470A	PPB		22 May 2019 13:32:15	S	Std
2.5 - 1	52-5	11153	-	2.50007470A	PPB		22 May 2019 13:34:23	S	Std
5.0 - 1	55-0	22196	-	5.00007470A	PPB		22 May 2019 13:36:31	S	Std
7.5 - 1	57-5	32068	-	7.50007470A	PPB		22 May 2019 13:38:38	S	Std
10.0 - 1	510-0	43318	-	10.00007470A	PPB		22 May 2019 13:44:47	S	Std
ICV56 - 1	ICV56	17071	3.9192	-7470A	PPB		22 May 2019 13:47:28	U	SMPL
ICB56 - 1	ICB56	11	-0.0388	-7470A	PPB		22 May 2019 13:49:34	U	SMPL
CCV73 - 1	CCV73	21952	5.0516	-7470A	PPB		22 May 2019 13:51:40	U	SMPL
CCB73 - 1	CCB73	-24	-0.0469	-7470A	PPB		22 May 2019 13:53:46	U	SMPL
CRA - 1	CRA	871	0.1607	-7470A	PPB		22 May 2019 13:55:52	U	SMPL
HighStd - 1	HighStd	42799	9.8881	-7470A	PPB		22 May 2019 13:57:58	U	SMPL
ChkStd - 1	ChkStd	29953	6.9078	-7470A	PPB		22 May 2019 14:00:04	U	SMPL
PB119977BL - 1	PBW	-33	-0.0490	-7470A	PPB		22 May 2019 14:02:12	U	SMPL
PB119977BS - 1	LCSW	16581	3.8055	-7470A	PPB		22 May 2019 14:04:20	U	SMPL
K2951-16 - 1	TP-14	-5	-0.0425	-7470A	PPB		22 May 2019 14:06:28	U	SMPL
K2951-16DUP - 1	TP-14DUP	60	-0.0274	-7470A	PPB		22 May 2019 14:08:35	U	SMPL
K2951-16MS - 1	TP-14MS	17688	4.0623	-7470A	PPB		22 May 2019 14:10:43	U	SMPL
K2951-16MSD - 1	TP-14MSD	17983	4.1307	-7470A	PPB		22 May 2019 14:12:51	U	SMPL
K2956-02 - 1	FM5210	-1	-0.0416	-7470A	PPB		22 May 2019 14:14:57	U	SMPL
CCV74 - 1	CCV74	21972	5.0562	-7470A	PPB		22 May 2019 14:17:03	U	SMPL
CCB74 - 1	CCB74	-25	-0.0472	-7470A	PPB		22 May 2019 14:19:10	U	SMPL
PB119978BL - 1	PBW	49	-0.0300	-7470A	PPB		22 May 2019 14:21:16	U	SMPL
PB119978BS - 1	LCSW	16973	3.8964	-7470A	PPB		22 May 2019 14:23:23	U	SMPL
K2951-04 - 1	TP-12-S	19	-0.0370	-7470A	PPB		22 May 2019 14:25:29	U	SMPL
K2951-04DUP - 1	TP-12-SDUP	79	-0.0230	-7470A	PPB		22 May 2019 14:27:36	U	SMPL
K2951-04MS - 1	TP-12-SMS	17421	4.0004	-7470A	PPB		22 May 2019 14:29:44	U	SMPL
K2951-04MSD - 1	TP-12-SMSD	17055	3.9154	-7470A	PPB		22 May 2019 14:31:51	U	SMPL
K2951-10 - 1	TP-14-S	102	-0.0177	-7470A	PPB		22 May 2019 14:33:58	U	SMPL
K2951-14 - 1	TP-15-S	89	-0.0207	-7470A	PPB		22 May 2019 14:36:06	U	SMPL
K2959-01 - 1	20657	114	-0.0149	-7470A	PPB		22 May 2019 14:38:13	U	SMPL
K2959-02 - 1	20658	138	-0.0093	-7470A	PPB		22 May 2019 14:40:20	U	SMPL
CCV75 - 1	CCV75	22143	5.0959	-7470A	PPB		22 May 2019 14:42:26	U	SMPL
CCB75 - 1	CCB75	4	-0.0404	-7470A	PPB		22 May 2019 14:44:32	U	SMPL
K2959-03 - 1	20659	104	-0.0172	-7470A	PPB		22 May 2019 14:46:39	U	SMPL
K2959-04 - 1	20660	116	-0.0145	-7470A	PPB		22 May 2019 14:48:46	U	SMPL
K2959-05 - 1	20661	148	-0.0070	-7470A	PPB		22 May 2019 14:50:52	U	SMPL
K2959-06 - 1	20662	99	-0.0184	-7470A	PPB		22 May 2019 14:52:58	U	SMPL
K2959-07 - 1	20663	83	-0.0221	-7470A	PPB		22 May 2019 14:55:05	U	SMPL
K2963-01 - 1	DRUM-1	221	0.0099	-7470A	PPB		22 May 2019 14:57:13	U	SMPL
PB119979BL - 1	PBW	28	-0.0349	-7470A	PPB		22 May 2019 14:59:20	U	SMPL
PB119979BS - 1	LCSW	17113	3.9289	-7470A	PPB		22 May 2019 15:01:29	U	SMPL
K2951-15 - 1	TP-14	2415	0.5189	-7470A	PPB		22 May 2019 15:03:38	U	SMPL
K2951-15DUP - 1	TP-14DUP	2437	0.5240	-7470A	PPB		22 May 2019 15:05:44	U	SMPL
CCV76 - 1	CCV76	21460	4.9374	-7470A	PPB		22 May 2019 15:07:51	U	SMPL
CCB76 - 1	CCB76	-12	-0.0441	-7470A	PPB		22 May 2019 15:09:57	U	SMPL
K2951-15MS - 1	TP-14MS	21713	4.9961	-7470A	PPB		22 May 2019 15:12:04	U	SMPL
K2951-15MSD - 1	TP-14MSD	21611	4.9725	-7470A	PPB		22 May 2019 15:14:11	U	SMPL
K2954-06 - 1	20190407-FIELD-BLANK-2	14	-0.0381	-7470A	PPB		22 May 2019 15:16:17	U	SMPL
K2956-01 - 1	FM5210	1037	0.1992	-7470A	PPB		22 May 2019 15:18:24	U	SMPL
K2964-01 - 1	COMP-1	510	0.0770	-7470A	PPB		22 May 2019 15:20:30	U	SMPL
K2964-02 - 1	COMP-2	946	0.1781	-7470A	PPB		22 May 2019 15:22:38	U	SMPL
K2964-03 - 1	COMP-3	243	0.0150	-7470A	PPB		22 May 2019 15:24:45	U	SMPL
K2964-04 - 1	51030	527	0.0809	-7470A	PPB		22 May 2019 15:26:53	U	SMPL
K2964-05 - 1	13818	44	-0.0312	-7470A	PPB		22 May 2019 15:29:01	U	SMPL
K2966-08 - 1	FB-20190520	14	-0.0381	-7470A	PPB		22 May 2019 15:31:07	U	SMPL
CCV77 - 1	CCV77	21577	4.9646	-7470A	PPB		22 May 2019 15:33:14	U	SMPL
CCB77 - 1	CCB77	-40	-0.0506	-7470A	PPB		22 May 2019 15:35:20	U	SMPL
K2970-17 - 1	TP-10	1064	0.2055	-7470A	PPB		22 May 2019 15:37:27	U	SMPL
PB119977TB - 1	PB119916TB	103	-0.0175	-7470A	PPB		22 May 2019 15:39:33	U	SMPL
PB119978TB - 1	PB119914TB	116	-0.0145	-7470A	PPB		22 May 2019 15:41:40	U	SMPL
K2951-16LX5 - 1		60	-0.0274	-7470A	PPB		22 May 2019 15:43:46	U	SMPL
K2951-16A - 1		18186	4.1778	-7470A	PPB		22 May 2019 15:45:53	U	SMPL
K2951-04LX5 - 1		-15	-0.0448	-7470A	PPB		22 May 2019 15:48:00	U	SMPL
K2951-04A - 1		17456	4.0085	-7470A	PPB		22 May 2019 15:50:08	U	SMPL
K2951-15LX5 - 1		395	0.0503	-7470A	PPB		22 May 2019 15:52:16	U	SMPL
K2951-15A - 1		20950	4.8191	-7470A	PPB		22 May 2019 15:54:23	U	SMPL
CCV78 - 1	CCV78	21462	4.9379	-7470A	PPB		22 May 2019 15:56:30	U	SMPL
CCB78 - 1	CCB78	-16	-0.0451	-7470A	PPB		22 May 2019 15:58:38	U	SMPL