



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB102924

Analyst By : JIGNESH

Supervisor Review By : apatel

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/22/2019	14:10
2	CAL2	1	Water	NA	NA	20.1	7.00	05/22/2019	14:11
3	CAL3	1	Water	NA	NA	20.1	10.03	05/22/2019	14:12
4	ICV	1	Water	NA	NA	20.1	7.01	05/22/2019	14:14
5	CCV1	1	Water	NA	NA	20.2	2.02	05/22/2019	14:15
6	K2970-18	1	Water	NA	NA	13.5	7.17	05/22/2019	14:16
7	K2994-18	1	Water	NA	NA	15.0	7.00	05/22/2019	14:18
8	K2994-18DUP	1	Water	NA	NA	15.1	7.01	05/22/2019	14:19
9	CCV2	1	Water	NA	NA	20.2	12.02	05/22/2019	14:20

75102924

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k2994

WorkList ID : 126428

Date : 05-22-2019 13:52:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/20/2019	Water	K2970-18-1A	Corrosivity	Cool 4 deg C	PSEG01	O11	TP-10	05/20/2019	9040C
05/21/2019	Water	K2984-01-1A	Corrosivity	Cool 4 deg C	PSEG01	O11	OILY-WATER	05/21/2019	9040C
05/21/2019	Water	K2994-18-1A	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-7	05/21/2019	9040C

Date/Time 0522-19 14:00
Received by: Jo
Relinquished by: Co

Date/Time 0522-19 15:00
Received by: [Signature]
Relinquished by: [Signature]