



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102929

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/22/2019	17:30
2	CAL2	1	Water	NA	NA	20.2	7.01	05/22/2019	17:31
3	CAL3	1	Water	NA	NA	20.2	10.02	05/22/2019	17:32
4	ICV	1	Water	NA	NA	20.1	6.99	05/22/2019	17:33
5	CCV1	1	Water	NA	NA	20.2	2.01	05/22/2019	17:35
6	K2647-12	1	Solid	20.03	20	21.3	8.08	05/22/2019	17:36
7	K2647-14	1	Solid	20.04	20	20.6	8.46	05/22/2019	17:37
8	K2984-03	1	Solid	20.02	20	20.1	3.97	05/22/2019	17:38
9	K2994-04	1	Solid	20.01	20	20.7	8.55	05/22/2019	17:40
10	K2994-08	1	Solid	20.03	20	21.1	7.84	05/22/2019	17:41
11	K2994-16	1	Solid	20.05	20	21.1	10.02	05/22/2019	17:43
12	K3008-07	1	Solid	20.02	20	21.8	10.44	05/22/2019	17:45
13	K3008-08	1	Solid	20.03	20	21.5	10.50	05/22/2019	17:46
14	K3008-09	1	Solid	20.05	20	21.2	10.48	05/22/2019	17:48
15	K3008-09DUP	1	Solid	20.04	20	21.3	10.49	05/22/2019	17:50
16	CCV2	1	Water	NA	NA	20.1	12.01	05/22/2019	17:53

WORKLIST(Hardcopy Internal Chain)

LB102929

WorkList Name : corrosivity k3008

WorkList ID : 126432

Date : 05-22-2019 13:55:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/22/2019	Solid	K2647-12	Corrosivity	Cool 4 deg C	PSEG05		SU-03-052219-A	05/22/2019	9045D
05/22/2019	Solid	K2647-14	Corrosivity	Cool 4 deg C	PSEG05		SU-03-052219-B	05/22/2019	9045D
05/21/2019	Solid	K2984-03	Corrosivity	Cool 4 deg C	PSEG01	O11	OILY-WATER	05/21/2019	9045D
05/21/2019	Solid	K2994-04	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-8-S	05/21/2019	9045D
05/21/2019	Solid	K2994-08	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-8-D	05/21/2019	9045D
05/21/2019	Solid	K2994-16	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-6-S	05/21/2019	9045D
05/22/2019	Solid	K3008-07	Corrosivity	Cool 4 deg C	TECT01	P63	COMP-19	05/22/2019	9045D
05/22/2019	Solid	K3008-08	Corrosivity	Cool 4 deg C	TECT01	P63	COMP-20	05/22/2019	9045D
05/22/2019	Solid	K3008-09	Corrosivity	Cool 4 deg C	TECT01	P63	COMP-21	05/22/2019	9045D

Date/Time 05/22/19 16:00
 Received by: J0
 Relinquished by: R.C

Date/Time 05/22/19 18:00
 Received by: R.C
 Relinquished by: J0