



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102933

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/22/2019	12:55
2	K2970-18		1	8	100.00	>212.0	>212.0	05/22/2019	13:20
3	K2970-18DUP		1	8	100.00	>212.0	>212.0	05/22/2019	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 102933

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP52219 WorkList ID : 126417 Date : 05-22-2019 09:17:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/27/2019	Water	K2970-18 A	Flash Point	Cool 4 deg C	PSEG01	O11	TP-10	05/20/2019	1010A

Date/Time 05/22/19 12:25
Received by: KS
Relinquished by: CP

Date/Time 05/22/19 13:55
Received by: CP
Relinquished by: KS