



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 5/24/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 15:10
In Date: 05/23/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:32
Out Date: 05/24/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB102943

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3016-01	A51C0	1	1.15	9.95	8.47	83.2
K3016-02	A51C1	2	1.12	9.64	8.51	86.7
K3016-03	A51C2	3	1.13	9.56	8.11	82.8
K3016-04	A51C3	4	1.14	9.68	8.07	81.1
K3016-05	A51C4	5	1.16	9.96	8.66	85.2
K3016-06	A51C5	6	1.18	9.52	8.03	82.1
K3016-07	A51C6	7	1.18	9.66	8.36	84.7
K3016-08	A51C7	8	1.11	9.68	8.75	89.1
K3016-09	A51C8	9	1.16	9.83	8.57	85.5
K3016-10	A51C9	10	1.18	9.79	8.35	83.3
K3016-11	A51C9MS	11	1.18	9.79	8.35	83.3
K3016-12	A51C9MSD	12	1.18	9.79	8.35	83.3
K3016-13	A51D0	13	1.14	9.93	8.12	79.4
K3016-14	A51D1	14	1.18	9.83	8.56	85.3
K3016-15	A51D2	15	1.15	9.96	8.36	81.8
K3016-16	A51D3	16	1.18	9.57	8.75	90.2
K3016-17	A51D4	17	1.16	9.98	8.92	88.0
K3016-18	A51D5	18	1.13	9.69	8.16	82.1
K3016-19	A51D6	19	1.11	9.75	7.82	77.7
K3016-20	A51D7	20	1.16	9.88	8.00	78.4
K3016-21	A51D8	21	1.16	9.54	8.16	83.5
K3016-22	A51D9	22	1.17	9.65	8.11	81.8
K3016-23	A51G5	23	1.00	2.00	2.00	100.0
K3016-24	A51G8	24	1.00	2.00	2.00	100.0
K3016-25	A51H1	25	1.00	2.00	2.00	100.0
K3016-26	A51H2	26	1.00	2.00	2.00	100.0
K3016-27	A51H5	27	1.00	2.00	2.00	100.0
K3016-28	VHBLK01	28	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$