



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102945

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/23/2019	13:15
2	CAL2	1	Water	NA	NA	20.1	7.00	05/23/2019	13:16
3	CAL3	1	Water	NA	NA	20.1	10.02	05/23/2019	13:17
4	ICV	1	Water	NA	NA	20.1	6.99	05/23/2019	13:19
5	CCV1	1	Water	NA	NA	20.1	2.02	05/23/2019	13:21
6	K3022-01	1	Water	NA	NA	15.2	6.03	05/23/2019	13:22
7	K3022-02	1	Water	NA	NA	14.0	6.42	05/23/2019	13:24
8	K3022-03	1	Water	NA	NA	14.8	6.36	05/23/2019	13:25
9	K3022-04	1	Water	NA	NA	13.0	6.40	05/23/2019	13:26
10	K3022-04DUP	1	Water	NA	NA	13.1	6.41	05/23/2019	13:27
11	CCV2	1	Water	NA	NA	20.2	12.01	05/23/2019	13:30

LB 102945

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3022

WorkList ID : 126495

Date : 05-23-2019 13:00:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/22/2019	Water	K3022-01 - 5	pH	Cool 4 deg C	LOCK01	P41	MW-1	05/22/2019	SM 4500-H B
05/22/2019	Water	K3022-02 - 5	pH	Cool 4 deg C	LOCK01	P41	MW-3	05/22/2019	SM 4500-H B
05/22/2019	Water	K3022-03 - 5	pH	Cool 4 deg C	LOCK01	P41	MW-4	05/22/2019	SM 4500-H B
05/22/2019	Water	K3022-04 - 5	pH	Cool 4 deg C	LOCK01	P41	MW-2	05/22/2019	SM 4500-H B

Date/Time 05-23-19 13:10
 Received by: JP
 Relinquished by: JP

Date/Time 05-23-19 14:15
 Received by: JP
 Relinquished by: JP