


1B102948

7470A

INSTRUMENT ID : CUI

Linear



A= 0.0000e+000

B= 2.3587e-004

C= -5.2025e-003

Rho= 0.9999331

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	67	0.000	67				
0.2	0.200	0.198	-0.002	861	0.0 %	861				
2.5	2.500	2.506	0.006	10645	0.0 %	10645				
5.0	5.000	4.938	-0.062	20956	0.0 %	20956				
7.5	7.500	7.578	0.078	32148	0.0 %	32148				
10.0	10.000	9.970	-0.030	42292	0.0 %	42292				

LB102948 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 May 2019 13:57:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0-0	67	-	0.00007470A	PPB		23 May 2019 13:59:59	S	Std
0.2 - 1	S0-2	861	-	0.20007470A	PPB		23 May 2019 14:02:07	S	Std
2.5 - 1	S2-5	10645	-	2.50007470A	PPB		23 May 2019 14:04:16	S	Std
5.0 - 1	S5-0	20956	-	5.00007470A	PPB		23 May 2019 14:06:24	S	Std
7.5 - 1	S7-5	32148	-	7.50007470A	PPB		23 May 2019 14:08:32	S	Std
10.0 - 1	S10-0	42292	-	10.00007470A	PPB		23 May 2019 14:23:03	S	Std
ICV58 - 1	ICV58	17383	4.0950	-7470A	PPB		23 May 2019 14:26:56	U	SMPL
ICB58 - 1	ICB58	-45	-0.0158	-7470A	PPB		23 May 2019 14:29:03	U	SMPL
CCV86 - 1	CCV86	21419	5.0470	-7470A	PPB		23 May 2019 14:31:10	U	SMPL
CCB86 - 1	CCB86	-41	-0.0149	-7470A	PPB		23 May 2019 14:33:15	U	SMPL
CRA - 1	CRA	817	0.1875	-7470A	PPB		23 May 2019 14:35:22	U	SMPL
HighStd - 1	HighStd	42194	9.9473	-7470A	PPB		23 May 2019 14:37:28	U	SMPL
ChkStd - 1	ChkStd	29840	7.0333	-7470A	PPB		23 May 2019 14:39:35	U	SMPL
PB120014BL - 1	PBW	-81	-0.0243	-7470A	PPB		23 May 2019 14:41:43	U	SMPL
PB120014BS - 1	LCSW	17223	4.0573	-7470A	PPB		23 May 2019 14:43:52	U	SMPL
K2642-06 - 1	OK-01-052119	-23	-0.0106	-7470A	PPB		23 May 2019 14:46:00	U	SMPL
K2642-06DUP - 1	OK-01-052119DUP	27	0.0012	-7470A	PPB		23 May 2019 14:48:07	U	SMPL
K2642-06MS - 1	OK-01-052119MS	17591	4.1441	-7470A	PPB		23 May 2019 14:50:15	U	SMPL
K2642-06MSD - 1	OK-01-052119MSD	16999	4.0044	-7470A	PPB		23 May 2019 14:52:23	U	SMPL
K2970-04 - 1	TP-11-S	-33	-0.0130	-7470A	PPB		23 May 2019 14:54:31	U	SMPL
CCV87 - 1	CCV87	21043	4.9583	-7470A	PPB		23 May 2019 14:56:38	U	SMPL
CCB87 - 1	CCB87	-50	-0.0170	-7470A	PPB		23 May 2019 14:58:44	U	SMPL
K2970-08 - 1	TP-11-D	8	-0.0033	-7470A	PPB		23 May 2019 15:00:51	U	SMPL
K2970-16 - 1	TP-9-S	50	0.0066	-7470A	PPB		23 May 2019 15:02:57	U	SMPL
K2971-18 - 1	SUP-2A-C-1-052019	34	0.0028	-7470A	PPB		23 May 2019 15:05:04	U	SMPL
K2971-19 - 1	SUP-2A-C-2-052019	37	0.0035	-7470A	PPB		23 May 2019 15:07:10	U	SMPL
K2971-20 - 1	SUP-2A-C-3-052019	56	0.0080	-7470A	PPB		23 May 2019 15:09:17	U	SMPL
K2971-21 - 1	SUP-2A-C-4-052019	24	0.0005	-7470A	PPB		23 May 2019 15:11:25	U	SMPL
K2971-22 - 1	SUP-2C-C-1-052019	36	0.0033	-7470A	PPB		23 May 2019 15:13:33	U	SMPL
K2978-03 - 1	RO-SOIL	30	0.0019	-7470A	PPB		23 May 2019 15:15:41	U	SMPL
K2978-08 - 1	TP-SOIL	22	-0.0000	-7470A	PPB		23 May 2019 15:17:50	U	SMPL
K2983-01 - 1	SOIL-GRAVEL	42	0.0047	-7470A	PPB		23 May 2019 15:19:56	U	SMPL
CCV88 - 1	CCV88	20949	4.9361	-7470A	PPB		23 May 2019 15:22:03	U	SMPL
CCB88 - 1	CCB88	-38	-0.0142	-7470A	PPB		23 May 2019 15:24:10	U	SMPL
K2983-03 - 1	OILY-DEBRIS	9	-0.0031	-7470A	PPB		23 May 2019 15:26:16	U	SMPL
K2985-01 - 1	05-21-19-A	15	-0.0017	-7470A	PPB		23 May 2019 15:28:23	U	SMPL
K2985-02 - 1	05-21-19-B	11	-0.0026	-7470A	PPB		23 May 2019 15:30:30	U	SMPL
K2985-03 - 1	05-21-19-C	40	0.0042	-7470A	PPB		23 May 2019 15:32:36	U	SMPL
PB120015BL - 1	PBW	30	0.0019	-7470A	PPB		23 May 2019 15:34:43	U	SMPL
PB120015BS - 1	LCSW	17073	4.0219	-7470A	PPB		23 May 2019 15:36:51	U	SMPL
K2968-01 - 1	EFFL-5B	3002	0.7029	-7470A	PPB		23 May 2019 15:38:58	U	SMPL
K2994-17 - 1	TP-7	1851	0.4314	-7470A	PPB		23 May 2019 15:41:06	U	SMPL
K2994-17DUP - 1	TP-7DUP	1920	0.4477	-7470A	PPB		23 May 2019 15:43:15	U	SMPL
K2994-17MS - 1	TP-7MS	21411	5.0451	-7470A	PPB		23 May 2019 15:45:22	U	SMPL
CCV89 - 1	CCV89	21318	5.0232	-7470A	PPB		23 May 2019 15:47:28	U	SMPL
CCB89 - 1	CCB89	-56	-0.0184	-7470A	PPB		23 May 2019 15:49:35	U	SMPL
K2994-17MSD - 1	TP-7MSD	21381	5.0380	-7470A	PPB		23 May 2019 15:51:41	U	SMPL
K3005-01 - 1	MW-1	0	-0.0052	-7470A	PPB		23 May 2019 15:53:48	U	SMPL
K3005-02 - 1	FB052119	19	-0.0007	-7470A	PPB		23 May 2019 15:55:55	U	SMPL
PB120014TB - 1	PB119941TB	39	0.0040	-7470A	PPB		23 May 2019 15:58:01	U	SMPL
K2642-06LX5 - 1		18	-0.0010	-7470A	PPB		23 May 2019 16:00:09	U	SMPL
K2642-06A - 1		17366	4.0910	-7470A	PPB		23 May 2019 16:02:16	U	SMPL
K2968-01LX5 - 1 K2994-17LX5		310	0.0679	-7470A	PPB		23 May 2019 16:04:24	U	SMPL
K2968-01A - 1 K2994-17A		21975	5.1781	-7470A	PPB		23 May 2019 16:06:33	U	SMPL
CCV90 - 1	CCV90	21114	4.9750	-7470A	PPB		23 May 2019 16:08:41	U	SMPL
CCB90 - 1	CCB90	-48	-0.0165	-7470A	PPB		23 May 2019 16:10:48	U	SMPL

LS 05-23-19