



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102954

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	28.00	82.4	82.4	05/23/2019	14:20
2	K2994-18		1	9	100.00	>212.0	>212.0	05/23/2019	14:45
3	K2994-18DUP		1	9	100.00	>212.0	>212.0	05/23/2019	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102954

WorkList Name : FP52319

WorkList ID : 126468

Date : 05-23-2019 09:15:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/28/2019	Water	K2994-18	C Flash Point	Cool 4 deg C	PSEG01	O21	TP-7	05/21/2019	1010A

Date/Time 05/23/19 13:45
 Received by: KS
 Relinquished by: CP

Date/Time 05/23/19 15:20
 Received by: CP
 Relinquished by: KS