



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102955

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/23/2019	12:45
2	K2984-03		1	8	100.00	>212.0	>212.0	05/23/2019	13:10
3	K2984-03DUP		1	8	100.00	>212.0	>212.0	05/23/2019	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102955

WorkList Name : FP52319-1

WorkList ID : 126486

Date : 05-23-2019 10:38:46

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/28/2019	Solid	K2984-03	B	B	Cool 4 deg C	PSEG01	O11	OILY-WATER	05/21/2019	1010A

Date/Time 05/23/19 12:15
 Received by: KS
 Relinquished by: CP

Date/Time 05/23/19 13:45
 Received by: CP
 Relinquished by: KS