

LB102959


ISM02.4

INSTRUMENT ID: CU1

Linear

 μ Abs.:

51477



A= 0.0000e+000

B= 1.9413e-004

C= 1.1566e-002

Rho= 0.9998904

Accept=Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.040	0.040	147	0.000	147					—
0.05	0.050										
0.20	0.200	0.215	0.015	1046	0.0 %	1046					8
2.50	2.500	2.392	-0.108	12263	0.0 %	12263					4
5.00	5.000	5.065	0.065	26029	0.0 %	26029					1
7.50	7.500	7.484	-0.016	38490	0.0 %	38490					0
10.0	10.000	10.005	0.005	51477	0.0 %	51477					0

RS 05.23.19

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	147	-		0 ISM02.4	PPB	5/23/2019 16:29	Std
	0.2 S0.2	1046	-		0.2 ISM02.4	PPB	5/23/2019 16:31	Std
	2.5 S2.5	12263	-		2.5 ISM02.4	PPB	5/23/2019 16:33	Std
	5 S5.0	26029	-		5 ISM02.4	PPB	5/23/2019 16:36	Std
	7.5 S7.5	38490	-		7.5 ISM02.4	PPB	5/23/2019 16:38	Std
	10 S10.0	51477	-		10 ISM02.4	PPB	5/23/2019 16:40	Std
ICV	ICV	20910	4.0708	-	ISM02.4	PPB	5/23/2019 16:44	SMPL
ICB	ICB	42	0.0197	-	ISM02.4	PPB	5/23/2019 16:47	SMPL
CCV091	CCV091	26176	5.0931	-	ISM02.4	PPB	5/23/2019 16:49	SMPL
CCB091	CCB091	40	0.0193	-	ISM02.4	PPB	5/23/2019 16:51	SMPL
PB120022BL	PBS022	103	0.0316	-	ISM02.4	PPB	5/23/2019 16:53	SMPL
K2720-01	MH0AQ3	1617	0.3255	-	ISM02.4	PPB	5/23/2019 16:56	SMPL
K2720-02	MH0AQ4	1896	0.3796	-	ISM02.4	PPB	5/23/2019 16:58	SMPL
K2720-03	MH0AQ5	1813	0.3635	-	ISM02.4	PPB	5/23/2019 17:00	SMPL
K2720-05	MH0AQ9	1592	0.3206	-	ISM02.4	PPB	5/23/2019 17:02	SMPL
K2720-06	MH0AR0	2609	0.5181	-	ISM02.4	PPB	5/23/2019 17:05	SMPL
K2720-07	MH0AR1	2423	0.4819	-	ISM02.4	PPB	5/23/2019 17:07	SMPL
K2720-08	MH0AR2	2834	0.5617	-	ISM02.4	PPB	5/23/2019 17:09	SMPL
K2720-09	MH0AR3	1628	0.3276	-	ISM02.4	PPB	5/23/2019 17:12	SMPL
K2720-10	MH0AR4	2122	0.4235	-	ISM02.4	PPB	5/23/2019 17:14	SMPL
K2720-11	MH0AR5	2404	0.4783	-	ISM02.4	PPB	5/23/2019 17:16	SMPL
K2720-12	MH0AR6	2806	0.5563	-	ISM02.4	PPB	5/23/2019 17:18	SMPL
K2720-13	MH0AR7	1423	0.2878	-	ISM02.4	PPB	5/23/2019 17:21	SMPL
K2720-14	MH0AR8	6496	1.2726	-	ISM02.4	PPB	5/23/2019 17:23	SMPL
K2720-15	MH0AR9	11787	2.2998	-	ISM02.4	PPB	5/23/2019 17:25	SMPL
K2720-16	MH0AS0	16472	3.2093	-	ISM02.4	PPB	5/23/2019 17:27	SMPL
K2720-17	MH0AS1	38830	7.5497	-	ISM02.4	PPB	5/23/2019 17:30	SMPL
K2720-18	MH0AS2	34292	6.6687	-	ISM02.4	PPB	5/23/2019 17:32	SMPL
K2720-19	MH0AS3	33317	6.4794	-	ISM02.4	PPB	5/23/2019 17:34	SMPL
K2720-20	MH0AS4	17553	3.4191	-	ISM02.4	PPB	5/23/2019 17:37	SMPL
CCV092	CCV092	26332	5.1234	-	ISM02.4	PPB	5/23/2019 17:39	SMPL
CCB092	CCB092	1	0.0118	-	ISM02.4	PPB	5/23/2019 17:41	SMPL
K2720-21	MH0AS4D	17757	3.4587	-	ISM02.4	PPB	5/23/2019 17:43	SMPL
K2720-22	MH0AS4S	29660	5.7695	-	ISM02.4	PPB	5/23/2019 17:46	SMPL
PB120023BL	PBS023	-7	0.0102	-	ISM02.4	PPB	5/23/2019 17:48	SMPL
K2787-19	MX2G93	40959	7.963	-	ISM02.4	PPB	5/23/2019 17:50	SMPL
K2787-20	MX2G93D	37284	7.2495	-	ISM02.4	PPB	5/23/2019 17:52	SMPL
K2787-21	MX2G93S	52649	10.2324	-	ISM02.4	PPB	5/23/2019 17:55	SMPL
PB120024BL	PBS024	-67	-0.0014	-	ISM02.4	PPB	5/23/2019 17:57	SMPL
K2812-01	MA4246	21926	4.2681	-	ISM02.4	PPB	5/23/2019 17:59	SMPL
K2812-02	MA4246D	24095	4.6891	-	ISM02.4	PPB	5/23/2019 18:02	SMPL
K2812-03	MA4246S	37408	7.2736	-	ISM02.4	PPB	5/23/2019 18:04	SMPL
PB120025BL	PBS025	-2	0.0112	-	ISM02.4	PPB	5/23/2019 18:06	SMPL
K2833-01	MA5174	169	0.0444	-	ISM02.4	PPB	5/23/2019 18:08	SMPL
K2833-02	MA5175	960	0.1979	-	ISM02.4	PPB	5/23/2019 18:11	SMPL
K2833-03	MA5180	140	0.0387	-	ISM02.4	PPB	5/23/2019 18:13	SMPL

INSTRUMENT ID : CV1

K2833-04	MA5189	106	0.0321 -	ISM02.4	PPB	5/23/2019 18:15	SMPL
K2833-05	MA5193	108	0.0325 -	ISM02.4	PPB	5/23/2019 18:17	SMPL
K2833-06	MA5198	88	0.0286 -	ISM02.4	PPB	5/23/2019 18:20	SMPL
K2833-07	MA51A5	106	0.0321 -	ISM02.4	PPB	5/23/2019 18:22	SMPL
K2833-08	MA51B0	87	0.0285 -	ISM02.4	PPB	5/23/2019 18:24	SMPL
K2833-09	MA51B0D	78	0.0267 -	ISM02.4	PPB	5/23/2019 18:27	SMPL
CCV093	CCV093	24688	4.8043 -	ISM02.4	PPB	5/23/2019 18:29	SMPL
CCB093	CCB093	-7	0.0102 -	ISM02.4	PPB	5/23/2019 18:31	SMPL
K2833-10	MA51B0S	15198	2.962 -	ISM02.4	PPB	5/23/2019 18:33	SMPL
PB120026BL	PBS026	61	0.0234 -	ISM02.4	PPB	5/23/2019 18:36	SMPL
K2930-01	MH0AA1	21299	4.1464 -	ISM02.4	PPB	5/23/2019 18:38	SMPL
K2930-02	MH0AA2	21137	4.1149 -	ISM02.4	PPB	5/23/2019 18:40	SMPL
K2930-03	MHA001	2689	0.5336 -	ISM02.4	PPB	5/23/2019 18:42	SMPL
K2930-04	MHA001D	2687	0.5332 -	ISM02.4	PPB	5/23/2019 18:45	SMPL
K2930-05	MHA001S	20403	3.9724 -	ISM02.4	PPB	5/23/2019 18:47	SMPL
K2930-06	MHA002	941	0.1942 -	ISM02.4	PPB	5/23/2019 18:49	SMPL
K2930-07	MHA003	713	0.15 -	ISM02.4	PPB	5/23/2019 18:52	SMPL
K2930-08	MHA004	8148	1.5933 -	ISM02.4	PPB	5/23/2019 18:54	SMPL
K2930-09	MHA005	3407	0.673 -	ISM02.4	PPB	5/23/2019 18:56	SMPL
K2930-10	MHA006	583	0.1247 -	ISM02.4	PPB	5/23/2019 18:58	SMPL
K2930-11	MHA015	3546	0.7 -	ISM02.4	PPB	5/23/2019 19:01	SMPL
K2930-12	MHA016	30767	5.9844 -	ISM02.4	PPB	5/23/2019 19:03	SMPL
K2930-13	MHA017	5137	1.0088 -	ISM02.4	PPB	5/23/2019 19:05	SMPL
K2930-14	MHA018	45521	8.8486 -	ISM02.4	PPB	5/23/2019 19:07	SMPL
K2930-15	MHA019	4471	0.8795 -	ISM02.4	PPB	5/23/2019 19:10	SMPL
K2930-16	MHA020	712	0.1498 -	ISM02.4	PPB	5/23/2019 19:12	SMPL
K2930-17	MHA021	12437	2.426 -	ISM02.4	PPB	5/23/2019 19:14	SMPL
K2930-18	MHA022	2605	0.5173 -	ISM02.4	PPB	5/23/2019 19:17	SMPL
CCV094	CCV094	24700	4.8066 -	ISM02.4	PPB	5/23/2019 19:19	SMPL
CCB094	CCB094	-6	0.0104 -	ISM02.4	PPB	5/23/2019 19:21	SMPL
K2930-19	MHA007	1864	0.3734 -	ISM02.4	PPB	5/23/2019 19:24	SMPL
K2930-20	MHA008	4238	0.8343 -	ISM02.4	PPB	5/23/2019 19:26	SMPL
K2930-21	MHA009	1330	0.2698 -	ISM02.4	PPB	5/23/2019 19:28	SMPL
K2930-22	MHA010	2457	0.4885 -	ISM02.4	PPB	5/23/2019 19:30	SMPL
K2787-19DLX2		20061	3.906 -	ISM02.4	PPB	5/23/2019 19:33	SMPL
K2787-20DLX2		18265	3.5574 -	ISM02.4	PPB	5/23/2019 19:35	SMPL
K2787-21DLX2		26065	5.0716 -	ISM02.4	PPB	5/23/2019 19:37	SMPL
CCV095	CCV095	26067	5.072 -	ISM02.4	PPB	5/23/2019 19:40	SMPL
CCB095	CCB095	42	0.0197 -	ISM02.4	PPB	5/23/2019 19:42	SMPL