



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102961

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/23/2019	17:20
2	CAL2	1	Water	NA	NA	20.2	6.99	05/23/2019	17:21
3	CAL3	1	Water	NA	NA	20.2	10.03	05/23/2019	17:22
4	ICV	1	Water	NA	NA	20.2	7.01	05/23/2019	17:24
5	CCV1	1	Water	NA	NA	20.2	2.01	05/23/2019	17:25
6	K2643-10	1	Solid	20.02	20	22.3	8.68	05/23/2019	17:26
7	K2643-10DUP	1	Solid	20.03	20	22.4	8.69	05/23/2019	17:27
8	K2643-12	1	Solid	20.04	20	22.3	8.86	05/23/2019	17:28
9	K3019-03	1	Solid	20.02	20	21.7	5.56	05/23/2019	17:30
10	K3019-06	1	Solid	20.01	20	21.4	5.58	05/23/2019	17:32
11	K3019-09	1	Solid	20.04	20	21.5	6.83	05/23/2019	17:34
12	K3023-01	1	Solid	20.02	20	21.4	9.92	05/23/2019	17:35
13	K3026-04	1	Solid	20.03	20	21.0	8.40	05/23/2019	17:36
14	K3026-08	1	Solid	20.04	20	21.4	8.02	05/23/2019	17:38
15	K3026-16	1	Solid	20.05	20	21.2	8.72	05/23/2019	17:40
16	CCV2	1	Water	NA	NA	20.1	12.02	05/23/2019	17:42
17	K3035-01	1	Solid	20.02	20	21.3	6.75	05/23/2019	17:43
18	CCV3	1	Water	NA	NA	20.1	2.02	05/23/2019	17:45

WORKLIST(Hardcopy Internal Chain)

LB 102961

WorkList Name : corrosivity k3026

WorkList ID : 126505

Date : 05-23-2019 14:39:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/23/2019	Solid	K2643-10	Corrosivity	Cool 4 deg C	PSEG05		OR-03-052319-A	05/23/2019	9045D
05/23/2019	Solid	K2643-12	Corrosivity	Cool 4 deg C	PSEG05		OR-03-052319-B	05/23/2019	9045D
05/23/2019	Solid	K3019-03	Corrosivity	Cool 4 deg C	PSEG01	K21	RT-3422	05/23/2019	9045D
05/23/2019	Solid	K3019-06	Corrosivity	Cool 4 deg C	PSEG01	K21	RBR200036	05/23/2019	9045D
05/23/2019	Solid	K3019-09	Corrosivity	Cool 4 deg C	PSEG01	K21	RT-1465	05/23/2019	9045D
05/23/2019	Solid	K3023-01	Corrosivity	Cool 4 deg C	USAR03	P41	SSSI-0523-S-TCLP-11	05/23/2019	9045D
05/22/2019	Solid	K3026-04	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-5-S	05/22/2019	9045D
05/22/2019	Solid	K3026-08	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-5-D	05/22/2019	9045D
05/22/2019	Solid	K3026-16	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-3-S	05/22/2019	9045D
05/21/2019	Solid	K3035-01	Corrosivity	Cool 4 deg C	CLEA11	P41	OIL-WATER-SEPERATOR(OV	05/21/2019	9045D

Date/Time 05-23-19 16:00

Received by: Jn

Relinquished by: ca

Date/Time 05-23-19 16:00

Received by: ca

Relinquished by: Jn