



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102970

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP75010
Residual chlorine std, 0PPM	WP75006
Residual chlorine std, 0.1PPM	WP75007
Residual chlorine std, 0.2PPM	WP75008
Residual chlorine std, and CCV , 0.4PPM	WP75009
Residual chlorine std, 1.6PPM	WP75011
Residual chlorine ICV-LCS, 0.4PPM	WP75012

Intercept: 0.0000

Slope: 1.0000

Regression: 1.000000

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	05/24/2019	09:26
2	CAL2	0.1	1	0.000	0.100	0.100	05/24/2019	09:27
3	CAL3	0.2	1	0.000	0.200	0.200	05/24/2019	09:28
4	CAL4	0.4	1	0.000	0.400	0.400	05/24/2019	09:29
5	CAL5	0.8	1	0.000	0.800	0.800	05/24/2019	09:30
6	CAL6	1.6	1	0.000	1.600	1.600	05/24/2019	09:31



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102970

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3900	05/24/2019	09:32
2	ICB				1	0.0000	0.0000	0.0000	0.0000	05/24/2019	09:33
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.4000	05/24/2019	09:34
4	CCB1				1	0.0000	0.0000	0.0000	0.0000	05/24/2019	09:35
5	LB102970BL	50	50		1	0.0000	0.0000	0.0000	0.0000	05/24/2019	09:36
6	LB102970BS	50	50	0.4	1	0.0000	0.3900	0.3900	0.3900	05/24/2019	09:37
7	K3015-01	50	50		1	0.0000	0.0000	0.0000	0.0000	05/24/2019	09:38
8	K3015-01DUP	50	50		1	0.0000	0.0000	0.0000	0.0000	05/24/2019	09:39
9	K3015-01MS	50	50	0.4	1	0.0000	0.4200	0.4200	0.4200	05/24/2019	09:40
10	K3015-01MSD	50	50	0.4	1	0.0000	0.4500	0.4500	0.4500	05/24/2019	09:41
11	CCV2			0.4	1	0.0000	0.3900	0.3900	0.3900	05/24/2019	09:42
12	CCB2				1	0.0000	0.0000	0.0000	0.0000	05/24/2019	09:43

CB 102976

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RCL052419

WorkList ID : 126534

Date : 05-24-2019 08:59:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/22/2019	Water	K3015-01	Residual Chlorine	Cool 4 deg C	SHAW22	P41	OUTFALL001-WK-4	05/22/2019	SM4500 Cl G

Date/Time 05/24/19 9:05
Received by: AG
Relinquished by: AG

Date/Time 05/24/19 10:00
Received by: AG
Relinquished by: AG