


LB102987

7471B

INSTRUMENT ID: CV1

 μ Abs.:

60453



Linear

A= 0.0000e+000

B= 1.6232e-004

C= -1.7471e-002

Rho= 0.9992722

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	88	0.000	88				
0.2	0.200	0.193	-0.007	1297	0.0 %	1297				
2.5	2.500	2.379	-0.121	14762	0.0 %	14762				
5.0	5.000	5.115	0.115	31621	0.0 %	31621				
7.5	7.500	7.721	0.221	47675	0.0 %	47675				
10.0	10.000	9.795	-0.205	60453	0.0 %	60453				

LB102987 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 May 2019 13:04:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	88	-	0.0000	7471B	PPB	24 May 2019 13:12:53	S	Std
0.2 - 1	50.2	1297	-	0.2000	7471B	PPB	24 May 2019 13:15:11	S	Std
2.5 - 1	52.5	14762	-	2.5000	7471B	PPB	24 May 2019 13:17:28	S	Std
5.0 - 1	55.0	31621	-	5.0000	7471B	PPB	24 May 2019 13:19:45	S	Std
7.5 - 1	57.5	47675	-	7.5000	7471B	PPB	24 May 2019 13:22:02	S	Std
10.0 - 1	510.0	60453	-	10.0000	7471B	PPB	24 May 2019 13:31:09	S	Std
ICV60 - 1	ICV60	24438	3.9493	-7471B	PPB	24 May 2019 13:35:03	U	SMPL	
ICB60 - 1	ICB60	17	-0.0147	-7471B	PPB	24 May 2019 13:37:20	U	SMPL	
CCV96 - 1	CCV96	31062	5.0245	-7471B	PPB	24 May 2019 13:39:35	U	SMPL	
CCB96 - 1	CCB96	-28	-0.0220	-7471B	PPB	24 May 2019 13:41:50	U	SMPL	
CRA - 1	CRA	1280	0.1903	-7471B	PPB	24 May 2019 13:47:50	U	SMPL	
HighStd - 1	HighStd	58499	9.4780	-7471B	PPB	24 May 2019 13:50:05	U	SMPL	
ChkStd - 1	ChkStd	43450	7.0353	-7471B	PPB	24 May 2019 13:52:21	U	SMPL	
PB120060BL - 1	PBS	-35	-0.0232	-7471B	PPB	24 May 2019 13:54:38	U	SMPL	
PB120060BS - 1	LCSS	22741	3.6738	-7471B	PPB	24 May 2019 13:56:56	U	SMPL	
K2643-09 - 1	OR-03-052319-A	4527	0.7173	-7471B	PPB	24 May 2019 13:59:13	U	SMPL	
K2643-09DUP - 1	OR-03-052319-ADUP	4511	0.7147	-7471B	PPB	24 May 2019 14:01:30	U	SMPL	
K2643-09MS - 1	OR-03-052319-AMS	25719	4.1572	-7471B	PPB	24 May 2019 14:03:46	U	SMPL	
K2643-09MSD - 1	OR-03-052319-AMSD	25965	4.1971	-7471B	PPB	24 May 2019 14:06:04	U	SMPL	
K2643-11 - 1	OR-03-052319-B	3414	0.5367	-7471B	PPB	24 May 2019 14:08:20	U	SMPL	
CCV97 - 1	CCV97	31181	5.0438	-7471B	PPB	24 May 2019 14:10:37	U	SMPL	
CCB97 - 1	CCB97	10	-0.0158	-7471B	PPB	24 May 2019 14:12:52	U	SMPL	
K3019-01 - 1	RT-3422	1262	0.1874	-7471B	PPB	24 May 2019 14:15:09	U	SMPL	
K3019-04 - 1	RBR200036	1720	0.2617	-7471B	PPB	24 May 2019 14:17:24	U	SMPL	
K3019-07 - 1	RT-1465	2090	0.3218	-7471B	PPB	24 May 2019 14:19:40	U	SMPL	
K3026-03 - 1	TP-5-S	9719	1.5601	-7471B	PPB	24 May 2019 14:21:56	U	SMPL	
K3026-07 - 1	TP-5-D	51512	8.3439	-7471B	PPB	24 May 2019 14:24:12	U	SMPL	
K3026-15 - 1	TP-3-S	28646	4.6323	-7471B	PPB	24 May 2019 14:26:28	U	SMPL	
K3036-01 - 1	FESB-26(2.5-3)	357	0.0405	-7471B	PPB	24 May 2019 14:28:46	U	SMPL	
K3036-02 - 1	FEQC-DUP-052319	945	0.1359	-7471B	PPB	24 May 2019 14:31:04	U	SMPL	
K2643-09LX5 - 1		851	0.1207	-7471B	PPB	24 May 2019 14:33:21	U	SMPL	
K2643-09A - 1		26640	4.3067	-7471B	PPB	24 May 2019 14:35:36	U	SMPL	
CCV98 - 1	CCV98	31810	5.1459	-7471B	PPB	24 May 2019 14:37:52	U	SMPL	
CCB98 - 1	CCB98	-9	-0.0189	-7471B	PPB	24 May 2019 14:40:08	U	SMPL	