



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102999

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/24/2019	18:00
2	CAL2	1	Water	NA	NA	20.1	6.99	05/24/2019	18:01
3	CAL3	1	Water	NA	NA	20.1	10.02	05/24/2019	18:02
4	ICV	1	Water	NA	NA	20.2	7.00	05/24/2019	18:04
5	CCV1	1	Water	NA	NA	20.1	2.01	05/24/2019	18:05
6	K3041-08	1	Solid	20.02	20	20.9	8.30	05/24/2019	18:06
7	K3041-08DUP	1	Solid	20.03	20	20.8	8.31	05/24/2019	18:08
8	K3041-12	1	Solid	20.01	20	20.2	8.03	05/24/2019	18:10
9	K3053-03	1	Solid	20.03	20	21.8	7.21	05/24/2019	18:11
10	K3065-13	1	Solid	20.04	20	21.4	11.20	05/24/2019	18:12
11	K3065-14	1	Solid	20.02	20	21.5	10.06	05/24/2019	18:14
12	K3065-15	1	Solid	20.03	20	20.9	11.12	05/24/2019	18:15
13	K3065-16	1	Solid	20.04	20	21.2	7.51	05/24/2019	18:16
14	K3065-17	1	Solid	20.05	20	20.4	9.14	05/24/2019	18:19
15	K3065-18	1	Solid	20.02	20	20.8	8.00	05/24/2019	18:20
16	CCV2	1	Water	NA	NA	20.1	12.01	05/24/2019	18:22

20102999

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3041

WorkList ID : 126571

Date : 05-24-2019 16:29:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/23/2019	Solid	K3041-08	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-2-S	05/23/2019	9045D
05/23/2019	Solid	K3041-12	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-2-D	05/23/2019	9045D
05/24/2019	Solid	K3053-03	Corrosivity	Cool 4 deg C	PSEG01	O31	CR-SOIL-PILE	05/24/2019	9045D
05/21/2019	Solid	K3065-13	Corrosivity	Cool 4 deg C	LIRO01	P51	B-16-COMP	05/21/2019	9045D
05/22/2019	Solid	K3065-14	Corrosivity	Cool 4 deg C	LIRO01	P51	B-15-COMP	05/22/2019	9045D
05/23/2019	Solid	K3065-15	Corrosivity	Cool 4 deg C	LIRO01	P51	B-14-COMP	05/23/2019	9045D
05/23/2019	Solid	K3065-16	Corrosivity	Cool 4 deg C	LIRO01	P51	B-14-11-12.0	05/23/2019	9045D
05/24/2019	Solid	K3065-17	Corrosivity	Cool 4 deg C	LIRO01	P51	B-13-COMP	05/24/2019	9045D
05/24/2019	Solid	K3065-18	Corrosivity	Cool 4 deg C	LIRO01	P51	B-12-COMP	05/24/2019	9045D

Date/Time 05-24-19 16:30

Received by: for

Relinquished by: for

Date/Time 05-24-19 16:35

Received by: an

Relinquished by: for