



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103030

Ambient Barometric Pressure (mmHg): 755.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	05/28/2019	12:25
2	K3041-14		1	9	100.00	>212.0	>212.0	05/28/2019	12:50
3	K3041-14DUP		1	9	100.00	>212.0	>212.0	05/28/2019	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB103030

WorkList Name : FP52819

WorkList ID : 126617

Date : 05-28-2019 09:15:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/30/2019	Water	K3041-14	C Flash Point	Cool 4 deg C	PSEG01	O21	TP-2	05/23/2019	1010A

Date/Time 05/28/19 12:00
 Received by: KS
 Relinquished by: JP

Date/Time 05/28/19 13:25
 Received by: JP
 Relinquished by: KS