



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 5/29/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 05/28/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 05/29/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103039

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2433-01	FK-BPM-04-002-ABCD	1	1.14	9.54	8.67	89.6
K2433-02	FK-BPM-04-002-ABCD	2	1.14	9.51	8.73	90.7
K3054-01	CULVER-CBTC-WC1B-052419	3	1.12	9.81	8.27	82.3
K3054-02	CULVER-CBTC-WC1A-052419	4	1.12	9.81	8.27	82.3
K3054-03	CULVER-CBTC-WC2A-052419	5	1.12	9.81	8.27	82.3
K3054-04	CULVER-CBTC-WC3A-052419	6	1.12	9.81	8.27	82.3
K3054-05	CULVER-CBTC-WC4A-052419	7	1.12	9.81	8.27	82.3
K3054-06	CULVER-CBTC-WC5A-052419	8	1.12	9.81	8.27	82.3
K3054-08	CULVER-CBTC-WC1B-052419	9	1.12	9.81	8.27	82.3
K3072-01	TP-1	10	1.15	9.64	6.64	64.7
K3072-02	TP-2	11	1.13	9.72	8.31	83.6
K3072-03	TP-3	12	1.13	9.6	7.91	80.0
K3072-04	TP-4	13	1.14	9.91	8.55	84.5
K3072-05	TP-5	14	1.13	9.98	6.57	61.5
K3072-06	TP-6	15	1.16	9.54	6.87	68.1
K3072-07	TP-7	16	1.14	9.97	6.75	63.5
K3072-08	TP-8	17	1.16	9.74	4.53	39.3
K3074-01	TP-3A-S1	18	1.15	9.79	7.4	72.3
K3074-02	TP-3A-S2	19	1.16	9.63	6.82	66.8
K3074-03	TP-3B-S1	20	1.13	9.71	6.39	61.3
K3074-04	TP-3B-S2	21	1.15	9.94	6.58	61.8
K3074-05	TP-3B-S	22	1.12	9.93	5.91	54.4
K3074-07	TP-3B-D1	23	1.15	9.82	8.47	84.4
K3074-08	TP-3B-D2	24	1.16	9.98	6.69	62.7
K3074-09	TP-3B-D	25	1.17	9.76	9.18	93.2
K3075-01	TEST-PIT-1	26	1.15	9.91	6.88	65.4
K3075-02	TEST-PIT-1	27	1.15	9.91	6.88	65.4
K3075-03	TEST-PIT-2	28	1.17	9.96	7.79	75.3
K3075-04	TEST-PIT-2	29	1.17	9.96	7.79	75.3
K3075-05	TEST-PIT-3	30	1.17	9.94	4.97	43.3
K3075-06	TEST-PIT-3	31	1.17	9.94	4.97	43.3
K3075-07	TEST-PIT-4	32	1.18	9.96	5.92	54.0
K3075-08	TEST-PIT-4	33	1.18	9.96	5.92	54.0
K3075-09	TEST-PIT-5	34	1.11	9.62	6.93	68.4
K3075-10	TEST-PIT-5	35	1.11	9.62	6.93	68.4
K3076-01	FESB-07 (5-5.5)	36	1.13	9.82	6.67	63.8
K3076-02	FESB-08 (6.5-7)	37	1.19	9.98	7.23	68.7

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$