



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103040

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/28/2019	17:05
2	CAL2	1	Water	NA	NA	20.1	6.99	05/28/2019	17:06
3	CAL3	1	Water	NA	NA	20.1	10.02	05/28/2019	17:07
4	ICV	1	Water	NA	NA	20.1	7.01	05/28/2019	17:08
5	CCV1	1	Water	NA	NA	20.1	2.02	05/28/2019	17:10
6	K3054-08	1	Solid	20.02	20	21.3	7.45	05/28/2019	17:11
7	K3054-08DUP	1	Solid	20.03	20	21.4	7.46	05/28/2019	17:12
8	K3074-06	1	Solid	20.05	20	22.1	7.41	05/28/2019	17:15
9	K3074-10	1	Solid	20.03	20	22.1	6.71	05/28/2019	17:16
10	CCV2	1	Water	NA	NA	20.1	12.02	05/28/2019	17:20

LB 103060

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3074

WorkList ID : 126646

Date : 05-28-2019 13:39:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/24/2019	Solid	K3054-08	Corrosivity	Cool 4 deg C	DIST01	P11	CULVER-CBTC-WC1A-05241	05/24/2019	9045D
05/28/2019	Solid	K3074-06	Corrosivity	Cool 4 deg C	PSEG01	O31	TP-3B-S	05/28/2019	9045D
05/28/2019	Solid	K3074-10	Corrosivity	Cool 4 deg C	PSEG01	O31	TP-3B-D	05/28/2019	9045D

Date/Time 05-27-19 16:10
 Received by: fn
 Relinquished by: ca

Date/Time 05-27-19 16:25
 Received by: ca
 Relinquished by: Jg