

LB103049


7471B

INSTRUMENT ID: CUI

Linear

 μ Abs.:

62775



A= 0.0000e+000

B= 1.5822e-004

C= -3.1430e-002

Rho= 0.9998472

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.024	-0.024	47	0.000	47				
0.2	0.200	0.165	-0.035	1243	0.0 %	1243				
2.5	2.500	2.535	0.035	16221	0.0 %	16221				
5.0	5.000	5.014	0.014	31890	0.0 %	31890				
7.5	7.500	7.609	0.109	48291	0.0 %	48291				
10.0	10.000	9.901	-0.099	62775	0.0 %	62775				

LB103049 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 May 2019 08:50:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50.0	47	-	0.00007471B	PPB	29 May 2019 08:52:09	S	Std
0.2 - 1	50.2	1243	-	0.20007471B	PPB	29 May 2019 08:54:27	S	Std
2.5 - 1	52.5	16221	-	2.50007471B	PPB	29 May 2019 08:56:44	S	Std
5.0 - 1	55.0	31890	-	5.00007471B	PPB	29 May 2019 08:59:02	S	Std
7.5 - 1	57.5	48291	-	7.50007471B	PPB	29 May 2019 09:01:19	S	Std
10.0 - 1	510.0	62775	-	10.00007471B	PPB	29 May 2019 09:03:36	S	Std
ICV63 - 1	ICV63	25282	3.9686	-7471B	PPB	29 May 2019 09:07:51	U	SMPL
ICB63 - 1	ICB63	-20	-0.0346	-7471B	PPB	29 May 2019 09:10:06	U	SMPL
CCV10 - 1	CCV10	32027	5.0358	-7471B	PPB	29 May 2019 09:12:22	U	SMPL
CCB10 - 1	CCB10	-29	-0.0360	-7471B	PPB	29 May 2019 09:14:37	U	SMPL
CRA - 1	CRA	1243	0.1652	-7471B	PPB	29 May 2019 09:16:54	U	SMPL
HighStd - 1	HighStd	62435	9.8468	-7471B	PPB	29 May 2019 09:19:10	U	SMPL
ChkStd - 1	ChkStd	45225	7.1239	-7471B	PPB	29 May 2019 09:21:25	U	SMPL
PB120138BL - 1	PBS	-56	-0.0403	-7471B	PPB	29 May 2019 09:23:43	U	SMPL
PB120138BS - 1	LCSS	23588	3.7006	-7471B	PPB	29 May 2019 09:26:01	U	SMPL
K3068-01 - 1	B-1N(3-4)	46289	7.2923	-7471B	PPB	29 May 2019 09:28:17	U	SMPL
K3068-01DUP - 1	B-1N(3-4)DUP	47357	7.4612	-7471B	PPB	29 May 2019 09:30:34	U	SMPL
K3068-01MS - 1	B-1N(3-4)MS	62431	9.8462	-7471B	PPB	29 May 2019 09:32:53	U	SMPL
K3068-01MSD - 1	B-1N(3-4)MSD	64354	10.1505	-7471B	PPB	29 May 2019 09:35:12	U	SMPL
K3068-02 - 1	B-1E(2.5-3.5)	118073	18.6497	-7471B	PPB	29 May 2019 09:37:30	U	SMPL
CCV11 - 1	CCV11	31969	5.0266	-7471B	PPB	29 May 2019 09:39:48	U	SMPL
CCB11 - 1	CCB11	-127	-0.0515	-7471B	PPB	29 May 2019 09:42:06	U	SMPL
K3068-03 - 1	B-1S(2.5-3.5)	31130	4.8939	-7471B	PPB	29 May 2019 09:44:23	U	SMPL
K3068-04 - 1	B-1W(4-5)	101266	15.9906	-7471B	PPB	29 May 2019 09:46:39	U	SMPL
K3068-05 - 1	B-3SE(4-5)	358881	HIGH	-7471B	PPB	29 May 2019 09:48:55	X	
K3068-06 - 1	B-3SW(3-4)	99886	15.7722	-7471B	PPB	29 May 2019 09:50:47	U	SMPL
K3068-07 - 1	B-4S(4-5)	101432	16.0168	-7471B	PPB	29 May 2019 09:53:07	U	SMPL
K3068-08 - 1	B-4W(3-4)	4871	0.7392	-7471B	PPB	29 May 2019 09:55:26	U	SMPL
K3068-09 - 1	B-6N(8.5-9.5)	59512	9.3844	-7471B	PPB	29 May 2019 09:57:46	U	SMPL
K3068-10 - 1	B-6E(7-8)	200715	31.7251	-7471B	PPB	29 May 2019 10:00:03	U	SMPL
K3068-11 - 1	B-6S(7-8)	98365	15.5316	-7471B	PPB	29 May 2019 10:02:22	U	SMPL
K3068-12 - 1	B-6W(8-9)	15797	2.4679	-7471B	PPB	29 May 2019 10:04:42	U	SMPL
CCV12 - 1	CCV12	32208	5.0644	-7471B	PPB	29 May 2019 10:07:00	U	SMPL
CCB12 - 1	CCB12	-71	-0.0427	-7471B	PPB	29 May 2019 10:09:16	U	SMPL
K3068-13 - 1	B-7N(3-4)	9374	1.4517	-7471B	PPB	29 May 2019 10:11:33	U	SMPL
K3068-14 - 1	B-7N(17-18)	228	0.0046	-7471B	PPB	29 May 2019 10:13:49	U	SMPL
K3068-15 - 1	B-7N(18-19)	315	0.0184	-7471B	PPB	29 May 2019 10:16:04	U	SMPL
K3068-16 - 1	B-7E(3-4)	2382	0.3454	-7471B	PPB	29 May 2019 10:18:20	U	SMPL
K3068-17 - 1	B-7E(17-18)	329	0.0206	-7471B	PPB	29 May 2019 10:20:37	U	SMPL
K3068-18 - 1	B-7E(19-20)	481	0.0447	-7471B	PPB	29 May 2019 10:22:53	U	SMPL
K3068-19 - 1	B-8N(3.5-4.5)	11819	1.8385	-7471B	PPB	29 May 2019 10:25:11	U	SMPL
K3068-20 - 1	B-8W(3.5-4.5)	163161	25.7834	-7471B	PPB	29 May 2019 10:27:28	U	SMPL
PB120139BL - 1	PBS	-330	-0.0836	-7471B	PPB	29 May 2019 10:29:46	U	SMPL
PB120139BS - 1	LCSS	24118	3.7844	-7471B	PPB	29 May 2019 10:32:05	U	SMPL
CCV13 - 1	CCV13	33403	5.2535	-7471B	PPB	29 May 2019 10:34:23	U	SMPL
CCB13 - 1	CCB13	-71	-0.0427	-7471B	PPB	29 May 2019 10:36:41	U	SMPL
K3041-07 - 1	TP-2-S	17864	2.7950	-7471B	PPB	29 May 2019 10:38:59	U	SMPL
K3041-11 - 1	TP-2-D	17547	2.7448	-7471B	PPB	29 May 2019 10:41:14	U	SMPL
K3047-01 - 1	SOIL-COMP	1556	0.2148	-7471B	PPB	29 May 2019 10:43:29	U	SMPL
K3051-02 - 1	SP-1-A	7527	1.1595	-7471B	PPB	29 May 2019 10:45:48	U	SMPL
K3051-04 - 1	SP-1-B	1754	0.2461	-7471B	PPB	29 May 2019 10:48:06	U	SMPL
K3052-01 - 1	FK-SF-03-007-A	7042	1.0827	-7471B	PPB	29 May 2019 10:50:23	U	SMPL
K3052-03 - 1	FK-SF-03-007-B	5724	0.8742	-7471B	PPB	29 May 2019 10:52:40	U	SMPL
K3053-01 - 1	CR-SOIL-PILE	1042	0.1334	-7471B	PPB	29 May 2019 10:54:57	U	SMPL
K3054-01 - 1	CULVER-CBTC-WC1B-052419	2066	0.2954	-7471B	PPB	29 May 2019 10:57:14	U	SMPL
K3055-01 - 1	COMP-1	3064	0.4533	-7471B	PPB	29 May 2019 10:59:30	U	SMPL
CCV14 - 1	CCV14	33422	5.2565	-7471B	PPB	29 May 2019 11:01:46	U	SMPL
CCB14 - 1	CCB14	-41	-0.0379	-7471B	PPB	29 May 2019 11:04:02	U	SMPL
K3065-02 - 1	B-16-COMP	6108	0.9350	-7471B	PPB	29 May 2019 11:06:19	U	SMPL
K3065-04 - 1	B-15-COMP	6538	1.0030	-7471B	PPB	29 May 2019 11:08:36	U	SMPL
K3065-06 - 1	B-14-COMP	27091	4.2548	-7471B	PPB	29 May 2019 11:10:52	U	SMPL
K3065-08 - 1	B-14-11-12.0	2960	0.4369	-7471B	PPB	29 May 2019 11:13:08	U	SMPL
K3065-10 - 1	B-13-COMP	448	0.0395	-7471B	PPB	29 May 2019 11:15:25	U	SMPL
K3065-12 - 1	B-12-COMP	558	0.0569	-7471B	PPB	29 May 2019 11:17:42	U	SMPL
K3066-01 - 1	FESB-27(3.5-4)	263	0.0102	-7471B	PPB	29 May 2019 11:19:59	U	SMPL
K3066-01DUP - 1	FESB-27(3.5-4)DUP	267	0.0108	-7471B	PPB	29 May 2019 11:22:16	U	SMPL
K3066-02 - 1	FESB-25(5.5-6)	568	0.0584	-7471B	PPB	29 May 2019 11:24:34	U	SMPL
K3066-03 - 1	K3066-01MS	25332	3.9765	-7471B	PPB	29 May 2019 11:26:51	U	SMPL
CCV15 - 1	CCV15	33728	5.3049	-7471B	PPB	29 May 2019 11:29:08	U	SMPL
CCB15 - 1	CCB15	-22	-0.0349	-7471B	PPB	29 May 2019 11:31:24	U	SMPL
K3066-04 - 1	K3066-01MSD	25755	4.0434	-7471B	PPB	29 May 2019 11:33:41	U	SMPL
K3068-01LX5 - 1		9597	1.4870	-7471B	PPB	29 May 2019 11:35:57	U	SMPL
K3068-01A - 1		63419	10.0025	-7471B	PPB	29 May 2019 11:38:14	U	SMPL
K3066-01LX5 - 1		-105	-0.0480	-7471B	PPB	29 May 2019 11:40:30	U	SMPL
K3066-01A - 1		24724	3.8803	-7471B	PPB	29 May 2019 11:42:48	U	SMPL
K3068-01DLX2 - 1		23080	3.6202	-7471B	PPB	29 May 2019 11:45:05	U	SMPL
K3068-01DUPDLX2 - 1		23431	3.6757	-7471B	PPB	29 May 2019 11:47:22	U	SMPL
K3068-01MSDLX2 - 1		30538	4.8002	-7471B	PPB	29 May 2019 11:49:39	U	SMPL
K3068-01MSDDLX2 - 1		31236	4.9106	-7471B	PPB	29 May 2019 11:51:57	U	SMPL
K3068-01ADLX2 - 1		31135	4.8947	-7471B	PPB	29 May 2019 11:54:14	U	SMPL
CCV16 - 1	CCV16	30497	4.7937	-7471B	PPB	29 May 2019 11:56:30	U	SMPL
CCB16 - 1	CCB16	-54	-0.0400	-7471B	PPB	29 May 2019 11:58:47	U	SMPL
K3068-02DLX4 - 1		29350	4.6122	-7471B	PPB	29 May 2019 12:01:04	U	SMPL
K3068-04DLX5 - 1		19805	3.1021	-7471B	PPB	29 May 2019 12:03:20	U	SMPL
K3068-05DLX10 - 1		52920	8.3414	-7471B	PPB	29 May 2019 12:05:37	U	SMPL
K3068-06DLX5 - 1		19745	3.0926	-7471B	PPB	29 May 2019 12:07:55	U	SMPL
K3068-07DLX4 - 1		25005	3.9248	-7471B	PPB	29 May 2019 12:10:13	U	SMPL
K3068-10DLX10 - 1		20212	3.1664	-7471B	PPB	29 May 2019 12:12:29	U	SMPL
K3068-11DLX5 - 1		18964	2.9690	-7471B	PPB	29 May 2019 12:14:46	U	SMPL
K3068-20DLX5 - 1		32109	5.0488	-7471B	PPB	29 May 2019 12:17:04	U	SMPL
CCV17 - 1	CCV17	30066	4.7255	-7471B	PPB	29 May 2019 12:19:21	U	SMPL
CCB17 - 1	CCB17	-6	-0.0324	-7471B	PPB	29 May 2019 12:21:39	U	SMPL