

LB103057


ISM02.4

INSTRUMENT ID: CU1

Linear

 μ Abs.:

45045



A= 0.0000e+000

B= 2.2347e-004

C= -6.2571e-002

Rho= 0.9999891

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.028	0.028	404	0.000	404				
0.05	0.050									
0.20	0.200	0.175	-0.025	1062	0.0 %	1062				
2.50	2.500	2.507	0.007	11497	0.0 %	11497				
5.00	5.000	4.983	-0.017	22576	0.0 %	22576				
7.50	7.500	7.504	0.004	33857	0.0 %	33857				
10.0	10.000	10.005	0.005	45049	0.0 %	45049				

 %D
 -13
 0
 0
 0
 0

LB103057
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	404	-		0 ISM02.4	PPB	5/29/2019 13:30	Std
	0.2 S0.2	1062	-		0.2 ISM02.4	PPB	5/29/2019 13:32	Std
	2.5 S2.5	11497	-		2.5 ISM02.4	PPB	5/29/2019 13:35	Std
	5 S5.0	22576	-		5 ISM02.4	PPB	5/29/2019 13:37	Std
	7.5 S7.5	33857	-		7.5 ISM02.4	PPB	5/29/2019 13:39	Std
	10 S10.0	45049	-		10 ISM02.4	PPB	5/29/2019 13:42	Std
ICV	ICV	18693	4.1148 -		ISM02.4	PPB	5/29/2019 13:47	SMPL
ICB	ICB	124	-0.0349 -		ISM02.4	PPB	5/29/2019 13:50	SMPL
CCV018	CCV018	22751	5.0217 -		ISM02.4	PPB	5/29/2019 13:52	SMPL
CCB018	CCB018	217	-0.0141 -		ISM02.4	PPB	5/29/2019 13:54	SMPL
PB120141BL	PBW141	292	0.0027 -		ISM02.4	PPB	5/29/2019 13:56	SMPL
K2940-01	MHA008	1428	0.2565 -		ISM02.4	PPB	5/29/2019 13:59	SMPL
K2940-02	MHA012	1200	0.2056 -		ISM02.4	PPB	5/29/2019 14:01	SMPL
K2940-03	MHA017	1252	0.2172 -		ISM02.4	PPB	5/29/2019 14:03	SMPL
K2940-04	MHA018	2484	0.4925 -		ISM02.4	PPB	5/29/2019 14:05	SMPL
K2940-05	MHA023	193	-0.0194 -		ISM02.4	PPB	5/29/2019 14:08	SMPL
K2940-06	MHA032	448	0.0375 -		ISM02.4	PPB	5/29/2019 14:10	SMPL
K2940-07	MHA036	321	0.0092 -		ISM02.4	PPB	5/29/2019 14:12	SMPL
K2940-08	MHA040	581	0.0673 -		ISM02.4	PPB	5/29/2019 14:15	SMPL
K2940-09	MHA043	728	0.1001 -		ISM02.4	PPB	5/29/2019 14:17	SMPL
K2940-10	MHA045	309	0.0065 -		ISM02.4	PPB	5/29/2019 14:19	SMPL
K2940-11	MHA048	619	0.0758 -		ISM02.4	PPB	5/29/2019 14:21	SMPL
K2940-12	MHA054	206	-0.0165 -		ISM02.4	PPB	5/29/2019 14:24	SMPL
K2940-13	MHA061	574	0.0657 -		ISM02.4	PPB	5/29/2019 14:26	SMPL
K2940-14	MHA062	316	0.008 -		ISM02.4	PPB	5/29/2019 14:28	SMPL
K2940-15	MHA063	244	-0.008 -		ISM02.4	PPB	5/29/2019 14:30	SMPL
K2940-16	MHA064	335	0.0123 -		ISM02.4	PPB	5/29/2019 14:33	SMPL
K2940-17	MHA068	172	-0.0241 -		ISM02.4	PPB	5/29/2019 14:35	SMPL
K2940-18	MHA069	240	-0.0089 -		ISM02.4	PPB	5/29/2019 14:37	SMPL
K2940-19	MHA070	220	-0.0134 -		ISM02.4	PPB	5/29/2019 14:40	SMPL
CCV019	CCV019	22850	5.0438 -		ISM02.4	PPB	5/29/2019 14:42	SMPL
CCB019	CCB019	219	-0.0136 -		ISM02.4	PPB	5/29/2019 14:44	SMPL
K2940-20	MHA080	253	-0.006 -		ISM02.4	PPB	5/29/2019 14:46	SMPL
K2940-21	MHA080D	257	-0.0051 -		ISM02.4	PPB	5/29/2019 14:49	SMPL
K2940-22	MHA080S	5170	1.0928 -		ISM02.4	PPB	5/29/2019 14:51	SMPL
PB120142BL	PBW142	301	0.0047 -		ISM02.4	PPB	5/29/2019 14:53	SMPL
K2941-01	MHA083	131	-0.0333 -		ISM02.4	PPB	5/29/2019 14:55	SMPL
K2941-02	MHA084	89	-0.0427 -		ISM02.4	PPB	5/29/2019 14:58	SMPL
K2941-03	MHA089	25	-0.057 -		ISM02.4	PPB	5/29/2019 15:00	SMPL
K2941-04	MHA093	27	-0.0565 -		ISM02.4	PPB	5/29/2019 15:02	SMPL
K2941-05	MHA096	1063	0.175 -		ISM02.4	PPB	5/29/2019 15:05	SMPL
K2941-06	MHA098	-25	-0.0682 -		ISM02.4	PPB	5/29/2019 15:07	SMPL
K2941-07	MHA106	49	-0.0516 -		ISM02.4	PPB	5/29/2019 15:09	SMPL
K2941-08	MHA106D	54	-0.0505 -		ISM02.4	PPB	5/29/2019 15:11	SMPL
K2941-09	MHA106S	5346	1.1321 -		ISM02.4	PPB	5/29/2019 15:14	SMPL
CCV020	CCV020	22538	4.9741 -		ISM02.4	PPB	5/29/2019 15:16	SMPL

LB103057

INSTRUMENT ID : CV1

CCB020

CCB020

193 -0.0194 -

ISM02.4 PPB

5/29/2019 15:18 SMPL