

LB103068


7470A

INSTRUMENT ID: CU1

Linear

 μ Abs.:

44091



A= 0.0000e+000

B= 2.2883e-004

C= -5.2016e-002

Rho= 0.9998545

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	158	0.000	158				
0.2	0.200	0.188	-0.012	1051	0.0 %	1051				
2.5	2.500	2.520	0.020	11241	0.0 %	11241				
5.0	5.000	5.088	0.088	22460	0.0 %	22460				
7.5	7.500	7.382	-0.118	32487	0.0 %	32487				
10.0	10.000	10.037	0.037	44091	0.0 %	44091				

LB103068 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 May 2019 15:40:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	158	-	0.0000	7470A	PPB	29 May 2019 15:44:37	S	Std
0.2 - 1	50-2	1051	-	0.2000	7470A	PPB	29 May 2019 15:46:45	S	Std
2.5 - 1	52-5	11241	-	2.5000	7470A	PPB	29 May 2019 15:48:53	S	Std
5.0 - 1	55-0	22460	-	5.0000	7470A	PPB	29 May 2019 15:51:01	S	Std
7.5 - 1	57-5	32487	-	7.5000	7470A	PPB	29 May 2019 15:53:09	S	Std
10.0 - 1	510-0	44091	-	10.0000	7470A	PPB	29 May 2019 15:55:15	S	Std
ICV65 - 1	ICV65	18928	4.2793	-	7470A	PPB	29 May 2019 15:59:48	U	SMPL
ICB65 - 1	ICB65	66	-0.0369	-	7470A	PPB	29 May 2019 16:01:54	U	SMPL
CCV21 - 1	CCV21	22071	4.9985	-	7470A	PPB	29 May 2019 16:04:00	U	SMPL
CCB21 - 1	CCB21	27	-0.0458	-	7470A	PPB	29 May 2019 16:06:06	U	SMPL
CRA - 1	CRA	1010	0.1791	-	7470A	PPB	29 May 2019 16:08:12	U	SMPL
HighStd - 1	HighStd	42489	9.6709	-	7470A	PPB	29 May 2019 16:10:18	U	SMPL
ChkStd - 1	ChkStd	30401	6.9047	-	7470A	PPB	29 May 2019 16:12:25	U	SMPL
PB120140BL - 1	PBW	44	-0.0419	-	7470A	PPB	29 May 2019 16:14:33	U	SMPL
PB120140BS - 1	LCSW	18083	4.0860	-	7470A	PPB	29 May 2019 16:16:41	U	SMPL
K3041-08 - 1	TP-2-S	123	-0.0239	-	7470A	PPB	29 May 2019 16:18:48	U	SMPL
K3041-12 - 1	TP-2-D	211	-0.0037	-	7470A	PPB	29 May 2019 16:20:56	U	SMPL
K3051-01 - 1	VNJ235	227	-0.0001	-	7470A	PPB	29 May 2019 16:23:04	U	SMPL
K3051-03 - 1	SP-1-A	304	0.0175	-	7470A	PPB	29 May 2019 16:25:12	U	SMPL
K3051-05 - 1	SP-1-B	50	-0.0406	-	7470A	PPB	29 May 2019 16:27:18	U	SMPL
CCV22 - 1	CCV22	22331	5.0580	-	7470A	PPB	29 May 2019 16:29:25	U	SMPL
CCB22 - 1	CCB22	63	-0.0376	-	7470A	PPB	29 May 2019 16:31:32	U	SMPL
K3053-03 - 1	CR-SOIL-PILE	288	0.0139	-	7470A	PPB	29 May 2019 16:43:31	U	SMPL
K3065-13 - 1	B-16-COMP	787	0.1281	-	7470A	PPB	29 May 2019 16:45:37	U	SMPL
K3065-14 - 1	B-15-COMP	463	0.0539	-	7470A	PPB	29 May 2019 16:47:44	U	SMPL
K3065-15 - 1	B-14-COMP	307	0.0182	-	7470A	PPB	29 May 2019 16:49:51	U	SMPL
K3065-16 - 1	B-14-11-12.0	377	0.0343	-	7470A	PPB	29 May 2019 16:51:58	U	SMPL
K3065-17 - 1	B-13-COMP	4	-0.0511	-	7470A	PPB	29 May 2019 16:54:05	U	SMPL
K3065-18 - 1	B-12-COMP	169	-0.0133	-	7470A	PPB	29 May 2019 16:56:13	U	SMPL
K3065-18MS - 1	B-12-COMPMS	18514	4.1846	-	7470A	PPB	29 May 2019 17:00:30	U	SMPL
K3065-18MSD - 1	B-12-COMPMSD	18731	4.2342	-	7470A	PPB	29 May 2019 17:02:37	U	SMPL
CCV23 - 1	CCV23	22716	5.1461	-	7470A	PPB	29 May 2019 17:04:43	U	SMPL
CCB23 - 1	CCB23	4	-0.0511	-	7470A	PPB	29 May 2019 17:06:50	U	SMPL
PB120143BL - 1	PBW	196	-0.0072	-	7470A	PPB	29 May 2019 17:08:56	U	SMPL
PB120143BS - 1	LCSW	18331	4.1427	-	7470A	PPB	29 May 2019 17:11:03	U	SMPL
K3036-03 - 1	FEQC-FB-052319	-31	-0.0591	-	7470A	PPB	29 May 2019 17:13:09	U	SMPL
K3037-01 - 1	MW6-R1-1	90	-0.0314	-	7470A	PPB	29 May 2019 17:15:16	U	SMPL
K3037-02 - 1	MW6-R1-2	61	-0.0381	-	7470A	PPB	29 May 2019 17:17:24	U	SMPL
K3037-03 - 1	FB-20190522	1	-0.0518	-	7470A	PPB	29 May 2019 17:19:31	U	SMPL
K3037-04 - 1	FB-20190523	-9	-0.0541	-	7470A	PPB	29 May 2019 17:21:39	U	SMPL
K3037-05 - 1	MW5-R1-1	-12	-0.0548	-	7470A	PPB	29 May 2019 17:23:47	U	SMPL
K3037-06 - 1	MW2-R1-1	22	-0.0470	-	7470A	PPB	29 May 2019 17:25:54	U	SMPL
K3037-06DUP - 1	MW2-R1-1DUP	21	-0.0472	-	7470A	PPB	29 May 2019 17:28:01	U	SMPL
CCV24 - 1	CCV24	21865	4.9514	-	7470A	PPB	29 May 2019 17:30:08	U	SMPL
CCB24 - 1	CCB24	53	-0.0399	-	7470A	PPB	29 May 2019 17:32:14	U	SMPL
K3037-07 - 1	K3037-06MS	17668	3.9910	-	7470A	PPB	29 May 2019 17:34:22	U	SMPL
K3037-08 - 1	K3037-06MSD	16843	3.8022	-	7470A	PPB	29 May 2019 17:36:28	U	SMPL
K3037-10 - 1	MW2-R1-2	117	-0.0252	-	7470A	PPB	29 May 2019 17:38:35	U	SMPL
K3037-11 - 1	DUP-20190523	250	0.0052	-	7470A	PPB	29 May 2019 17:40:42	U	SMPL
K3037-12 - 1	MW2-R1-3	263	0.0082	-	7470A	PPB	29 May 2019 17:42:49	U	SMPL
K3040-01 - 1	0444-FIELD-BLANK-1	45	-0.0417	-	7470A	PPB	29 May 2019 17:44:57	U	SMPL
K3040-02 - 1	0445-FIELD-BLANK-2	36	-0.0438	-	7470A	PPB	29 May 2019 17:47:05	U	SMPL
K3041-13 - 1	TP-2	1041	0.1862	-	7470A	PPB	29 May 2019 17:49:12	U	SMPL
K3073-01 - 1	TWP-1B	5	-0.0509	-	7470A	PPB	29 May 2019 17:51:21	U	SMPL
K3073-02 - 1	TWP-6B	118	-0.0250	-	7470A	PPB	29 May 2019 17:53:28	U	SMPL
CCV25 - 1	CCV25	21859	4.9500	-	7470A	PPB	29 May 2019 17:55:35	U	SMPL
CCB25 - 1	CCB25	18	-0.0479	-	7470A	PPB	29 May 2019 17:57:42	U	SMPL
PB120158BL - 1	PBW	100	-0.0291	-	7470A	PPB	29 May 2019 17:59:49	U	SMPL
PB120158BS - 1	LCSW	18218	4.1169	-	7470A	PPB	29 May 2019 18:01:56	U	SMPL
K3041-14 - 1	TP-2	-44	-0.0621	-	7470A	PPB	29 May 2019 18:04:03	U	SMPL
K3041-14DUP - 1	TP-2DUP	37	-0.0435	-	7470A	PPB	29 May 2019 18:06:10	U	SMPL
K3041-14MS - 1	TP-2MS	18322	4.1407	-	7470A	PPB	29 May 2019 18:08:17	U	SMPL
K3041-14MSD - 1	TP-2MSD	18189	4.1102	-	7470A	PPB	29 May 2019 18:10:24	U	SMPL
PB120163BL - 1	PBW	58	-0.0387	-	7470A	PPB	29 May 2019 18:12:32	U	SMPL
PB120163BS - 1	LCSW	18144	4.0999	-	7470A	PPB	29 May 2019 18:14:40	U	SMPL
K3054-08 - 1	CULVER-CBTC-WC1B-052419	-38	-0.0607	-	7470A	PPB	29 May 2019 18:16:48	U	SMPL
K3074-06 - 1	TP-3B-S	48	-0.0410	-	7470A	PPB	29 May 2019 18:18:55	U	SMPL
CCV26 - 1	CCV26	22353	5.0631	-	7470A	PPB	29 May 2019 18:21:02	U	SMPL
CCB26 - 1	CCB26	79	-0.0339	-	7470A	PPB	29 May 2019 18:23:09	U	SMPL
K3074-10 - 1	TP-3B-D	30	-0.0452	-	7470A	PPB	29 May 2019 18:25:22	U	SMPL
K3074-10DUP - 1	TP-3B-DDUP	8	-0.0502	-	7470A	PPB	29 May 2019 18:27:32	U	SMPL
K3074-10MS - 1	TP-3B-DMS	17890	4.0418	-	7470A	PPB	29 May 2019 18:29:39	U	SMPL
K3074-10MSD - 1	TP-3B-DMSD	17040	3.8473	-	7470A	PPB	29 May 2019 18:31:49	U	SMPL
PB120140TB - 1	PB120072TB	-83	-0.0710	-	7470A	PPB	29 May 2019 18:33:58	U	SMPL
PB120158TB - 1	PB120115TB	206	-0.0049	-	7470A	PPB	29 May 2019 18:36:05	U	SMPL
PB120163TB - 1	PB120116TB	426	0.0455	-	7470A	PPB	29 May 2019 18:38:13	U	SMPL
K3065-18LX5 - 1		176	-0.0117	-	7470A	PPB	29 May 2019 18:40:21	U	SMPL
K3065-18A - 1		18954	4.2853	-	7470A	PPB	29 May 2019 18:42:29	U	SMPL
K3037-06LX5 - 1		-33	-0.0596	-	7470A	PPB	29 May 2019 18:44:37	U	SMPL
CCV27 - 1	CCV27	22567	5.1121	-	7470A	PPB	29 May 2019 18:46:44	U	SMPL
CCB27 - 1	CCB27	87	-0.0321	-	7470A	PPB	29 May 2019 18:48:52	U	SMPL

LB103068 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 May 2019 15:40:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
K3037-06A - 1		16675	3.7638		-7470A	PPB	29 May 2019 18:50:59	U	SMPL
K3041-14LX5 - 1		35	-0.0440		-7470A	PPB	29 May 2019 18:53:06	U	SMPL
K3041-14A - 1		18179	4.1079		-7470A	PPB	29 May 2019 18:55:14	U	SMPL
K3074-10LX5 - 1		13	-0.0490		-7470A	PPB	29 May 2019 18:57:21	U	SMPL
K3074-10A - 1		17853	4.0333		-7470A	PPB	29 May 2019 18:59:28	U	SMPL
K3065-18DUP - 1	B-12-COMPDUP	84	-0.0328		-7470A	PPB	29 May 2019 19:01:36	U	SMPL
CCV28 - 1	CCV28	22237	5.0365		-7470A	PPB	29 May 2019 19:03:44	U	SMPL
CCB28 - 1	CCB28	34	-0.0442		-7470A	PPB	29 May 2019 19:05:52	U	SMPL