



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 5/30/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 16:45
In Date: 05/29/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:10
Out Date: 05/30/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103069

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2642-07	OK-02-052919	1	1.17	9.88	9.04	90.4
K2648-03	TR-04-052919	2	1.15	9.9	9.47	95.1
K3017-25	A51G4	3	1.00	2.00	2.00	100.0
K3017-26	A51G9	4	1.00	2.00	2.00	100.0
K3080-01	MASONRY-COMPOSITE	5	1.00	2.00	2.00	100.0
K3084-01	TEST-PIT-6	6	1.15	9.82	8.33	82.8
K3084-02	TEST-PIT-6	7	1.15	9.82	8.33	82.8
K3084-03	TEST-PIT-7	8	1.12	9.67	7.65	76.4
K3084-04	TEST-PIT-7	9	1.12	9.67	7.65	76.4
K3084-05	TEST-PIT-8	10	1.15	9.75	7.92	78.7
K3084-06	TEST-PIT-8	11	1.15	9.75	7.92	78.7
K3084-07	TEST-PIT-9	12	1.18	9.92	7.79	75.6
K3084-08	TEST-PIT-9	13	1.18	9.92	7.79	75.6
K3084-09	TEST-PIT-10	14	1.15	9.65	8.44	85.8
K3084-10	TEST-PIT-10	15	1.15	9.65	8.44	85.8
K3089-01	T-2A	16	1.11	9.59	8.68	89.3
K3089-02	T-3A	17	1.13	9.8	8.74	87.8
K3090-01	TP-4-S1	18	1.16	9.57	9.26	96.3
K3090-02	TP-4-S2	19	1.19	9.66	9.44	97.4
K3090-03	TP-4-S	20	1.14	9.83	9.54	96.7
K3090-05	TP-6-S1	21	1.15	9.94	9.28	92.5
K3090-06	TP-6-S2	22	1.16	9.7	9.44	97.0
K3090-07	TP-4-D1	23	1.13	9.96	9.41	93.8
K3090-08	TP-4-D2	24	1.1	9.93	9.52	95.4
K3090-09	TP-4-D	25	1.13	9.8	9.53	96.9

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$