



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103072

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/29/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	6.99	05/29/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	05/29/2019	17:03
4	ICV	1	Water	NA	NA	20.1	7.01	05/29/2019	17:04
5	CCV1	1	Water	NA	NA	20.2	2.01	05/29/2019	17:05
6	K3041-13	1	Water	NA	NA	11.0	6.40	05/29/2019	17:07
7	K3041-13DUP	1	Water	NA	NA	11.1	6.41	05/29/2019	17:08
8	CCV2	1	Water	NA	NA	20.2	12.02	05/29/2019	17:10

LB 103072

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K3085	WorkList ID :	126708	Date :	05-29-2019 15:33:57				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/23/2019	Water	K3041-13 -17	pH	Cool 4 deg C	PSEG01	O21	TP-2	05/23/2019	9040C

Date/Time 05-29-19 15:40
Received by: JD
Relinquished by: CD

Date/Time 05-29-19 17:00
Received by: CR
Relinquished by: JG