

LB103083


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46411



A= 0.0000e+000

B= 2.1816e-004

C= -4.6141e-002

Rho= 0.9999029

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.037	0.037	379	0.000	379					1
0.05	0.050										5
0.20	0.200	0.210	0.010	1173	0.0 %	1173					0
2.50	2.500	2.503	0.003	11684	0.0 %	11684					-1
5.00	5.000	4.941	-0.059	22862	0.0 %	22862					-1
7.50	7.500	7.431	-0.069	34273	0.0 %	34273					-1
10.0	10.000	10.079	0.079	46411	0.0 %	46411					1

RS 05-30-19

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	379	-		0 ISM02.4	PPB	5/30/2019 11:28	Std
	0.2 S0.2	1173	-		0.2 ISM02.4	PPB	5/30/2019 11:31	Std
	2.5 S2.5	11684	-		2.5 ISM02.4	PPB	5/30/2019 11:33	Std
	5 S5.0	22862	-		5 ISM02.4	PPB	5/30/2019 11:35	Std
	7.5 S7.5	34273	-		7.5 ISM02.4	PPB	5/30/2019 11:37	Std
	10 S10.0	46411	-		10 ISM02.4	PPB	5/30/2019 11:40	Std
ICV	ICV	19368	4.1791 -		ISM02.4	PPB	5/30/2019 11:42	SMPL
ICB	ICB	236	0.0053 -		ISM02.4	PPB	5/30/2019 11:45	SMPL
CCV029	CCV029	22712	4.9086 -		ISM02.4	PPB	5/30/2019 11:47	SMPL
CCB029	CCB029	161	-0.011 -		ISM02.4	PPB	5/30/2019 11:49	SMPL
PB120141BL	PBW141	361	0.0326 -		ISM02.4	PPB	5/30/2019 11:52	SMPL
K2940-01	MHA008	1598	0.3025 -		ISM02.4	PPB	5/30/2019 11:54	SMPL
K2940-02	MHA012	1245	0.2255 -		ISM02.4	PPB	5/30/2019 11:56	SMPL
K2940-03	MHA017	1552	0.2924 -		ISM02.4	PPB	5/30/2019 11:58	SMPL
K2940-04	MHA018	2784	0.5612 -		ISM02.4	PPB	5/30/2019 12:01	SMPL
K2940-05	MHA023	215	0.0008 -		ISM02.4	PPB	5/30/2019 12:03	SMPL
K2940-06	MHA032	442	0.0503 -		ISM02.4	PPB	5/30/2019 12:05	SMPL
K2940-07	MHA036	306	0.0206 -		ISM02.4	PPB	5/30/2019 12:07	SMPL
K2940-08	MHA040	577	0.0797 -		ISM02.4	PPB	5/30/2019 12:10	SMPL
K2940-09	MHA043	706	0.1079 -		ISM02.4	PPB	5/30/2019 12:12	SMPL
K2940-10	MHA045	316	0.0228 -		ISM02.4	PPB	5/30/2019 12:14	SMPL
K2940-11	MHA048	627	0.0906 -		ISM02.4	PPB	5/30/2019 12:17	SMPL
K2940-12	MHA054	207	-0.001 -		ISM02.4	PPB	5/30/2019 12:19	SMPL
K2940-13	MHA061	569	0.078 -		ISM02.4	PPB	5/30/2019 12:21	SMPL
K2940-14	MHA062	314	0.0224 -		ISM02.4	PPB	5/30/2019 12:23	SMPL
K2940-15	MHA063	241	0.0064 -		ISM02.4	PPB	5/30/2019 12:26	SMPL
K2940-16	MHA064	338	0.0276 -		ISM02.4	PPB	5/30/2019 12:28	SMPL
K2940-17	MHA068	186	-0.0056 -		ISM02.4	PPB	5/30/2019 12:30	SMPL
K2940-18	MHA069	247	0.0077 -		ISM02.4	PPB	5/30/2019 12:32	SMPL
K2940-19	MHA070	262	0.011 -		ISM02.4	PPB	5/30/2019 12:35	SMPL
CCV030	CCV030	23093	4.9918 -		ISM02.4	PPB	5/30/2019 12:37	SMPL
CCB030	CCB030	219	0.0016 -		ISM02.4	PPB	5/30/2019 12:39	SMPL
K2940-20	MHA080	236	0.0053 -		ISM02.4	PPB	5/30/2019 12:42	SMPL
K2940-21	MHA080D	213	0.0003 -		ISM02.4	PPB	5/30/2019 12:44	SMPL
K2940-22	MHA080S	5266	1.1027 -		ISM02.4	PPB	5/30/2019 12:46	SMPL
PB120142BL	PBW142	341	0.0283 -		ISM02.4	PPB	5/30/2019 12:48	SMPL
K2941-01	MHA083	93	-0.0259 -		ISM02.4	PPB	5/30/2019 12:51	SMPL
K2941-02	MHA084	78	-0.0291 -		ISM02.4	PPB	5/30/2019 12:53	SMPL
K2941-03	MHA089	25	-0.0407 -		ISM02.4	PPB	5/30/2019 12:55	SMPL
K2941-04	MHA093	23	-0.0411 -		ISM02.4	PPB	5/30/2019 12:57	SMPL
K2941-05	MHA096	842	0.1375 -		ISM02.4	PPB	5/30/2019 13:00	SMPL
K2941-06	MHA098	-19	-0.0503 -		ISM02.4	PPB	5/30/2019 13:02	SMPL
K2941-07	MHA106	82	-0.0283 -		ISM02.4	PPB	5/30/2019 13:04	SMPL
K2941-08	MHA106D	71	-0.0307 -		ISM02.4	PPB	5/30/2019 13:07	SMPL
K2941-09	MHA106S	5275	1.1046 -		ISM02.4	PPB	5/30/2019 13:09	SMPL
PB120141TB	LEB141	349	0.03 -		ISM02.4	PPB	5/30/2019 13:11	SMPL

LB103083

INSTRUMENT ID : CV1

PB120142TB	LEB141	358	0.032 -	ISM02.4	PPB	5/30/2019 13:13	SMPL
CCV031	CCV031	22859	4.9407 -	ISM02.4	PPB	5/30/2019 13:16	SMPL
CCB031	CCB031	244	0.0071 -	ISM02.4	PPB	5/30/2019 13:18	SMPL